Materials and Methods: Patients were treated using the gated RapidArc technique with a Varian Novalis Truebeam linear accelerator and a novel application, Intrafraction Motion Review (IMR, Varian medical systems) allowing kV/MV images acquisition during dose delivery. Before treatment simulation, three gold markers were implanted inside the prostate using transrectal ultrasound guidance. Hyaluronic acid was injected to separate the rectum from the prostate. Dose delivery consisted in two arcs using RapidArc modality with 600 MU/min maximum dose rate. An equivalent dose of 84 Gy over 20 fractions at 3.1 Gy/fraction was delivered.

Before the treatment delivery, the RPM block placed on patient abdominal surface generated the gating signal use for IMR acquisition. The patient alignment was then set based on markers using daily cone beam CT (CBCT) and kV/kV matching. During the treatment, kV images were acquired at each exhale phase of the breathing cycle and the positions of the fiducial markers were compared to their expected positions represented by a green circle with a diameter corresponding to the fiducial length.

We reported here differences between expected and real gold seeds position for seven patients.

Results: Fraction time delivery (time to deliver the total dose) was 16 ± 6 min. The number of KV images acquired during VMAT delivery was 26 ± 15 with a maximum of 64. Average deviation in cranio caudal direction was analyzed on each images leading to a mean deviation of 0.29 [0-1.4] mm. Antero posterior or lateral shifts were difficult to analyzed in a post processing phase as they were not visible on all KV images but online correction with a shift of fiducial outside the green circles was necessary for only two sessions out 120

Conclusions: Intrafraction KV verification is possible with the IMR system available on Truebeam. Increased time session is acceptable even if the process has to actually be associated to a respiratory gate treatment. New version of the software be available soon and will allow this kind of online verification without the need of RPM system.

EP-1657

Classification of radiotherapy lung patients by their characterized respiratory patterns

J. Cortés-Rodicio¹, G. Sanchez-Merino¹, R. Lope-Lope¹, T. Martín-González¹, M.A. García-Fidalgo¹

¹Hospital Universitario Alava, Física médica y protección radiológica, Vitoria-Gasteiz, Spain

Purpose/Objective: Obtaining the respiratory patterns of radiotherapy lung patients is mandatory to know the uncertainties associated to radiotherapy treatments. It gives also the possibility to study the stability of the patients and classify them through parameters like the amplitude and frequency of the pattern. The aim of this work is to develop a patient classification method from the respiratory pattern parameters.

Materials and Methods: The treatment position of radiotherapy patients is followed-up using kV guidance system XVI CBCT (Elekta Oncology Systems, Crawley, UK). A home-made software based on ImageJ has been developed to extract the respiratory patterns from each kV imageset. This software based on the Amsterdam Shroud method, also analyzes the pattern, given as a result a statistical table with the amplitudes and respiratory cycle time as well as their standard deviations.

A patient classification system has been created. On this system we distinguish four different groups based on the amplitudes and cycle of the respiratory motion (Fig. 1a):

- Standard patients: amplitude ~ 6 mm, cycle ~ 4 s
- Polypnea patients: amplitude < 4 mm, cycle < 3 s
- Hyperpnea patients: amplitude > 10 mm, cycle < 3 s
- Bradypnea patients: amplitude > 10 mm, cycle > 5 s

Results: Nine patients with lung cancer with a total number of 153 XVI acquisitions were studied. The XVI - CBCT's were analyzed to extract the respiratory patterns and their statistical parameters. The amplitudes and cycle times were plotted on Fig. 1b. It can be noticed that there is a cluster of patients with standard respiratory patterns, with amplitudes around 6mm and cycle times around 4s. Apart from that, there is a patient that can be classified as 'bradypnea' and another one as 'hyperpnea'. In both cases the amplitude is > 10 mm because the studied lung tries to compensate the respiratory deficit of the cancer damaged lung. There is another patient who presents 'polypnea' in which both amplitude and cycle time are below the standard parameters, because the lung studied was the damaged one. This patient classification gives us extra information to perform more suitable radiotherapy treatments.

Conclusions: A patient classification system based on the amplitude and time cycle of the respiratory pattern has been shown. We found that the nine patients studied can be classified according to one of the four established groups. Anyway, more patients should be analyzed to improve the classification system and perform further studies.

EP-1658

Does the delivery technique impact the effect of respiratory motion in stereotactic ablative body radiotherapy?

A. Dimitriadis¹, M. Hussein², S. Jafari³, K.J. Kirkby⁴, A. Nisbet³, C.H. Clark²

¹University of Surrey, Centre of Doctoral Training for Next Generation Accelerators, Guildford, United Kingdom

²Royal Surrey County Hospital, Medical Physics, Guildford, United Kingdom

³University of Surrey, Centre for Nuclear and Radiation Physics, Guildford, United Kingdom

⁴University of Surrey, Ion Beam Centre, Guildford, United Kingdom

Purpose/Objective: To examine whether the impact of respiratory motion on the delivered dose distributions for a range of tumour sizes in SABR, differs between 3DCRT, sliding window IMRT and RapidArc.

Materials and Methods: Three patient data sets with different lung tumour sizes were selected: T1=6cc, T2=31cc and T3=60cc. Each was planned in Eclipse for SABR using 3DCRT, sliding window IMRT and VMAT, creating 9 treatment plans which were then delivered to a dynamic thorax phantom. The phantom was programmed to move at a typical patient breathing amplitude of 15mm with a period of 5 seconds and Varian linacs were used for the delivery. EBT3 Gafchromic film was used in coronal and sagittal planes for measuring dose distributions. Static phantom measurements were compared with the TPS calculated plans to establish agreement between expected and measured dose distributions without motion, using the software OmniPro ImRT. Comparisons of static and dynamic phantom measurements followed. Global gamma analysis was used to carry out a relative comparison between the three delivery techniques. Five regions of the gamma index map (Middle, Proximal Left, Proximal Right, Distal Left, Distal Right) were analysed to quantify the differences along the axis of motion. The criteria used for the gamma analysis were 3%/3mm, with a threshold of 20%.

Results: The setup and delivery accuracy was confirmed by the agreement between planned and static delivered dose distributions. The average percentage of pixels passing was 100% (T1), 100% (T2) and 98.46% (T3). The comparison of films with and without motion gave lower percentages of pixels passing, ranging between 33.68 - 59.94% (T1), 47.86 - 61.77% (T2) and 43.44 - 64.32% (T3). Comparison of the delivery technique, showed passing rates of 33.63 - 52.25% (3DCRT), 43.44 - 64.52% (IMRT) and 46.58 - 56.08% (VMAT). Analysis of the five regions for

all delivery techniques gave averages of 93.76% (Middle), 58.7% (Proximal) and 12.8% (Distal). For 3DCRT results were 87.08% (Middle), 46.52% (Proximal) and 12.54% (Distal), for IMRT were 96.45%, 69.20%, 14.14% and for VMAT 97.75%, 60.39% and 11.71%, respectively.

Figure showing (a) a static phantom dose distribution map, (b) the corresponding dynamic phantom dose distribution map and (c) the gamma index map with blue pixels passing and red pixels failing. The five regions of interest and their borders are also annotated on (c).

Conclusions: The results are indicative of the intra-fractional respiratory motion-induced dosimetric inaccuracies caused in three SABR delivery techniques. On average, the impact is greatest in the distal regions, significant in the proximal regions, whereas the middle region is less susceptible to these effects. No noticeable difference was observed between coronal and sagittal planes. The results also suggest that the effect of motion is greater in the proximal regions for 3DCRT in relation to the other techniques, particularly with smaller tumour sizes.

EP-1659

An observatory study of organ motion in patients with cervical cancer using cinematic-magnetic resonance imaging

N. Jastaniyah¹, B. Murray², K. Wachowicz², A. Yahya², R. Pearcey¹

¹Cross Cancer Institute, Radiation Oncology, Edmonton, Canada

²Cross Cancer Institute, Medical Physics, Edmonton, Canada

Purpose/Objective: To observe the nature and speed of movements in different parts of the uterus in patients with cervical cancer undergoing radiation therapy.

Materials and Methods: Three patients with biopsy-proven cervical cancer underwent baseline and weekly cinematic-magnetic resonance imaging scans of the pelvis during the period of radiation therapy delivery. In each scan, three planes were obtained (sagittal, uterine, and cervical) each acquired over a period of 6 seconds and the sequencing cycle was repeated over a period of 20 minutes. Points of interests representing different parts of the uterus were identified and traced in all planes. The positional variation of each point of interest was measured in comparison to the location of the reference point of the first plane acquired. The measurements were then graphed over the scanning period.

Results: Three main types of motion can be observed within the chosen temporal resolution: fine oscillatory movements reflecting propagation of bowel peristalsis, abrupt more significant shifts secondary to changes in rectal filling with gas, and gradual transition in the position of the uterus due to bladder filling. The pattern of motion varies considerably

between patients and changes during the treatment delivery. Multiple factors govern organ motion in the female pelvis and these can be summarized as follows: (1) anatomical orientation of the uterus, (2) pre-existing pathologies in the pelvis, such as endometriosis or uterine fibroids, resulting in anatomical aberrations of pelvic organs, (3) disease extent at the time of diagnosis, (4) response to treatment, (5) pathological effects of radiation on adjacent organs e.g. radiation cystitis, (6) patients compliance with bladder or bowel preparation instructions. Fig.1 shows quantitative examples of some of the POIs at week 1 and week 6 among the three patients.

Conclusions: Movements of the different parts of the uterus are complex and some of its components are random and difficult to predict. The nature of this motion should be taken into consideration when developing systems that aim at tumor tracking.

EP-1660

2D/3D image registration for various gantry angles for online tumor motion tracking

E. Steiner¹, H. Furtado², C. Viehböck¹, W. Birkfellner², D. Georg¹, M. Stock¹

¹Medical University of Vienna, Department of Radiation Oncology & Christian Doppler Laboratory for Medical Radiation Research for Radiation Oncology, Vienna, Austria

²Medical University of Vienna, Center for Medical Physics and Biomedical Engineering & Christian Doppler Laboratory for Medical Radiation Research for Radiation Oncology, Vienna, Austria

Purpose/Objective: Several intrafraction motion management strategies for lung tumors aim to minimize uncertainties during irradiation in order to improve the therapeutic ratio. Breathing motion models fail in the presence of irregular motion and marker implantation can lead to severe complications. Therefore, we evaluated a purely intensity based 2D/3D real-time image registration as a promising alternative. For delivery of SBRT lung treatments typically seven to 13 different gantry angles are used and for certain projection angles the overlap of different structures might limit the intensity based approach. Thus, we aimed to answer the question: Is automatic 2D/3D registration possible for angles other than AP or PA direction?

Materials and Methods: Fluoroscopic movies were acquired during SBRT irradiation (7 beams over $\sim 180^\circ$) of 5 patients suffering from NSCLC using the Elekta XVI system and retrospectively analyzed. Patients were treated in a vacuum pillow using an abdominal pressure plate. For each patient the planning CT scan was registered to the CBCT scan for setup. Fluoroscopic images were then registered to DRRs generated from the planning CT for a region of interest (PTV or CTV) using an in-house developed software. In the next step the motion in CC direction was extracted. The registration was verified by comparison with the diaphragm motion, which was annotated in addition, and visual inspection of the moving CTV contours, resulting from the registration process. A Fourier transform was performed to analyze the motion frequency components and the Pearson correlation between diaphragm motion and the registration results was calculated.

Results: 2D/3D image registration was successful for 4-7 gantry angles per patient. Difficulties in the registration were observed where other radioopaque structures (e. g. heart) overlaid the tumor in the projection images. Fig. 1 illustrates the tumor motion in CC direction, the