NATURAL FREQUENCY ANALYSIS FOR OSSEOINTEGRATION TRANS- FEMORAL IMPLANT

By: Fei Shao
Supervisors: Dr Wei Xu
Dr David Ewins

**Thesis submitted for the Degree of Doctor of
Philosophy from the University of Surrey**

**Centre for Biomedical Engineering,
School of Engineering, University of Surrey,
Guildford, Surrey, GU2 7XH, UK.**

**-August 2007-
© Fei Shao 2007**

ProQuest Number: U229454

All rights reserved

INFORMATION TO ALL USERS

The quality of this reproduction is dependent upon the quality of the copy submitted.

In the unlikely event that the author did not send a complete manuscript and there are missing pages, these will be noted. Also, if material had to be removed, a note will indicate the deletion.

ProQuest U229454

Published by ProQuest LLC (2019). Copyright of the Dissertation is held by the Author.

All rights reserved.

This work is protected against unauthorized copying under Title 17, United States Code
Microform Edition © ProQuest LLC.

ProQuest LLC.
789 East Eisenhower Parkway
P.O. Box 1346
Ann Arbor, MI 48106 – 1346

Abstract

Osseointegrated trans-femoral implants are a promising orthopaedic anchoring method to attach prosthetic limbs. The clinical success of this technique depends on the biological osseointegration achieved between the implant and bone. In this thesis, resonant characteristics of the trans-femoral osseointegration implant systems were investigated thoroughly by means of natural frequency (NF) analysis.

The study consists of three parts, Finite Element (FE) simulation, *in-vitro* and *in-vivo* experiments. In the FE analysis study, various clinical conditions were simulated, such as, size of contact area, mechanical properties of the interface, femur length, and mechanical properties of soft tissue. The results showed that both the area and elastic modulus of the osseointegration interface affect the NF of the implant system.

The *in-vitro* experiments were carried out using a silicone interface layer to simulate an implant-bone interface. The results showed a linear correlation between the NF and the elastic modulus of the interface. Meanwhile, the NF was found decreased when the length of the abutment increased.

The *in vivo* experiments were carried out on three patients during their rehabilitation programme. Gradual NF increases were found with patients who finished the rehabilitation program within normal time. NF decreases were commonly observed after the first weight bearing. Moreover, all of the NF results after the weight bearing were significant lower than the NF results before the weight bearing ($p<0.01$).

The correlation found between the NF and mechanical characteristics, such as contact area and elastic modulus, of osseointegration interface in all of above three studies suggested that the NF analysis is an effective method to reveal the osseointegration condition during the rehabilitation. As the future work the author suggested further *in vivo* investigations with established techniques including radiography and histology would help to confirm the correlation uncovered in study. The suggestion of further improvement of the NF analysis software and FE models were also given.

Acknowledgments

I would like to acknowledge the help and contribution of the following people and organizations:

I would like to thank my supervisors Dr Wei Xu and Dr Ewins for their continuous support, advice, encouragement and patience. This thesis and work would not have been possible without his help.

I would like to thank the School of Engineering at the University of Surrey for the financial support towards the cost of my studies.

I would like to thank the Henry Lester Trust for the financial support during my studies.

I would like to thank friends, colleagues and staff at the Centre for Biomedical Engineering for their advices in my studies and helps during my difficult time.

I am also very grateful to all the patients who took part in my experimental studies and their willingness to offer their help and time generously.

I would like to thank my parents for their encouragements and financial support. I would like to thank my wife for her caring, love, endless support, and understanding. I would like to express my sincere appreciation to her for sharing this time with me.

Finally, I would like to thank God for giving me the wisdom, hope and comfort and thank the brothers and sisters in Christ for their prayers and helps.

Table of Contents

ABSTRACT	I
ACKNOWLEDGMENTS	II
TABLE OF CONTENTS.....	III
LIST OF FIGURES.....	VIII
LIST OF TABLES	XV
CHAPTER 1 INTRODUCTION	1
1.1 Background	1
1.2 Aim and objectives.....	2
1.3 Layout of the thesis	3
CHAPTER 2 LITERATURE REVIEW	4
2.1 Background	4
2.1.1 Definition of osseointegration.....	4
2.1.2 Experimental studies	4
2.1.3 Titanium properties	5
2.1.3.1 Material response	7
2.1.3.2 Host response	7
2.1.4 Clinical applications of osseointegration	9
2.1.4.1 Dental	9
2.1.4.2 Facial prostheses	9
2.1.4.3 Finger joint prostheses	10
2.1.5 Rehabilitation of the trans-femoral amputee with an osseointegrated prosthesis.....	11
2.2 Applications of natural frequency measurement in clinical studies	14

2.2.1 Assessment of bone healing after fracture	15
2.2.2 Assessment of dental implant stability	22
2.2.3 Diagnosis the loosening of the hip prosthesis	29
2.3 Methods of Vibration Measurement	31
2.3.1 Excitation Methods	32
2.3.2 Vibration measurement	33
2.4 Models for vibration analysis in clinical studies	34
2.5 Comparison of in vitro, in vivo and FEA results	44
2.6 Summary	45
 CHAPTER 3 FUNDAMENTAL THEORIES AND COMPUTATION	
METHOD OF SUPPORTING THIS RESEARCH	47
3.1 Vibration theory	47
3.2 Fast Fourier transform	49
3.2.1 Fourier series	49
3.2.2 Discrete Fourier Transform and the FFT	50
3.2.2.1 The Discrete Fourier Transform	50
3.2.2.2 The Fast Fourier Transform	51
3.3 Finite element method	52
3.3.1 Pre-processing	52
3.3.2 Analysis	53
3.3.3 Visualisation	53
3.3.4 Modal analysis using FEM	54
 CHAPTER 4 FE MODELLING OF OSSEOINTEGRATION IMPLANT	
SYSTEM	56
4.1 Finite element model generation	56
4.1.1 Real geometry amputated distal femur model generation	57
4.1.2 Edge detection	58

**CHAPTER 5 MEASUREMENT SYSTEM DEVELOPMENT AND
EXPERIMENT PROCEDURE 92**

5.3.3 Signal conditioner	98
5.3.4 A/D converter and Pcmcia Card	98
Voltage signals	98
Resolution	98
Input channels	99
Throughput.....	100
5.3.5 Battery supply	101
5.4 Software for data collection and analysis.....	102
5.4.1 InstruNet World	102
5.4.2 FFT analyzer	103
5.5 Validation and calibration of the data collection system	105
5.5.1 Accelerometer calibration	105
5.5.2 Validation of the data collection system	105
5.6 In vitro experiment.....	107
5.6.1 The setup of the <i>in vitro</i> experiments.....	107
5.6.2 The procedure of the <i>in vitro</i> experiments	109
5.6.3 Testing the elastic modulus of three silicone rubbers	109
5.7 In vivo experiment	111
5.7.1 The preparation of in vivo experiment.....	111
5.7.2 The procedure of in vivo experiment.....	113
5.7.3 The rehabilitation process of the three patients.....	114
CHAPTER 6 RESULTS	117
6.1 In vitro results	117
6.1.1 Detection of NF.....	117
6.1.2 The NF test of different abutment length and implant diameter.....	121
6.1.3 NF test on different elastic modulus interface	122
6.2 In vivo results.....	124
6.2.1 Detection of NF.....	124
6.2.2 Results of patient A.....	129

6.2.3 Results of patient B	132
6.2.4 Results of patient C	134
CHAPTER 7 DISCUSSION.....	138
7.1 In vitro study	138
7.1.1 Detection of NF.....	138
7.1.2 Implant size	139
7.1.3 Interface between the implant and the bone.....	140
7.2 In vivo study.....	142
7.2.1 The detection of NF	142
7.2.2 NF and rehabilitation time	144
7.2.3 NF and the weight bearing.....	146
7.2.4 NF and pain index	148
7.3 Limitations	149
CHAPTER 8 CONCLUSIONS AND FUTURE WORK.....	151
REFERENCES.....	155
APPENDIX1 CALCULATION OF THE IMPULSE	166
APPENDIX2 PROPOSAL OF THE CLINICAL STUDY	168
APPENDIX3 INFORMATION SHEET	174
APPENDIX 4 CONSENT FORM.....	177
APPENDIX 5 CERTIFICATE OF THE CALIBRATION OF ACCELEROMETERS	179
APPENDIX 6 PAIN INDEX FORM.....	185
APPENDIX 7 C++ CODE FOR EDGE DETECTION.....	189
APPENDIX 8 EVALUATION OF THE VIBRATION EQUATION	201
APPENDIX 9 PUBLICATION	207

List of Figures

Figure 1. An artist's view of the titanium/tissue interface.	6
Figure 2. Representation of events at the bone-implant interface. (a) Protein adsorption from blood and tissue fluids, (b) protein desorption, (c) surface changes and material release, (d) inflammatory and connective tissue cells approach the implant, (e) possible targeted release of matrix proteins and selected adsorption of proteins such as BSP and OPN, (f) formation of <i>lamina limitans</i> and adhesion of osteogenic cells, (g) bone deposition on both the exposed bone and implant surfaces, (h) remodeling of newly formed bone.	6
Figure 3. Electron micrographs illustrating the presence of a thin, electron-dense cement line (CL) at the interface between titanium implants and bone in (A) a rat tibial model and (B) an implant retrieved from a human. IS, implant space; Oc, Osteocyte.	8
Figure 4. Bar graph comparing the rate of bone deposition in a direction away from and towards the implant. (mean and SD)	8
Figure 5. Major maxillofacial defect after tumor resection and radiation therapy.	10
Figure 6. Example of osseointegration arthroplasty.	11
Figure 7. Trans-femoral amputee with osseointegrated limb prosthesis.	11
Figure 8 Schematic drawing of socket prosthesis (left) and osseointegrated prosthesis . The arrows show the reaction force support the body weight.	12
Figure 9 Operation for stage 1 and 2; (a) implant insertion, (b) abutment installation.....	13
Figure 10 The apparatus used to obtain the measurements.	16
Figure 11 Experimental set up used for the study of sound propagation parameters of bone.	16
Figure 12 Simulation of the fracture bone; (a) normal healing (b) non-uniform healing	17
Figure 13 Bone clamping for axial (a) and lateral (b) exitaiton.	18
Figure 14 Block diagram of the instrument set-up to measure the vibration of an excised human tibia by the impulse response method.	18
Figure 15 The vibrational mode of a dry tibia.	19

Figure 16 The experimental set-up for the simulation of callus consolidation.	19
Figure 17 Block diagram of the natural frequency measurement system.	20
Figure 18 Plot of natural frequencies that increased steadily and reached a plateau during the healing.	21
Figure 19 Diagram of instrumentation used and natural frequency transducer mounted on implant in porcine mandible.	23
Figure 20 The transducer design and analysis.	23
Figure 21 Schematic diagram of data acquirement and treatment system of <i>in vitro</i> test.	24
Figure 22. Schematic diagram of data acquirement and treatment system of <i>in vivo</i> test.	25
Figure 23 The implant was placed into the tibia of the tested rabbits by a general procedure. (a) Before surgery, the testing rabbit was anesthetized. (b) The left tibia of the rabbit was scraped. (c) Implant was placed at the site near the proximal metaphysis. (d) An abutment screw was exposed for future natural frequency analysis.	26
Figure 24 Variation of natural frequencies (mean and SD) (a) the results of the <i>in vitro</i> tests (b) the results of the <i>in vivo</i> tests.	27
Figure 25 Implant stability with time for implants that eventually failed. The bold line indicates the mean stability for all successful implants.	28
Figure 26 Diagram describing the change of implant stability quotient values over time.	29
Figure 27 Photograph of vibration testing on total hip replacement.	30
Figure 28. Thompson <i>et al.</i> impedance mode to represent the human <i>in vivo</i> model.	35
Figure 29 Mode shape predicted by finite element model developed by Hobatho <i>et al.</i>	36
Figure 30 Discretization of the mandible and dental pins.	37
Figure 31 Sagittal views of the four configurations of the model. Elements with dark and light colour represent compact bone and trabecular bone, respectively	38
Figure 32. The THA model generated by Pro/Engineer. Left: assembled model; right: exploded view of the THA model. Dimensions are all in mm	40

Figure 33. Examples of hypothetical interface failures. (a) 1/3 distal interface failure and (b) finite element model of a femoral component of THA	40
Figure 34 The first six vibration modes of the control model	42
Figure 35 Vibration pattern in the case of cantilever linked to the implant; (a) in bucco-lingual direction (b) in mesio-distal direction.	43
Figure 36 The TXT file created by an edge detection software package.....	57
Figure 37 File converting software programmed by using Builder C++.....	58
Figure 38 Final ANSYS file for FE modelling.	58
Figure 39 (a) The original selection for the edge of the femur head. (b) The revised selection for the edge of the femur head after lowering the grey scale of the selected point.....	59
Figure 40 (a) The original selection including both the femur head and the pelvis (b) the revised selection for the femur head only by using the separation line.....	60
Figure 41 (a) The original selection of the femur (b) the revised selection of complete femur head by drawing some white lines in lower grey scale area. ..	60
Figure 42 (a) One of eight CT scan slices taken from the first 16cm of the distal part of the femur (b) Input points to ANSYS from CT scan.....	61
Figure 43 (a) One of two CT scan slices taken from the next 2cm of the femur (b) Input points to ANSYS from CT scan.....	61
Figure 44 (a) One of four CT scan slices taken from the following 2cm of the femur (b) Input points to ANSYS from CT scan.	62
Figure 45 (a) One of ten CT scan slices taken from the next 2cm of the femur (b) Input points to ANSYS from CT scan.....	62
Figure 46 (a) One of ten CT scan slices taken from the last 2cm of the femur (b) Input points to ANSYS from CT scan.....	63
Figure 47 The CT scan slices with unclear edge taken from the head part of the femur	63
Figure 48 The display of all the key points created by using CT scan slices. ..	64
Figure 49 Finished amputated distal part femur model in ANSYS.	64
Figure 50 Simplified model of implant and abutment in ANSYS.....	65
Figure 51 The hypothetical interface layer between femur and implant.	66
Figure 52 Complete model of osseointegration residual limb.	67
Figure 53 SOLID92 Geometry.....	69

Figure 54 (a) Meshed implant and abutment, (b) meshed interface, (c) meshed femur model with the implant and the abutment (d) The final meshed FE model for this study.....	70
Figure 55 The constrain of the femur-one key point on the top of the femur head was selected to fix in all directions.....	71
Figure 56 The demonstration of the permitted movement of a hip joint (ball-and-socket joint); three axes plus all directions in between.....	71
Figure 57 The constrain of the tight- the proximal part of the thigh was fixed with the area.	72
Figure 58 The first tenth vibration modes of the control model. (a) First bending mode in XZ plane, (b) first bending mode in YZ plane, (c) second bending mode in XZ plane, (d) second bending mode in YZ plane, (e) Stretch and compression, (f) third bending mode in XZ plane, (g) third bending mode in YZ plane (h) rotation (i) forth bending mode in XZ plane (j) forth bending mode in YZ plane.....	75
Figure 59 The spectrum of harmonic analysis; A, B and C are three obvious peaks between 1Hz to 3000Hz.....	76
Figure 60 Plot of NF differences between the different elastic modulus of the model and the classification of three vibration bands.....	77
Figure 61 The simulation for a 25% failure in the central part of the interface.....	78
Figure 62 The effect of the middle interface failure.	79
Figure 63 The simulation for a 50% failure in the proximal part of the interface	79
Figure 64 The effect of the proximal interface failure.....	80
Figure 65 The simulation of a 75% contact failure from the distal end.....	81
Figure 66 The effect of the distal interface failure.....	82
Figure 67 The plot of the NFs against different femur length.	83
Figure 68 The NF results of 3 rd vibration mode in Sensitive Band when the femur length changed from 14cm to 24cm.	83
Figure 69 The NF results of the first 10 vibration modes for the effect of different soft tissue condition.....	84
Figure 70 The NF of the third vibration mode with soft tissue with different young's modulus	84
Figure 71 The excitation method in this study.....	94

Figure 72 The vibration measurement system	94
Figure 73 The structure of the Shear Mode ICP accelerometer	96
Figure 74 The two different types of input channels.	99
Figure 75 The InstruNet Model100 A/D Box.	100
Figure 76. Network page with signal from all the channels.	102
Figure 77. Record page.	103
Figure 78 Record setup.	103
Figure 79 The control panel of the FFT analyser.....	104
Figure 80 (a) A schematic diagram of the NF measurement system, (b) <i>In vitro</i> experiment setup.	108
Figure 81. Mechanical properties test machine.....	110
Figure 82. Tensile test of the silicone rubber.....	110
Figure 83 The position of the patient in the <i>in vivo</i> experiments.....	112
Figure 84 The excitation method in the <i>in vivo</i> experiments	112
Figure 85 The weight bearing with a short training prosthesis.....	113
Figure 86 Weight bearing record for the patient A.	114
Figure 87 Weight bearing record for the patient B.	115
Figure 88 Weight bearing record for the patient C.	116
Figure 89 A typical vibration signal of a tested implant model obtained in 90th minute with SILASTIC T4 silicone rubber as interface.	117
Figure 90. The same vibration signal in frequency domain obtained in 90th minute with SILASTIC T4 silicone rubber.....	118
Figure 91 Three obvious peaks in the spectrum window of the signal obtained in 90 th minute with SILASTIC T4 silicone rubber as interface.	118
Figure 92 Three obvious peaks in the spectrum window of the signal obtained in 3000 th minute with SILASTIC T4 silicone rubber as interface.	119
Figure 93 No obvious peaks in the Over Sensitive Band in some measurements. The signal was obtained in 3000 th minute with SILASTIC T4 silicone rubber as interface.....	120
Figure 94 A spectrum of filtered signal obtained in 3000 th minute with SILASTIC T4 silicone rubber as interface.....	121
Figure 95 The NF values of different implant diameter and abutment length (Mean NF $\pm$ SD).....	122

Figure 96 The NF values of using SILASTIC T4, MOLDSIL MIDDLE, and MOLDSIL SOFT silicone rubber as interface during the curing time (Mean NF $\pm$ SD).....	123
Figure 97. Plot of NF values of three interface materials with different Young's Modulus at the time of 3000minutes (Mean NF $\pm$ SD)	123
Figure 98 A typical vibration signal spectrum taken from patient C at day 1 after the weight bearing.....	124
Figure 99 A typical vibration signal spectrum taken from patient C at day 45 before the weight bearing.....	124
Figure 100. A typical vibration signal spectrum from <i>in vivo</i> test. (a) Patient A (b) Patient B (c) Patient C.	125
Figure 101. Common features in the FFT pattern. (a) Patient A (b) Patient B (c) Patient C.	127
Figure 102. A vibration signal from tri-axial accelerometer taken from patient A on the first measurement.	127
Figure 103. The FFT of tri-axial accelerometer. (a) Horizontal direction, the same direction as the single axial accelerometer (b) Vertical direction (c) Axial direction.	128
Figure 104 A schematic diagram of how the excitation force produced vibrations in different directions.	129
Figure 105 The NF results from single accelerometer measurement of Patient A (Mean NF $\pm$ SD).....	130
Figure 106. The results of the tri-axial accelerometer of Patient A. (a) Horizontal direction (b) Vertical direction (c) Axial direction.	132
Figure 107 The <i>in vivo</i> NF results of the patient B (before weight bearing). .	133
Figure 108 The natural frequencies of the patient B (before and after weight bearing).	133
Figure 109 Pain scale data of patient B.....	134
Figure 110 The graph of pain scale data averaged between the day of last test and the day before new test.....	134
Figure 111 The <i>in vivo</i> NF measurement results of the patient C before weight bearing.....	136

Figure 112 A plot of the natural frequency results of patient C (before and after weight bearing).	136
Figure 113 A plot of the pain scale data of patients C	137
Figure 114 The graph of the pain scale data averaged each time between the last and present test.	137
Figure 115. The force analysis of an element in a vibrational beam.	202

List of Tables

Table 1 Detail dimensions for implant and abutment.	65
Table 2 Dimensions of the soft tissue in FE model.	66
Table 3 The fundamental frequencies of the control model.	73
Table 4 The NF results of different elastic modulus interface.....	77
Table 5 The NF of different contact area in the middle part of interface.	78
Table 6 The NF results of different contact area in the proximal part of interface.....	80
Table 7 The NF results of different contact area in the distal part of interface. The contact area is from full contact to 25% contact.....	81
Table 8 The NF results of different femur length.	82
Table 9 The NF results for the effect of different soft tissue condition.....	84
Table 10 Proposed specifications for the vibration measurement system.	93
Table 11 Specification of Model 352A21 accelerometer.	97
Table 12 Specification of Model 356A24 accelerometer.	97
Table 13 Specification of the signal conditioner.....	98
Table 14 The differences of single-ended and differential input channels.	99
Table 15 Specification of A/D box.	101
Table 16 Accelerometer calibration data.	105
Table 17 Validation of the data collection system results.....	106
Table 18 The <i>in vitro</i> experiment setup	109
Table 19 The clinical records of candidates.....	111
Table 20 NF of 17mm and 18mm implant. (Mean NF $\pm$ SD, n=12)	121
Table 21 NF of abutment with different interfaces. ST, SILASTIC T4. MM, MOLDSIL MIDDLE. MS, MOLDSIL SOFT. (Mean NF $\pm$ SD, n=12).....	122
Table 22 <i>In vivo</i> natural frequency results of patient A. (Mean NF $\pm$ SD, n=12)	129
Table 23 <i>In vivo</i> natural frequency results from patient B. (Mean NF $\pm$ SD, n=12)	132
Table 24 <i>In vivo</i> natural frequency results of patient C. (Mean NF $\pm$ SD, n=12)	135

Chapter 1 Introduction

1.1 Background

Osseointegrated trans-femoral implants present a way forward in dealing with certain above knee amputees for whom a conventional socket prosthesis has not been a successful treatment. In this two-stage surgical procedure, an implant is inserted into the femur (stage one) and then typically 6 months later an abutment is fitted to the implant (stage two). A conventional prosthesis may then be attached to the abutment, which protrudes through the soft tissue and skin. Full load bearing is achieved normally 6 months after the fitting of the abutment. The results from a small number of patients (near 100 world-wide) who have undergone this treatment are encouraging, however the 6-month fixed length rehabilitation programme seems not fit for all patients due to individually varying health conditions, medical histories and osseointegration achieved between the bone and the implant. Therefore quantitative assessment was sought to support the clinical decision in managing rehabilitation progress. One important element of this is to determine the degree of osseointegration, which influences the anchorage strength between the implant and the bone, as the degree of osseointegration could be used to make a more informed judgement about the post surgery rehabilitation progress.

To assess the bone-implant interface condition and the degree of osseointegration, a number of methods have been used. These include: histology and histomorphometry observations, removal torque analysis, pull and push through tests, and X-ray examination. However, due to problems of invasiveness and accuracy, these methods could not be used to assess the osseointegration interface in the current clinical trials. Mechanical vibration analysis, a traditional non-destructive testing method, has been successfully used by mechanical engineers to inspect structural integrity. In recent years, it has been used in clinical research, such as dental implants, stability studies of orthopaedic implants, the assessment of bone healing after fracture, and the diagnosis of total hip arthroplasty loosening.

Vibration analysis involves the mechanical excitation of an object followed by recording and analysing the subsequent response. The natural frequency (NF) is the key parameter in vibration analysis. It is related to certain important characteristics of the vibration object, such as its dimension, mass, boundary conditions, and structural integrity. The changes of NF were observed by other investigators while the osseointegration process took place.

1.2 Aim and objectives

The aim of this study was to use the NF of the implant system to determine the change of mechanical characteristics as a reflection of interface between the implant and the bone. To achieve this goal, the current study was divided into five objectives:

- Creating finite element (FE) model to analyse the dynamic behavior of the osseointegration implant system. The purpose of the FE modelling was to find out what factors influence the NF and carry out simulations which are unable to be duplicated in experimental studies.
- Designing and constructing a vibration measurement hardware system for *in vitro* and *in vivo* studies.
- Developing a Fast Fourier transform (FFT) signal processing software for the NF analysis of the signal measured from the abutment.
- Applying the designed vibration measurement system in the *in vitro* experiments. The purpose of the *in vitro* experiments was to assess the operation of the hardware system, to establish and evaluate protocol of the *in vivo* measurement and to simulate the changes of the interface using different silicone rubbers.

- Applying the vibration measurement system in the *in vivo* experiments. The results form the basis for further studies of the relationship between NF and osseointegration interface.

1.3 Layout of the thesis

After this introductory chapter, Chapter 2 provides the background of osseointegration and its clinical applications and reviews the development of the natural frequency measurement in clinical studies. It outlines the need for the current work and the possibility of using natural frequency analysis to assess the osseointegration interface.

Chapter 3 provides the fundamental theories supporting this research and a description of vibration, FFT and FEA theory is presented in the second part of this chapter.

Chapter 4 describes the 3D FE simulations of the osseointegration trans-femoral implant and the methodology to create the FE model is presented. This chapter primarily focuses on the factors that affect the natural frequency of the implant abutment system.

Chapter 5 describes the hardware and the software development for the vibration measurement and signal analysis. Further details regarding the excitation method and vibration measurement equipments are presented. The software for collecting and analyzing signal are explained. The *in vitro* and *in vivo* experimental procedures are also explained in this chapter.

Chapter 6 presents the results from the *in vitro* and the *in vivo* studies.

Discussions of the results are provided in Chapter 7 with limitations of this work also being presented. Chapter 8 presents the conclusions from this work whilst key areas of future work are identified in the final part of this chapter.

Chapter 2 Literature Review

2.1 Background

The loosening of implants from bone tissues has been a cause of problems in reconstructive surgery and joint replacement. It had been thought that the layer of fibrous tissue that develops around the implant diminishes the integrity and mechanical stability of the implant/bone interface (Southam 1971). During the 1950s it had been shown by Brånemark in experiments focused on observing the microcirculation of bone that chambers made of titanium could become permanently incorporated with bone. The discovery showed that the living bone could become so fused with the titanium oxide layer of the implant that the two could not be separated without fracture (Branemark 1983). Brånemark introduced a term 'osseointegration' to describe this modality for stable fixation of titanium to bone tissue. From this discovery, through its laboratory development and initial application in the clinical dentistry, osseointegration has become a realized phenomenon of importance also in the orthopedics and rehabilitation.

2.1.1 Definition of osseointegration

Osseointegration was defined as a direct structural and functional connection between ordered living bone and the surface of a load-carrying implant (Branemark 1977). In practice, this means that in osseointegration there is an anchorage mechanism whereby non-vital components can be reliably and predictably incorporated into living bone and that this anchorage can persist under all normal conditions of loading (Worthington 1997).

2.1.2 Experimental studies

The initial observations of osseointegration were made in the 1950s during the study of the circulation in bone marrow (Branemark 1959). In a modification of the rabbit ear chamber, a titanium implant with a central canal and a transverse opening at one level was threaded into bone to allow both bone and vessels to

grow into the chamber. The investigator felt that such integration of titanium screws and bone might be useful for supporting dental prostheses on a long-term basis.

The study of the biomechanics of osseointegration was a key early research activity, and was overseen by Skalak (1991). Detailed biomechanical tests were performed by Brånemark and coworkers to evaluate implants during healing, after irradiation, in experimental arthritis, in osteoarthritis and rheumatoid arthritis, and *in vivo* in rat, rabbit, dog, and man (Brånemark, 1996). This series of studies provided evidence that the biomechanics of bone-anchored implants are complex. In dental applications, the clinical experience is now sufficient in length of time and in total patient numbers to conclude that neither stress shielding nor fatigue appear to be limiting factors in the long-term successful function of titanium dental fixtures (Skalak, 1997).

2.1.3 Titanium properties

Titanium is believed to have been discovered by Gregor in Cornwell, UK, in 1790 (Nishimura, 1991). Titanium forms 0.6-0.7 percent of the earth crust. In the presence of oxygen, as in air, a highly corrosive resistant oxide is rapidly formed which is very resistant to acids as well as alkalis (Ferguson 1959). Various types of oxides such as TiO , Ti_2O_3 , and TiO_2 are formed when titanium is machined in open air (McLean 1994). The oxide layer adheres closely to the bulk metal and is also very dense which makes titanium one of the most corrosion-resistant metals in existence even when exposed to the chemical surroundings in the body. Another interesting property of TiO_2 is its high dielectric constant. This results in stronger van der Waals bonds for TiO_2 than with other oxides, and may contribute to the biocompatibility of titanium. (Kasemo, 1985) This thin oxide layer covers the surface of pure titanium after being spontaneously formed under atmospheric conditions. More extensive oxide growth occurs on titanium implants subjected to biological tissues (Sundgren, 1986). Inflammatory cells, especially macrophages, may contribute to development of the oxide layer by excreting proteolytic enzymes, cytokines, superoxide, and hydrogen peroxide (Bjurstén, 1991). It is hypothesized that the actual interface of the titanium implant to the living tissue is a hydrated

titanium peroxy matrix as shown in Fig. 1. The formation of such a matrix is unique to titanium. The other possible transition metals either have too low solubility of their peroxy complex or too low stability of the complex.

Figure 1. An artist's view of the titanium/tissue interface. (Bjursten, 1991)

Figure 2. Representation of events at the bone-implant interface. (a) Protein adsorption from blood and tissue fluids, (b) protein desorption, (c) surface changes and material release, (d) inflammatory and connective tissue cells approach the implant, (e) possible targeted release of matrix proteins and selected adsorption of proteins such as BSP and OPN, (f) formation of *lamina limitans* and adhesion of osteogenic cells, (g) bone deposition on both the exposed bone and implant surfaces, (h) remodeling of newly formed bone. (Puleo, 1999)

The performance of biomaterials comprises two components, the response of the host to the implant and the behavior of the material in the host (Black 1992) as shown in Fig. 2.

2.1.3.1 Material response

The almost immediate event that occurs upon implantation of metals, as with other biomaterials, is adsorption of proteins (Andrade 1985). These proteins first come from blood and tissue fluids at the wound site and later from cellular activity in the periprosthetic region. Once on the surface, proteins can desorb (native or denatured, intact or fragmented) or remain to mediate tissue implant interactions. In fact, the nature of this 'conditioning film' deposited on biomaterials is believed to be responsible for the host response.

2.1.3.2 Host response

The host response to implants placed in bone involves a series of cell and matrix events, ideally culminating in intimate apposition of bone to biomaterial, for example, osseointegration. For this intimate contact to occur, gaps between bone and implant must be filled, and bone damaged during preparation of the implant site must be repaired. During this time, hostile conditions, such as, premature loading leading to micromotion, will disrupt the newly forming tissue, leading to formation of a fibrous capsule (Brunski, 1991). Morphological studies have revealed the heterogeneity of the bone-implant interface. One feature often reported is an afibrillar interfacial zone. Although its thickness and appearance vary, this zone forms regardless of the type of biomaterial implanted, including cpTi, stainless steel, and hydroxyapatite as shown in Fig. 3 as reviewed in Nanci et al. (1998)). It has been postulated that this interfacial zone provides a mechanism for 'bonding' between natural hard tissue and biomaterial.

Figure 3. Electron micrographs illustrating the presence of a thin, electron-dense cement line (CL) at the interface between titanium implants and bone in (A) a rat tibial model and (B) an implant retrieved from a human. IS, implant space; Oc, Osteocyte. (Nanci, 1998)

Osteoblasts, osteoid, and mineralized matrix are observed adjacent to the *lamina limitans*-like layer (Albrektsson 1987), suggesting bone is deposited directly on the surface of the implant, extending outward from the biomaterial. Thus, bone formation in the periprosthetic region occurs in two directions; not only does the healing bone approach the biomaterial, but bone also extends from the implant toward the healing bone. Interestingly, fluorochrome labeling indicates that the bone extending away from the implant forms at a rate about 30% faster than that moving toward the biomaterial as shown in Fig. 4.

Figure 4. Bar graph comparing the rate of bone deposition in a direction away from and towards the implant. (mean and SD) (Puleo, 1999)

Understandably, because of the complexities of the *in vivo* environment, the bone-implant interface is not fully characterized. The heterogeneity and patchy immunolabeling observed in morphological studies suggest that, even though several biomolecules have been identified at the interface, there are likely to be more than the ones presented. These other biomolecules may have essential roles in directing the bone response to the implant, and further work is needed to identify them and determine their functions at the interface (Puleo, 1999).

2.1.4 Clinical applications of osseointegration

2.1.4.1 Dental

In the field of oral surgery, the most common application of osseointegration has been the dental and oral reconstruction of patients who have lost their teeth. The anatomical and functional rehabilitation requires not only replacement of the teeth but also part of the surrounding tissues because the loss of teeth results in involution of periodontal tissues. Osseointegration has been used for the replacement of missing single teeth, for the restoration of the partially edentulous segment of the mouth, and for the reconstruction of the completely edentulous patient by means of implant-supported fixed bridges or removable overdentures that attach to an implant-supported framework (Worthington, 1991).

Since 1965, there are more than 800,000 patients worldwide who had been treated with osseointegration dental reconstructions (Brånemark, 2001). It apparently indicates the superior performance in long term success rate of osseointegration in dental applications.

2.1.4.2 Facial prostheses

Complex problems of facial tissue loss are often amenable to management by means of implant-supported maxillofacial prostheses as shown in Fig. 5. Many previously irradiated patients have been treated successfully with implant

reconstructions. One such patient had a hemimandibulectomy performed as part of her ablation and the surgery was followed by radiation. After reconstruction of the mandible with a bone graft and prophylactic hyperbaric oxygen, implants were used to stabilize a full fixed lower partial denture (Brånemark, 1983).

Figure 5. Major maxillofacial defect after tumor resection and radiation therapy.
(Tjellstrom, 1992)

2.1.4.3 Finger joint prostheses

Patients suffering from rheumatoid arthritis, osteoarthritis, or posttraumatic/postinfectious arthrosis often present with considerable impairment in hand function. Traditional solutions employed silicon rubber implants (Swanson 1972). Although initial results usually are satisfactory, the friction of silicone moving against bone may eventually result in progressive bone destruction with subsidence and fracture of the implant as well as progressive stiffness of the hand (Hagert, 1975).

The artificial joint consists of two components, the joint mechanism and the titanium fixtures. The screw-shaped titanium fixtures are essentially of the same design as the anchorage elements used for edentulous patients. The artificial joint mechanism consists of a constrained flexible silicone spacer, each end being mounted on a titanium plate with a short stem, designed to fit into a central longitudinal cylindrical channel in each fixture. As shown in Fig. 6, a 25-year-old female with juvenile rheumatoid arthritis and severe subluxation of the third to fifth metacarpophalangeal joints with ulnar drift of

the fingers (left). Radiograph two years following arthroplasty. (Lundborg, 1997)

Figure 6. Example of osseointegration arthroplasty. (Lundborg, 1997)

2.1.5 Rehabilitation of the trans-femoral amputee with an osseointegrated prosthesis

Osseointegration has more recently been extended to orthopaedic applications (Branemark, 1994). Of particular interest is the treatment of trans-femoral amputees where traditional rehabilitation using socket prostheses causes complications related to prosthesis retention and function, as shown in Fig. 7.

Figure 7. Trans-femoral amputee with osseointegrated limb prosthesis. (Brånemark, 1999)

As shown in Fig. 8, in the case of a conventional socket prosthesis, the load is transferred via skin and subcutaneous tissue, while in the case of an osseointegrated prosthesis the load is transferred directly to the skeleton. The first trans-femoral osseointegration procedure took place in Sweden in 1990. In 1997, this procedure was extended to selected patients in the U.K at Queen Mary's Hospital, Roehampton, in collaboration with the team in Sweden. Currently, osseointegration is only considered for trans-femoral amputees who have been unable to achieve a satisfactory level of rehabilitation using conventional socket techniques (Brånemark, 2001). Reasons for this include: recurrent skin infections and ulceration in the socket contact area, a short stump, volume fluctuation of the stump, soft tissue scarring and extensive area of skin grafting or socket retention problems due to excessive perspiration (Sullivan, 2003).

Figure 8 Schematic drawing of socket prosthesis (left) and osseointegrated prosthesis .
The arrows show the reaction force support the body weight. (Cunningham, 1990)

Osseointegration in the above knee amputation involves a two-stage surgical procedure. In stage 1, a titanium fixture, basically designed as a screw as shown in Fig. 9a, is inserted using a particular technique and special instruments to optimise osseointegration. The soft tissue is then closed so

covering the fixture which is then left unloaded for a period of 4 to 8 months. The patient may use a conventional prosthesis when the soft tissue has healed. In stage 2, the soft tissue is trimmed and an abutment penetrating the skin is attached to the fixture, as shown in Fig.9b. Approximately six weeks following stage 2 surgery, osseointegration trans-femoral implant patients wear a short training prosthesis to complete an exercise program. The short training prosthesis allows the patients to weight bear through the osseointegrated structure in standing only. Scales are used to monitor the loading. Bending moment is also applied to the short training prosthesis throughout rehabilitation progress. Gradually increased loading of the system is then permitted under the supervision of a physiotherapist. The final prosthesis is usually fitted after a further four months, with the full rehabilitation around 12-18 months following the stage 1 procedures.

Figure 9 Operation for stage 1 and 2; (a) implant insertion, (b) abutment installation

It is through this period of progression that patients begin to experience some of the physical and prosthetic benefits of osseointegration.

Physical benefits including:

1. A full, unrestricted range of movement around the hip joint, unimpeded by a socket brim;
2. Osseoperception i.e. improved sensory feedback as a result of the direct link with the skeletal system (Brånemark, 2001; Ysander, 2001);
3. Reduced soft tissue problem;

Prosthetic benefits include:

1. Improved suspension of the prosthesis, with no functional lengthening during swing phase and direct transmission of movements;
2. No fitting problems due to fluctuation of stump volume;
3. No skin problems from the stump/socket interface;
4. Ease of donning and doffing (Bergkvist).

As a result of these benefits, candidates experience functional improvements in their general activities of daily living. Examples include sitting comfortably on a low chair, fully flexing the hip to tie shoelaces and cycling.

The quality of life for osseointegrated trans-femoral amputees increases markedly, but some patients feel pain or discomfort during the rehabilitation program (Sullivan, 2003). A medical review is then required, which may be followed by a decrease or cessation of weight bearing, and consequently further work is required to support the clinical decision making process. One element of this is determining the degree of osseointegration achieved between the implant and the femur, as this could be used to provide a more informed judgement on post surgery rehabilitation time.

2.2 Applications of natural frequency measurement in clinical studies

To assess the interface of the implant and the bone, various methods, histology and histomorphometry observations (Albrektsson, 1987; Sennerby, 1992a; Sennerby, 1992b; Ericsson, 1994), removal torque analysis (Carlsson, 1988; Johansson, 1991; Wennerberg, 1995), pull and push through tests (Dhert, 1992; Shirazi-Adl, 1992), and X-ray examination (Meredith, 1998), have been proposed. However, it is difficult to avoid the problems of invasiveness and accuracy when these methods apply.

Mechanical vibration analysis, a traditional non-destructive testing tool, has been widely used by mechanical engineers to inspect structural integrity. In recent years, it has been used in biomedical research, such as in bone dentistry,

the assessment of bone healing after fracture, and the diagnosis of total hip arthroplasty loosening.

2.2.1 Assessment of bone healing after fracture

It was found that clinical methods of assessment of fracture union were subjective and liable to observer bias (Nokes, 1985). Moreover, the radiological appearance relies on the mineralization of callus, which itself lags behind the mechanical events of union by up to six weeks. As a result, the radiological assessment is not helpful until late in the course of healing (Nokes, 1985). Therefore, it is necessarily to have a method of assessing fracture union, which is able to record at a specific time, the actual state of the callus in term of its mechanical properties (Nokes, 1985).

Vibration measurements were early used in the assessment of bone healing after fracture. Lippmann (1932) applied oscillatory percussion to the femur, humerus and clavicle using his finger. The responses were recorded via a stethoscope. He concluded that the pitch and quality changes resulted from free vibration of the separate fragments and, accordingly, signified a completely healed fracture or an incomplete union. Also, Dencker and Moberg (1967) described a technique for analyzing femoral and humeral fracture healing by vibrating the bone in 1967. Even though, this method was not refined enough and limited the study to fractures with soft-tissue interposition or cases where the two major fragments were not in contact. In addition, Lewis and Tarr (1975) used vibration measurement technique to monitor fracture healing in humans.

In 1984 Nokes (1984) used two accelerometers to record the vibration signal from the fracture tibias as shown in Fig. 10. He found that middle third fractures of the human tibia manifested specific changes in the natural frequency of the fracture fragments as the fracture united. The difference in frequency between the proximal and distal fragments diminished with time,

and the natural frequency of each tends was close to a common value. The differences could be used to construct a healing curve against to measure the healing of other fractures.

Figure 10 The apparatus used to obtain the measurements. (Nokes, 1985)

In the experiments carried out by Singh (1989), twelve animal fresh and dry tibias were used for the NF measurements. The vibrations were excited by an electro-mechanical vibrator. Sensor pick-up output was fed to a CRO (cathode-ray oscilloscope) as shown in Fig. 11. It was found the reduction in natural frequency and sound intensity, and the increase in the output voltage for fractured bone in comparison to a normal bone.

Figure 11 Experimental set up used for the study of sound propagation parameters of bone. (Singh, 1989)

Nikiforidis et al., (1990) used two accelerometers to record the lateral and axial vibration signal of a fracture tibiae model. The fresh human tibiae were obtained from amputations. An osteotomy in the middle third of the tibiae was then created. To simulate a healing fracture, the osteotomy gap was filled with silicone rubber whose mechanical properties are similar to those of the soft callus of healing fracture as shown in Fig 12. To simulate the non-uniform formation of new bone, the material representing new bone was applied in a non-uniform shape. The materials chosen to simulate the fracture callus in different stages of healing were based on available information regarding the mechanical characteristics of the callus during healing (Davy; Yamada; Yamagishi, 1955; Ham, 1972; Perren, 1979).

Nikiforidis (1990) investigated the use of vibrational analysis of fracture healing of *in vitro* and *in vivo* bone using an electromagnetic shaker as an excitation source with a slowly sweeping frequency ranging from 200 to 1600 Hz as shown in Fig. 13. The phase meter, in order to investigate the phase shifts between normal and fractured bone and to monitor the phase, shifted as healing progressed.

Figure 12 Simulation of the fracture bone; (a) normal healing (b) non-uniform healing (Nikiforidis, 1990)

Figure 13 Bone clamping for axial (a) and lateral (b) excitation. (Nikiforidis, 1990)

In two papers written by Nakatsuchi et al. (1996) vibrational modes were examined on excised human tibias both *in vitro* and *in vivo*. For the *in vitro* test, two separate modes were found, one vibrating in the lateral direction and the other more weaker one in the anterior posterior direction as shown in Fig. 14.

Figure 14 Block diagram of the instrument set-up to measure the vibration of an excised human tibia by the impulse response method. (Nakatsuchi, 1996)

Figure 15 The vibrational mode of a dry tibia. (Nakatsuchi, 1996)

By identifying the primary vibrational mode in the lateral direction they speculated that this would be a more viable clinical tool to investigate bone fracture healing as shown in Fig 15. Also in the test they performed a vibration experiment, which simulated callus consolidation in fracture healing with the wet tibias mounted on a strip of foam rubber and external fixation as shown in Fig. 16.

Figure 16 The experimental set-up for the simulation of callus consolidation. (Nakatsuchi, 1996)

The results showed that the impulse response method can be useful for the quantitative assessment of fracture healing even if the various materials for fracture fixation such as a plate, Ender's intramedullary nail or an external fixator were utilized (Nakatsuchi, 1996).

For their *in vivo* test, natural frequency measurements were applied on 65 tibia fracture patients as shown in Fig.17 (Nakatsuchi, 1996).

Figure 17 Block diagram of the natural frequency measurement system. (Nakatsuchi, 1996)

The results showed that the changes in natural frequency demonstrated various patterns during the overall course of fracture healing. Excluding five tibias in which the natural frequency was difficult to interpret, 55 tibias (88.7%) showed an increase in the natural frequency with healing as shown in Fig.18.

Figure 18 Plot of natural frequencies that increased steadily and reached a plateau during the healing. (Nakatsuchi, 1996)

These previous works have demonstrated that the vibration technique enables a comprehensive picture of fracture healing to be obtained since the frequencies of several modes of vibration are determined. The vibration measurement as an assessment of fracture healing has several advantages. Firstly it is a non-invasive examination; the magnitude of the impact has little effect on the stability of the fracture. Secondly, the examination can be performed repeatedly in a short time frame. Thirdly the instrumentation is not expensive and finally a quantitative evaluation of fracture healing is possible, especially in the early stage of callus consolidation (Nakatsuchi, 1996). However, the vibration measurement in the assessment of bone healing after fracture has following two problems. The first problem is the lack of an adequate theoretical model for the behaviour of bone *in vivo* (Jurist, 1970). The second problem is the difficulties in obtaining the reproducible data in living subjects because of the interposition of soft tissue between the accelerometers and the bone (Pugh, 1975). Although, it is desirable to have a method of assessment that should allow the early and accurate prediction of failure of progress to union, so that the treatment designed to augment healing may be instituted without the delay which is inevitable with current clinical techniques of fracture healing assessment (Nokes, 1985).

2.2.2 Assessment of dental implant stability

The development of effective and durable dental implants has become increasingly important in oral rehabilitation along with the rapid increase in the number of dental implants installed. However, implant treatments still fail frequently. Some researchers have attempted to use vibrational measurements to determine whether sufficient osseointegration has occurred between the dental implant and the bone.

In 1973, Noyes and Solt excited central incisors to vibrate by applying a sinusoidal force to the teeth. They found that the vibrational characteristics of a tooth are related to the mass and the nature of the surrounding tissue, although, the detailed relationship between NF and boundary conditions were not discussed in their report (Noyes, 1973).

In previous research in oral rehabilitation by Kaneko, the sensitivity of a vibrational test for *in vivo* assessing the interfacial rigidity of a biocompatible dental root implant was studied using models. The use of vibration theory has proved to be a feasible method for assessing the stability of dental implants (Kaneko, 1987; Kaneko, 1989; Kaneko, 1991).

In a series of publications, Meredith and coworkers (Meredith, 1994; Meredith, 1996) reported on the use of a transducer that could be directly attached to an implant body or to the abutment on the implant as shown in Fig. 19. The study was using resonance frequency (RF) analysis to make quantitative measurements of the stability of the implant tissue interface *in vitro* and *in vivo*.

Figure 19 Diagram of instrumentation used and natural frequency transducer mounted on implant in porcine mandible. (Meredith, 1998)

Resonance frequency analysis offers a clinical, noninvasive measurement of stability of the osseointegration of implants. Initial *in vitro* studies demonstrated the ability of the device to assess changes in interfacial stiffness (Meredith, 1998). Clinically, RF values have been correlated with changes of implant stability during osseointegration healing, failure of implants to integrate and the supracrestal dimensions of the implant (Meredith, 1997). The results of a recent histomorphometric study suggested that RF values correlated well with levels of bone-implant contact (Rasmusson, 1999). These findings support the use of RF analysis in assessing changes in the bone healing and osseointegration process following implant placement.

Figure 20 The transducer design and analysis.

In their study, a beam-like transducer for clinical trials was analyzed as shown in Fig. 20 (Cawley, 1998). They drew the conclusion that the sensitivity of the transducer natural frequency to the changes of interest is dependent on the thickness and length of the beam element. The research results have been used to finalize a transducer design for the clinical trials. The fundamental cantilever theory was used for their vibration analysis.

The transducer behaved like a cantilever beam, which is free to vibrate but the other end was attached to the implant. The natural frequency of such a beam can be calculated as;

$$R_f = \frac{1}{2\pi} \sqrt{\frac{3EI}{l^3 m}} \quad (1)$$

where: R_f = natural frequency (Hz), l = effective length of beam, m = mass of beam (g), E = Young's modulus (GN m^{-2}), I = moment of inertia.

Recently Huang et al., have improved the natural frequency measurement by using FFT software. In their studies, the vibration behaviors of dental implants were analyzed using a non-contact model testing method both *in vitro* as shown in Fig. 21 and *in vivo* as shown in Fig. 22.

Figure 21 Schematic diagram of data acquisition and treatment system of *in vitro* test. (Huang, 2000)

Figure 22. Schematic diagram of data acquisition and treatment system of *in vivo* test. (Huang, 2001 Oct.)

From the *in vitro* results (Huang, 2000; Huang, 2003.), the natural frequencies of the implants were concentrated from 8 to 19 kHz under different boundary conditions. The natural frequency values decreased when the attachment levels and boundary force were reduced. Linear relationships were found between response frequencies and the degree of implant stability.

In their *in vivo* study (Huang, 2001), 698 anterior teeth were examined by a conventional probing method and also by natural frequency analyzer. The results showed that the natural frequencies of teeth with healthy condition were significant higher than those of teeth with periodontal disease. Therefore this can be used as an effective method for the early quantifiable testing and prevention of periodontal disease.

In 2003, a complete osseointegration progress was monitored by Huang et al. (2003) both *in vitro* and *in vivo*. To simulate the bone healing process, stones with a density of 1.90 g/cm³ were used to fill the inter-space between the implant and the bone block. Before filling, the stone was mixed in a vacuum mixer with a water/stone ratio of 3/10 for 45 s. The RF values of the samples were measured immediately after the implants were installed in the bone grafts and every 3 min until 75 min after the placement. The results showed that the NF increased with the bone healing process. To validate such an *in vitro* test,

animal experiments, as shown in Fig. 23 were performed (Huang 2002). Implants were placed in the left tibias of six rabbits using a general surgery procedure and the NF values of the implants were measured. The NF values of the implants increased significantly ($P<0.05$) during the healing period and reached a plateau when the implant-bone interface was united. The variations of NF values of the testing implants showed a similar trend to the results of *in vitro* tests, as shown in Fig. 24, The implants with higher initial NF values had shorter simulated healing times. Based on these findings, they concluded that NF analysis is a reliable and accurate method for early assessment of the osseointegration process. Although this technique has been shown to be a useful tool for implant stability examination, many factors, including bone density and bone type, which could affect NF of an implant, were not investigated and their effects remain unclear.

Figure 23 The implant was placed into the tibia of the tested rabbits by a general procedure. (a) Before surgery, the testing rabbit was anesthetized. (b) The left tibia of the rabbit was scraped. (c) Implant was placed at the site near the proximal metaphysis. (d) An abutment screw was exposed for future natural frequency analysis. (Huang, 2003)

Figure 24 Variation of natural frequencies (mean and SD) (a) the results of the *in vitro* tests (b) the results of the *in vivo* tests. (Huang, 2003)

In 2004, Glauser et al. (2004) analysed the development of implant stability by repeated NF analysis measurements during 1 year in 23 patients treated according to an immediate/early-loading protocol. The objective was to evaluate the possible differences between failing and successful implants. Nine implants failed (11.2%) during the 1 year of loading, as shown in Fig 25. The NF analysis showed a distinctly different pattern between the implants that remained stable and the ones which were lost. The implants that failed during the course of the study showed a significantly lower stability already after 1 month. Within the limitations of this study, it was concluded that failing

implants show a continuous decrease of stability until failure. Low NF levels after 1 and 2 months seem to indicate an increased risk of future failure.

Figure 25 Implant stability with time for implants that eventually failed. The bold line indicates the mean stability for all successful implants. (Glauser, 2004)

In 2006, Crismani et al. (2006) investigated the behaviour of early loaded palatal implants when observed with NF analysis. The NF value was automatically translated into an implant stability quotient (ISQ) value by an Osstell device. The behaviour of early loaded palatal implants showed an initial decrease of the ISQ values. From 6 weeks post-surgery onwards the ISQ values increased, as shown in Fig 26. Within the limitations of their study, the results suggest that the healing time of palatal implants reported in the literature (12 weeks) should be discussed. An orthodontic loading of palatal implants 6 weeks post-surgery with a force of up to 0.4 N seems to be justified. During the whole observation time, the mean ISQ value dropped significantly after 2 and 3 weeks post-surgery when compared with the one of first loading. A possible reason for this phenomenon is a decrease of the mechanical stability of the implants and the surrounding hard tissue relaxation. They suggested that further investigations were necessary to evaluate the behaviour of early loaded palatal implants considering observation periods of over 12 weeks.

Figure 26 Diagram describing the change of implant stability quotient values over time. (Crismani, 2006)

In these studies, it was widely found that the NF increased in healthy dental implants and decreased in unstable or failure implants. According to this, the use of vibration theory was proved to be a feasible method for assessing the stability of dental implant. It is also concluded that the NF measurement can be used as an effective method for the early quantifiable testing and the NF analysis is a reliable and accurate method for early assessment of the osseointegration. Furthermore, the results indicate the possibility that loading on the implant might cause the decrease of the NF.

Even though, it is proved that the NF measurement and analysis were the useful methods for examining the implant stability. It still remains unclear how the factors, such as the bone density and bone type, could affect the NF of the implant. Nevertheless, it is obviously that this technique greatly contributed to assess the stability of dental implants. Therefore, in osseointegration trans-femoral implants, it brings the possibility to apply this technique to assess the implant stability and the interface between the implant and the bone.

2.2.3 Diagnosis the loosening of the hip prosthesis

An application of vibration measurement in an *in vitro* setting showed that models at early and late loosening stages could be reliably detected (Li, 1995;

Li, 1996). There was, however, poor sensitivity for early loosening unless there was obvious prosthetic instability. The first documentation using this technique on total hip arthroplasty patients is by Georgiou and Cunningham (2001). They compared vibration test results with radiographs of 23 patients with total hip replacements experiencing hip pain without X-ray evidence of loosening. As shown in Fig. 27, the vibrator was applied to the knee, and held at 90° to the femur in both planes. The accelerometer was applied to the subcutaneous greater trochanter. Their results showed a sensitivity of 80% and a specificity of 89%. The positive predictive value was 92% and the negative predictive value was 73%; it was unable to produce a definitive diagnosis in 8% of patients. When compared with radiographs from the same patients, vibration testing was shown to be 20% more sensitive and able to diagnose 13% more patients. In another words, it can be concluded in this study that the vibration testing method was more sensitive than the traditional X-ray.

Figure 27 Photograph of vibration testing on total hip replacement. (Georgiou, 2001)

Recently Qi et al., (2003) investigated how much information a vibrational technique can reveal regarding the loosening of the femoral component of a total hip arthroplasty (THA) by numerical model using the finite element method. A numerical model analysis was undertaken to obtain the natural frequencies and mode shapes of femoral models under free vibration. An absent interface portion was hypothesized at specified locations. The data from

modal analysis were then used in a computer simulation to visualize the vibration diagnosis.

In their FE simulation, the effects of interface failure on the femoral component stiffness vary with the locations and sizes of failure. Nominal critical, reliable, and non-detectable levels of interface failure were found. When the size of failure is more than one third of the stem length, this tool is able to provide a reliable diagnosis. The critical detectable interface size is one fifth of the stem length. It was found that the motion at higher harmonics is the most sensitive to the interface failure. The tool is unlikely to differentiate the slight differences in natural frequencies that occur when the interface failure length is less than one sixth of the stem length. It is likely to fail in detecting failures located at the stem central portion regardless of their length.

In this area, it was confirmed that the vibration measurement could reliably detect the late loosening with an unstable prosthesis. Furthermore, it was concluded that the vibration testing method was more sensitive than the traditional X-ray through the testing on the 23 patients. Recently, using finite element method, it was found that the vibration measurement was able to provide a reliable diagnosis when the size of the interface failure was more than one third of the stem. Therefore, it is worthwhile to apply the vibration measurement on the osseointegration trans-femoral implant.

2.3 Methods of Vibration Measurement

Generally speaking, vibration analysis involves two basic steps, which are the mechanical excitation of an object and the recording and analysis of the subsequent response (Noke, 1999). Theoretically, it is feasible to monitor various pathological and trauma-induced conditions by applying a stimulus to the bone and analyzing the response. With the rapid development in the technology both in soft and hardware, it brings great possibility to develop a useful clinical diagnostic tool based on vibrational analysis techniques.

2.3.1 Excitation Methods

In practice, there were two basic types of stimuli which the human skeletal system can be excited in order to deduce some of its vibrational characteristics (Noke, 1999). The first type was applying impulsive stimuli to excite the system and the response yielded with consistently inputting a number of independent fundamental frequencies. It may be analyzed with the aid of the fast Fourier transform.

Thomsen, in 1990, applied an input impulse of approximately 0.5 ms to sufficiently excite a tibia in the frequency range of up to 450 Hz. A simple beam model was developed to represent the tibia under vibratory conditions (Thomsen, 1990). In addition, Nakatsuchi et al. (1996) used an elaborate feedback mechanism to incorporate an electrically mounted impulse hammer, which was triggered to give a short impulse to the tibia. More recently, Huang et al. (2003) used an impulse hammer in their *in vitro* and *in vivo* test on dental implants.

The second type was frequency input, which consisted of variable frequency cycles of pure sine waves (Noke, 1999). Most investigation applied the direct frequency inputs by using a mechanical vibrator. Nikiforidis et al. (1990) used an electromagnetic shaker as an excitation source to investigate the use of vibrational analysis of fracture healing of *in vitro* and *in vivo* bone. Additionally, Li et al. (1996) evaluated the implant loosening by applying a force of a sinusoidal profile to the distal end of an *in vitro* femur containing an implant prosthesis. Moreover, Meredith et al. (1996) used a frequency response analyzer, which was programmed on a personal computer.

In summary, it has been shown that the method of impulse was usually chosen to start a vibration-based research. The impulsive force was often generated either by using a ball of a known mass dropped down a tube of a fixed height or a hammer, which had attached to it a force transducer. On the other hand, for frequency input, the use of a mechanical vibrator was widely applied. The vibrator generates a range of frequencies on the object, which the natural

frequency range must be known. With the help of a FFT analyzer, some recent studies were carried out using impulse excitation because the excitation is easy to apply and the NF of the targeted object is not affected by the excitation impulse.

2.3.2 Vibration measurement

There were several methods, which were used to record vibration signals in previous studies. Firstly, stethoscope was early used by Lippmann in 1932 to monitor the healing of fracture long bones. The stethoscope was placed proximally and then distally to the fracture site on bony prominences to record the characteristic osteal sound.

Secondly, with the advent of increased technology, piezoelectric transducers were introduced to monitor the skeletal vibrations. There are many researches used accelerometers to examine the bone fractures. Indeed, piezoelectric sensors were employed widely as the most versatile means of converting vibrations into electrical energy. Noke pointed out (1999) that successful operation of a piezoelectric gauge depended upon the coupling of the gauge to the structure under investigation. Meredith and coworkers (1996) used piezoelectric beam-like transducers to assess the dental osseointegration interface. Also, Georgiou and Cunningham (2001) used a vibrator to excite the femur with total hip replacement, and collected the vibration signal by using an accelerometer. Loosening of the hip were successfully diagnosed.

Thirdly, strain gauges were used by Pelker and Saha (1983) to monitor the stress wave propagation in bones. The gauges were bonded near the proximal end of *in vitro* bone to reduce the interference from the wave reflected from the distal end. It is found that the strain gauge was not accurate enough when the NF of the object was higher than 1000Hz.

More recently, microphone was used in the dental implant researches. Huang and his coworkers (Huang, 1999) used an acoustic sensor to reduce the contact

effects on the measured natural frequency value of tested implant. In 2003, Huang and his coworkers used a microphone to measure the NF of dental implant *in vivo*. However, it is difficult to entirely extinguish the vibration of the dental implant from the sound picked up by the microphone.

Among these methods, the piezoelectric transducer was more widely used than others due to its accuracy and reliability, as well as the inexpensive cost. Accelerometers using piezoelectric technique are more accepted in recent NF studies because they were easy to attach and remove on the vibration object.

2.4 Models for vibration analysis in clinical studies

The early vibration modeling started off in 1970's. The mechanical or electrical equivalents were often involved such as dash-pots and springs, and resistors and capacitors. With the support from the experimental results, the vibratory characteristics of bone can be represented by the modeling system.

In 1976, Thompson *et al.* (1976) investigated the vibratory properties of the human ulna *in vivo* by mechanical impedance measurements. The modeling system used springs and masses, as shown in Fig. 28. The constant values used in the model were obtained by analyzing the experimental data and the mathematical models were improved by the results.

Figure 28. Thompson *et al.* (Thomsen 1990) impedance mode to represent the human *in vivo* model

Later, Hight *et al.* (1980) modeled the *in vivo* tibia based on Timoshenko uniform beam theory. They proposed that natural frequencies were heavily influenced by the boundary conditions, curvature and twists of the bone.

Peters (1982) investigated the vibrational characteristics of the human tibia by undertaking Fourier analysis. An impulse was applied to the medial tibial condyle and a response signal picked up at the medial tibial malleolus by accelerometers. The frequency components of the two signals were then determined using fast Fourier analysis. The accelerometer transfer function $H(E)$ was used as an index of fracture healing and is defined as

$$H(E) = \frac{Y(F)}{S(F)} \quad (2)$$

where $Y(F)$ and $S(F)$ are the frequency responses of the medial tibial condyle and the medial tibial malleolus respectively. He compared the transfer function of the healthy contralateral bone with the healing bone. It was observed that as healing occurred the transfer function approached that of the healthy limb.

Thomsen (1990) used excised and embalmed human tibia to simulate free-free support conditions. A model was developed using the Timoshenko beam theory. Derivation of the model based on the geometry of the tibia resulted in a series of constants including Young's modulus and shear modulus of cortical and cancellous bone. Densities of cortical and cancellous bone and bone marrow were also required for modeling. The approach to the human tibia was simplified by making several assumptions, such as the neglect of the stiffness of cancellous bone, the mass of bone marrow, the twist of the tibia along its principal planes, shear deformations and rotational inertia. The results, after a series of experimental investigations, proved that the simplified approach is adequate to model the vibratory response of an *in vitro* bone.

Figure 29 Mode shape predicted by finite element model developed by Hobatho *et al.* (1991)

In 1991, Hobatho *et al.* (1991) reported the development of a three-dimensional finite element model of the human tibia using experimental modal analysis, as shown in Fig. 29. The experiment involved using a dry bone and a simplistic set-up consisting of an instrumented hammer and an accelerometer. This was used in order to obtain the modal shapes associated with the vibrating bone.

The result of the experiment was used to optimize the FE model. The percent relative error between the two methods was approximately 3%.

In a paper by Roberts *et al.* (1996), a technique described as mechanical response tissue analysis (MRTA) offered a method of separating the effects of soft tissue on bone. A theoretical analysis of the wave propagation through the compressed tissue and bone revealed a strong mass effect dependent on the relative accelerations of the vibration and bone. The experiment involved the use of monkey tibias, which were tested *in vivo* to obtain vibration response. The monkeys were subsequently sacrificed and the tibias excised and tested to failure in three point bending. The *in vitro* and *in vivo* results were compared in order to generate a model that could be clinically relevant to assess long bone mechanical properties.

Figure 30 Discretization of the mandible and dental pins. (Williams, 1997)

In 1997, Williams (1997) created a mandible model with dental pins as shown in Fig. 30. The conditions of a dental pin and a dental pin with attached cantilever integrated in the mandible have been examined. The dynamic analysis by impulse excitation of an osseointegrated dental pin indicated a

sensitivity of the implant response to impulse duration and direction. The analysis provided response data from the implant for a range of impulse strengths. Williams concluded that the data might be used with clinically measured experimental data for an interpretation of implant osseointegration, rate of bone healing, the onset of infection and pin failure.

Figure 31 Sagittal views of the four configurations of the model. Elements with dark and light colour represent compact bone and trabecular bone, respectively (Lee, 2000).

Lee *et al.* (2000) reported that finite element analysis could serve as a tool to aid in the accuracy of natural frequency analysis in periodontal problems whilst Huang *et al.* (2001) established a 3D FE model of a dental implant with healing abutment. It was used to determine the effects of different interfaces on the natural frequency of a dental implant. The finite element analysis package,

ANSYS (Rev. 5.4, Swanson Analysis Systems Inc., Houston, PA, USA), was used on a personal computer for pre-processing and modal analysis. The fixture body was 3.75 mm in diameter and 10 mm in length. The healing abutment had a length of 3 mm. As shown in Fig 34, the implant was embedded into a bone block (10 mm×10 mm in cross-section and 15 mm in length with a compact bone of 2 mm thickness on the top side) with 0.8 mm exposed height. As shown in Fig 32a, the term "exposed height" was defined as the distance from the upper bone surface to the cervical portion of the implant. According to the classification of Lekholm and Zarb (1985), four types of bone quality were simulated. In the type I model, the mechanical properties of the entire bone block were set as compact bone, as shown in Fig. 32a. In the type II model, a core of dense trabecular bone was covered by a thick layer of compact bone with a width of 1.2-1.6 mm, as shown in Fig. 32b. The geometric configuration of the type III model was similar to the type II model, but the width of the compact bone was reduced to 0.6-0.8 mm as shown in Fig. 32c. The type IV model had the same configuration as the type III model, however the density of trabecular bone was decreased by one half, as shown in Fig. 32d. The thickness of compact bone set in the FE models were according to the previous study of Holmes and Loftus (1997). To validate the reliability of the FE model, the resonance frequencies of actual dental implants were measured by a series of *in vitro* experiments.

After the reliability of the FE model was verified, two situations were simulated to assess the effects of surrounding bone. The first situation simulated different bone types, and evaluated their effect on the resonance frequency of the dental implant. The second situation simulated continuous bone loss after implant treatment. In the first situation, the resonance frequencies and vibration mode shapes of the model were also computed with four different bone types. In the second situation, to determine the effect of bone density on resonance frequency, the entire bone (both compact bone and trabecular bone) density in the type III model was altered from 100% to 10% in increments of 10%.

The results of the study (Huang, 2001) showed that the natural frequency of an implant was affected by its marginal bone characteristics including type, density and level. Thus, the natural frequency could be useful as a parameter in future development of devices for assessing surrounding problems. It indicated that numerical simulation could offer reliable results in the analysis of the vibrational characteristics of a dental implant. The methodology used in the study provided a useful reference for future study in the natural frequency analysis.

Figure 32. The THA model generated by Pro/Engineer. Left: assembled model; right: exploded view of the THA model. Dimensions are all in mm (Qi 2003).

Figure 33. Examples of hypothetical interface failures. (a) 1/3 distal interface failure and (b) finite element model of a femoral component of THA (Qi 2003).

As mentioned before, Qi et al. (2003) created a finite element model for total hip arthroplasty (THA). The computer model was constructed in a standard manner for biomechanical situations, which consists of four components: bone, cement mantle, interface, and implant stem with a hypothetical interface layer between cement and stem. The CT scans of human femoral cadaver were used and converted into digital cross-sectional profiles such that these profiles were edited by a computer aided design (CAD) package, Pro/Engineer. Using these cross-sections, the shape of a femur was created by the advanced feature generation methods of the CAD system. An uncollared straight stem was selected and modeled and the cement mantle was generated so it fits the profiles of the stem and the bone. A nominal interface layer of 0.2 mm thickness was generated between the stem and cement mantle. The four components were then assembled automatically. The interfaces shared the same dimensions. This choice assures zero clearance and no interference because quantity was allowed in the final assembly of the THA CAD model. The THA model, its assembly relationship, and key dimensions are given in Fig. 32. This model was then ready for the finite element analysis. Fig. 33 shows examples of the models with hypothetical stem cement interface failures and their locations. Fig 34 shows the deformation of the first six vibration mode. The study showed that a vibrational diagnostic tool was able to provide a reliable diagnosis, however, the tool was unlikely to differentiate the slight differences in NF that occurred when the interface failure was less than one-sixth of the stem length.

Figure 34 The first six vibration modes of the control model (Qi 2003).

In 2006, Natali et al. (2006) used the finite element method for assessing the ability of frequency analysis in detecting the degree of oral implant osseointegration. A 3D model, as shown in Fig 35, of a mandible was obtained from tomographic survey. Magnitude displacements are shown on the amplified deformed shapes in the case of a maximum interfacial bone elastic modulus of 15 GPa. The base frame is used to show the un-deformed configuration and a single implant was considered in canine region. A good correlation between frequency response and osseointegration level was obtained. This was carried out by providing an impulse excitation of the implant that resulted in a vibration pattern. Within the limit of finite element analysis, the outcomes showed that numerical investigation provided

understanding the behaviour of testing devices based on frequency measurements. Also, it confirmed that the potential of the vibrations technique as non-invasive analysis for osseointegration process.

Figure 35 Vibration pattern in the case of cantilever linked to the implant; (a) in bucco-lingual direction (b) in mesio-distal direction. (Natali, 2006).

According to the reviews above, finite element analysis has been proved to be a useful tool to investigate the effect of varying geometries of implant and the area of the bone-implant interface on the NF. Although, it was questioned that FE analysis might not be sufficiently accurate due to complex non-linear properties of both skin and bone. With the advent of increased technology, today, most of the FE work can be facilitated by readily available commercial FE computer software (Lee, 1992). The accuracy of the solution is largely dependent upon the number of elements of the model, a higher number of elements potentially enabling a more accurate solution to be found (Zhang, 1998). Moreover, other factors affecting the accuracy of the analysis involve model assumptions and parametric inputs of the geometry, material properties, boundary conditions. Successfully, FE analyses make possible the simulation of clinical situations involving dental implants with different percentages and patterns of osseointegration. In this study, the development of the FE model is going to be carried out very carefully and validated experimentally.

2.5 Comparison of *in vitro*, *in vivo* and FEA results

In the research related to dental application, there were several investigators, who have compared the results between FE modelling, *in vitro* and *in vivo* studies. In 2003, Huang tested dental implants on both *in vitro* and *in vivo* (Huang 2003). For the 3.75mm implant, the NFs of the *in vitro* study were between 9630Hz and 10140Hz. The NFs of the *in vivo* study were between 4590Hz and 6960Hz. In 2006, Natali (2006) created a FE model of a 4mm dental implant to carry out natural frequency analysis. The NF values were between 12900Hz and 13350Hz under different bone conditions.

In the research related to total hip replacement, the closest results between FEA modelling and an *in vivo* study were reported by Qi (2003). In his study, FEA modelling results were compared with other clinical studies (Li, 1995; Li, 1996; Georgiou, 2001). In his work, the fundamental frequency of the control model was 115.75Hz. The clinical data, which were collected by Georgiou and Cunningham, measured it at 102.3Hz. Qi expected the difference because of factors such as femur length, material properties, type of the stem, and contiguous joints. In comparison with *in vitro* data which were measured by Li (1995 and 1996), the difference was insignificant. Qi concludes that the models of his study can be considered as a fair representation of the experimental models of Li's study. There were no higher frequency harmonic peaks in Li's study and therefore, the comparison for the NF in high frequency range was kept unknown.

2.6 Summary

During the last three decades, osseointegration has been used increasingly in clinical applications. More recently, Osseointegration was extended in trans-femoral implants and become a new orthopaedic anchoring method to attach prosthetic limbs. The results from the small number of patients who have undergone this treatment are encouraging. However, the fixed 6-month rehabilitation time scale has caused frustration to patients. On the other hand, this time scale may not be sufficient for the patient with poor bone quality. Therefore, quantitative data is required to support the clinical decision concerning the rehabilitation process.

Due to the problems of invasiveness and accuracy, the traditional methods including X-ray examination, are not entirely suitable to assess the osseointegration interface. The vibration analysis was used as an alternative clinical tool to diagnosis bone fractures in early stage. Since the mid-1990s, there was a significant increase in the research on the vibration analysis of the osseointegration interface in dental and total hip replacement area.

A considerable number of investigations have been carried out and proved that the vibration measurement could be used to reliably detect the degree of osseointegration. Evidently, the vibration analysis has been already widely used to assess the osseointegration interface in the dental area. For osseointegration trans-femoral implant, some amputees feel pain or discomfort during the rehabilitation program. Therefore, it is necessary to assess the osseointegration interface between the implant and the femur, as this could be used to provide a more informed judgement on post surgery rehabilitation time.

In the previous researches, the NF was used as a key parameter, which was related to the mechanical properties of the osseointegration interface. Normally, two excitation methods were used to obtain the NF. One is the frequency input method, which required installing a transducer to the implant to apply programmed frequency vibration. When the frequency of the

excitation vibration matched the NF of the implant system, the implant system had the highest amplitude. Then, the NF could be measured by the highest response amplitude. The other is the impulse method which could be applied by an impact force. The NF could be obtained from the free vibrational signal by using an FFT analyzer. The impulse method has been used more often in some most recent studies because the excitation was easy to apply and the NF of the implant system was not affected by the excitation method. There were several sensors to collect vibration signal. Among these sensors, accelerometers using piezoelectric technique which was more accepted because of their accuracy and reliability. In several reports, the osseointegration process was entirely monitored and the increasing NFs were found. In other reports, the failures of the osseointegration were found when the NFs were decreased.

Due to rapid technology improvement both in soft and hardware, finite element modeling is able to not only simulate complex geometric shapes and material properties, but also simulate various clinical situations. As a result, it has gradually become an important tool to assess the osseointegration interface in NF analysis. In previous studies, the mechanical properties of the soft tissue had not been simulated in any FE models. Understandably, the soft tissue surrounding a dental implant is only a thin layer, which might not cause noticeable effect. However, there is necessity to simulate the soft tissue in this FE study since the surrounding soft tissue is thick layer comparing with the size of the femur and could affect the NF of osseointegration trans-femoral implant system significantly.

This study was to use the NF to determine the change of the interface of the osseointegration trans-femoral implant. The FE modelling was used to find out the factors, including soft tissue, that influence the natural frequency of the implant system and to simulate different clinical situations. The impulse excitation was applied on the abutment by a pendulum whilst the vibration was recorded by accelerometers both in *in vivo* and *in vitro* experiments. The NFs were obtained by using an FFT analyzer.

Chapter 3 Fundamental theories and computation method of supporting this research

Chapter 2 illustrated that vibration analysis had been used widely in clinical application of osseointegration with three fundamental theories and method being confirmed to support this research. This chapter presents the details of these methods: vibration theory, fast Fourier transform and finite element modelling. In vibration theory, the equation for osseointegration trans-femoral implant system is evaluated. In fast Fourier transform, the theory is explained, and in FE modelling, both the theory and the application are described in detail.

3.1 Vibration theory

From a mechanical standpoint, an implant system is a simple cantilever. The natural frequencies of the abutment are computed as (See Appendix 8)

$$F = \frac{\omega}{2\pi} = \frac{\beta^2}{2\pi} \sqrt{\frac{EI}{\rho A}} = \frac{(\beta l)^2}{2\pi} \sqrt{\frac{EI}{\rho A l^4}} = \frac{\alpha^2}{2\pi} \sqrt{\frac{EI}{\rho A l^4}} \quad (1)$$

Where F = Natural Frequency (Hz); l =effective length of beam; A =area of cross section; E = Young's modulus; I = moment of inertia; ρ = the vibration mess per unit 1, and α is a constant related to the boundary conditions.

Meredith (1996) was the first researcher to introduce cantilever vibration theory in dentistry vibrational study. He assumed that the implant behaved as a cantilever beam, which is free to vibrate but is supported at the end attached to the bone. The natural frequency of such beam was calculated as:

$$R_f = \frac{1}{2\pi} \sqrt{\frac{3EI}{l^3 m}} \quad (2)$$

where: R_f = natural frequency (Hz), l =effective length of beam, m = mass of beam (g), E = Young's modulus (GN m^{-2}), I = moment of inertia.

Since
$$m = \rho Al \quad (3)$$

The natural frequency given by Meredith could be changed as:

$$R_f = \frac{\sqrt{3}}{2\pi} \sqrt{\frac{EI}{\rho Al^4}} \quad (4)$$

Huang and his coworker evaluated the vibration equation as

$$F = \alpha \sqrt{\frac{EI}{\rho l^4}} \quad (\text{Huang 2003}) \quad (5)$$

where: F = natural frequency (Hz), l =effective length of beam, E = Young's modulus (GN m^{-2}), I = moment of inertia, ρ =mass per unit effective vibrational length, α = constant related to boundary conditions. In Equation (5) A , which is the area of the cross section, was treated as a constant.

Therefore, Equation (1) which was evaluated was consistent with the equations tested by Meredith and Huang. It was anticipated that the natural frequency of the system would vary with the length of vibrational object, moment inertial and the surrounding bone tissue.

3.2 Fast Fourier transform

A vibration signal can be viewed from two different standpoints: the time domain and the frequency domain. Time domain is like the trace on an oscilloscope where the vertical deflection is the signal amplitude, and the horizontal deflection is the time variable.

The second representation is the frequency domain. This is like the trace on a spectrum analyzer, where the horizontal deflection is the frequency variable and the vertical deflection is the signal amplitude at that frequency. The FFT provides more accurate and reliable natural frequency analysis result to the study.

3.2.1 Fourier series

According to the theory developed by Fourier, any periodic function $F(t)$, with period T , may be represented by an infinite series of the form

$$F(t) = \frac{a_0}{2} + \sum_{n=1}^{\infty} (a_n \cos n\omega_T t + b_n \sin n\omega_T t) \quad (6)$$

where the coefficients a_0 , a_n , and b_n for a given periodic function $F(t)$ are calculated by the formulas

$$\omega_T = \frac{2\pi}{T} \quad (7)$$

$$a_0 = \frac{2}{T} \int_0^T F(t) dt \quad (8)$$

$$a_n = \frac{2}{T} \int_0^T F(t) \cos n\omega_T t dt \quad n=1,2,\dots \quad (9)$$

$$b_n = \frac{2}{T} \int_0^T F(t) \sin n\omega_T t dt \quad n=1,2,\dots \quad (10)$$

It should be noted that the first coefficient a_0 is twice the average of the function $F(t)$ over one period. This series is called the Fourier series and the coefficients are called the Fourier coefficients.

A line graph of the amplitudes of the Fourier series components can be drawn as a function of frequency. Such a graph is called a spectrum or frequency spectrum. The periodic time T defines completely which spectral components occur in the spectrum as $f_0 = 1 / T$. The component f_0 is called the fundamental frequency or first harmonic. All the higher components are multiples of f_0 and are called the higher harmonics. The smallest spacing that can occur between frequency components in the spectrum is $f_0 = 1 / T$.

3.2.2 Discrete Fourier Transform and the FFT

The Fourier transform provides the means of transforming a signal defined in the time domain into one defined in the frequency domain. When a function is evaluated by numerical procedures, it is always necessary to sample it in some fashion. This means that in order to fully evaluate a Fourier transform with digital operations, it is necessary that the time and frequency functions be sampled in some form or another. Thus the digital or Discrete Fourier Transform (DFT) is of primary interest.

3.2.2.1 The Discrete Fourier Transform

This is used in the case where both the time and the frequency variables are discrete (which they are if digital computers are being used to perform the analysis). Let $x(nT)$ represent the discrete time signal, and let $X(mF)$ represent the discrete frequency transform function. The DFT is given by

$$X(mF) = \sum_n x(nT) e^{-inm2\pi FT} \quad (11)$$

where

$$x(nT) = \frac{1}{N} \sum_m X(mF) e^{inm2\pi FT} \quad (12)$$

3.2.2.2 The Fast Fourier Transform

The FFT is simply a class of special algorithms, which implement the discrete Fourier transform with considerable saving in computational time. The FFT is not a different transform from the DFT, but rather just a means of computing the DFT with a considerable reduction in the number of calculations required. While it is possible to develop FFT algorithms that work with any number of points, maximum efficiency of computation is obtained by constraining the number of time points to be an integer power of two, e.g. 1024 or 2048.

3.3 Finite element method

Finite element analysis (FEA) is a powerful technique originally developed for numerical solution of complex problems in structural mechanics, and it remains the method of choice for complex systems. In FEA, the structural system is modeled by a set of appropriate finite elements interconnected at points called nodes. Elements may have physical properties such as thickness, density, Young's modulus and Poisson's ratio.

The FEA is an approximate solution method whereby a complex structure is discretized or divided into a number of regularly shaped pieces, known as elements. Each element is defined by several nodes or points, the coordinates of which establish the geometry of the structure. The behaviour of the complex structure is then approximated as the sum of the respective responses of each of the regularly shaped elements.

There are three phases in FE software--ANSYS: pre-processing, which defines the finite element model and environmental factors to be applied to it, analysis solver (solution of finite element model) and post-processing of results (using visualization tools).

3.3.1 Pre-processing

Pre-processing in using FEA is constructing a finite element model of the structure to be analysed. The input of a topological description of the structure's geometric features is required in most FEA packages. This can be in either 1D, 2D or 3D form, modeled by lines, shapes or surfaces representation respectively, although nowadays 3D models are predominantly used. The primary objective of the model is to realistically replicate the important parameters and features of the real model. The simplest mechanism to achieve modeling similarity in structural analysis is to utilise pre-existing digital blueprints, design files, CAD models and/or data by importing that into a FE

analysis environment. Once the finite element geometric model has been created, a meshing procedure is used to define and break the model up into small elements. In general a finite element model is defined by a mesh network, which is made up of the geometric arrangement of elements and nodes. Nodes represent points at which features such as displacements are calculated. FE packages use node numbers to serve as an identification tool in viewing solutions in structures such as deflections. Elements are bounded by sets of nodes, and define localised mass and stiffness properties of the model. Elements are also defined by mesh numbers which allow reference to be made to corresponding deflections or stresses at specific model locations.

3.3.2 Analysis

The next stage of the FEA process is analysis. The FEM conducts a series of computational procedures involving applied forces, and the properties of the elements which produces a model solution. The modal analysis used in this study is to determine the NFs of the osseointegration implant system during free vibration.

3.3.3 Visualisation

These results can then be studied using visualisation tools within FEA environment to view and to fully identify the implications of the analysis. Numerical and Graphical tools allow the precise location of data such as stresses and deflections to be identified.

FE investigations may include structural (for example, dynamic analysis, stress analysis, deformation analysis, fatigue analysis, crack propagation), heat transfer, fluid flow and electro-magnetic analysis. According to the literature review in Chapter 2, FE modeling not only can simulate complex geometric shapes and material properties, but also can simulate various clinical conditions, which are difficult to replicate in experiments. As a result, it has

gradually become an important tool to assess the degree of osseointegration. (Papavasiliou, 1997; Qi, 2003 June; Natali, 2006) In this study, dynamic analysis of osseointegrated residual limb has been carried out to find out what factors influence the natural frequency measurement whilst modal and harmonic analysis were used to find the NF of the model. Steps taken in generating and investigating the model are generation of the geometry of the model, assigning material properties to the model, deciding on elements to be used, loading of the model, solution method, and investigation of the outcome of the modelling.

3.3.4 Modal analysis using FEM

The goal of modal analysis is to determine the natural mode shapes and frequencies of an object or structure during free vibration. It is common to use the FEM to perform this analysis because the object being analyzed can have arbitrary shape and the results of the calculations are acceptable. The types of equations which arise from modal analysis are those seen in eigensystems. The physical interpretation of the eigenvalues and eigenvectors which come from solving the system are that they represent the frequencies and corresponding mode shapes. Usually, the only desired modes are the smallest because they are commonly the most prominent modes at which the object will vibrate dominating all the higher modes.

For the most basic problem involving a linear elastic material which obeys Hooke's Law, the matrix equations take the form of a dynamic three dimensional spring mass system. The generalized equation of motion is given as:

$$[M][\ddot{U}] + [C][\dot{U}] + [K][U] = [F] \quad (13)$$

where $[M]$ is the mass matrix, $[\ddot{U}]$ is the 2nd time derivative of the displacement (i.e. the acceleration), $[\dot{U}]$ is the velocity, $[C]$ is a damping matrix, $[K]$ is the stiffness matrix, and $[F]$ is the force vector. The only

terms kept are the 1st and 3rd terms on the left hand side which give the following system:

$$[M][\ddot{U}] + [K][U] = [0] \quad (14)$$

This is the general form of the eigensystem encountered in structural engineering using the FEM.

In this study, because of the complexity of the osseointegration structure, multiple-degree-of-freedom and different mode shape are used in to the simulations

Chapter 4 FE modelling of osseointegration implant system

As previously reviewed, FE modelling has increasingly become a very useful tool to analyze vibration characteristics of osseointegration implant and simulate different clinical conditions. This chapter firstly explains how the FE model of osseointegration trans-femoral implant system was generated. Secondly, four factors are simulated, such as elastic modulus of the implant interface, contact areas of the interface, lengths of the femur and soft tissue. Thirdly, the results of the harmonic and modal analysis are presented. Fourthly, the results are discussed in order to examine to what extent factors could influence the natural frequency of the implant system. Finally, the suggestion was made for the further *in vivo* and *in vitro* experiments in the part of summary.

4.1 Finite element model generation

As it was mentioned in literature review, finite element method (FEM) is a numerical modelling method. Therefore, it is important to take the validity of FEM results into account. Comprehensively, it is mainly dependent upon the following three types of modelling factor, which were involved in Finite element model generation. The first factor is the similarity of the finite element model to the real structure of the object to be analyzed, the second is the effectiveness of modelling the boundary conditions, whilst the third is associated with the modelling of the material properties of the structure studied.

In this study, the osseointegrated residual limb model was constructed in a standard manner for biomedical situations in term of mechanical characteristics. It consists of five components: amputated distal femur, implant, abutment, soft tissue, and a hypothetical interface between bone and implant.

4.1.1 Real geometry amputated distal femur model generation

Clinical Computer Tomography (CT) scans of osseointegration patient A and an edge-detection programme make reconstruction of the real-geometry model a feasible process. In this study, two software packages were used for the reconstruction. One is edge-detection software, which was developed by Dr J Gao (2006), who worked on a previous osseointegration project. The software can detect the edge of the bone from CT scan slices and write the position of the points into a TXT file as shown in Fig 37. The function of the other software, as shown in Fig 38, is to select a number of these points from the bone edge and to convert the data as an ANSYS command file as shown in Fig 39. The code is presented in **Appendix 6**

Figure 36 The TXT file created by an edge detection software package.

Figure 37 File converting software programmed by using Builder C++.

Figure 38 Final ANSYS file for FE modelling.

4.1.2 Edge detection

Compared to other radiography examination methods, the CT scan is the best method for investigating the condition of the skeletal system with an excellent grey scale. The edge detection software could automatically trace a closed loop, which was reconstructed with same grey scale points. Fig 40a shows that edge of the femur was selected when clicked on in the area of the femur head. Due to the complexity of bones and soft tissue, it is very difficult to select a line just near the edge of the bone especially on the CT scan slice near the femur head. To overcome this problem, the software was developed to allow users to identify the edge of the bone.

After selecting the point in the CT scan slice, the selected grey scale can be increased or decreased. Fig 40b shows how decreasing the grey scale of the selected point can help to select a better edge for the femur head.

Figure 39 (a) The original selection for the edge of the femur head. (b) The revised selection for the edge of the femur head after lowering the grey scale of the selected point.

Some areas of the femur head can be very close to the pelvis. It is impossible to only select the femur without selecting the pelvis as shown in Fig 41a. The “Separation” command can separate the femur head and the pelvis by drawing a black line in the gap as shown in Fig 41b.

Figure 40 (a) The original selection including both the femur head and the pelvis (b) the revised selection for the femur head only by using the separation line

In some parts of the femur, especially near the femoral head, a lower grey scale was found due to lower density of the bone, as shown in Fig 42a. This problem was solved by using “Connection” command, which could draw white lines to the area where it was needed to be connected with the femur head as shown in Fig 42b.

Figure 41 (a) The original selection of the femur (b) the revised selection of complete femur head by drawing some white lines in lower grey scale area.

4.1.3 3D modelling

According to the CT scan record of the patient, the length of the amputated distal femur was 22.4cm. Because the shape of the femur is not uniform, a different number of CT scan slices were chosen for 3D modelling

reconstruction on each occasion. As shown in Fig 43, reconstructed by him, the lower part of the femur is more uniform. Therefore, CT scan slices were chosen to create the ANSYS 3D model. 8 CT scan slices were chosen to create first 14cm of the femur model.

Figure 42 (a) One of eight CT scan slices taken from the first 16cm of the distal part of the femur (b) Input points to ANSYS from CT scan.

Because the femur structure was more complex near the femur head, closer CT scan slice was needed to have an accurate dimension for FE analysis. Therefore in the next part of the femur model, a CT scan slice was chosen every 1cm to create the 3D model. 2 CT scan slices were chosen to create next 2cm of the femur model. One of the CT scan slice is shown in Fig 44a. The key points created in ANSYS were shown in Fig 44b.

Figure 43 (a) One of two CT scan slices taken from the next 2cm of the femur (b) Input points to ANSYS from CT scan.

Then, every 0.5cm one CT scan slice was chosen to be converted into the 3D model. 4 CT scan slices were chosen to create next 2 cm of the femur model.

Fig 45 shows one of the CT scan slice for this part of the femur model construction.

Figure 44 (a) One of four CT scan slices taken from the following 2cm of the femur (b) Input points to ANSYS from CT scan.

Then, every 0.2cm one CT scan slice was chosen to produce the ANSYS 3D file. 10 CT scan slices were chosen to create next 2cm of the femur model. One example is shown in Fig 46.

Figure 45 (a) One of ten CT scan slices taken from the next 2cm of the femur (b) Input points to ANSYS from CT scan.

Finally, 10 CT scan slices were chosen to create the head of the femur. The last part is 2cm. Fig 47a shows one CT scan slice for this part of femur model reconstruction, and Fig 47b shows the key points, which were created in ANSYS.

Figure 46 (a) One of ten CT scan slices taken from the last 2cm of the femur (b) Input points to ANSYS from CT scan.

As shown in Fig 48, the femur head and the pelvis were closely near to each other and no gap can be seen to separate them and as a result it is difficult to detect this part of the femur. Therefore, there was a 0.4cm part of the femur that could not be converted to ANSYS by using CT slices. Instead, it had to be entered manually in ANSYS.

Figure 47 The CT scan slices with unclear edge taken from the head part of the femur

At last, all the key points converted from the CT scan slices are displayed in Fig 49. The key points were then connected into lines manually and the 3D

femur model was reconstructed with these lines. The final amputated distal part femur model is shown in Fig 50.

Figure 48 The display of all the key points created by using CT scan slices.

Figure 49 Finished amputated distal part femur model in ANSYS.

4.1.4 Titanium implant and abutment model generation

The implant model for this study was generated according to the real implant in the hospital. The abutment dimensions were measured to create an abutment model. The “Glue” command in ANSYS was used to connect the implant and

the abutment together as one rigid part as shown in Fig 51. The detail dimensions of the implant and the abutment model are shown in Table 1.

Figure 50 Simplified model of implant and abutment in ANSYS.

Table 1 Detail dimensions for implant and abutment.

Implant	Length		80mm
	Pitch size		2mm
	Diameter		18mm
Abutment	Cylinder	Diameter	9mm
		Length	40mm
	Block	Width	14mm
		Height	14mm
		Depth	20mm
	Hexagon	Diameter	12mm
		Length	6mm

To insert the implant into the femur model, “Subtract” command was used. In addition, a hypothetical interface layer between femur and implant was created as shown in Fig 52. The thickness of the layer was 0.2mm. This layer separated along the femur model into eight equal parts. In interface failure analysis, a selected number of the interface layer can be removed to simulate partial

contact situation. The mechanical properties of the layer can also be changed to find out how the interface layer affects natural frequency of the implant system.

Figure 51 The hypothetical interface layer between femur and implant.

4.1.5 Soft tissue model generation

In the CT scan, the soft tissue was largely deformed and some parts of the soft tissue can not be seen in the CT scan image. Therefore in this study, the soft tissue was created according to the mean diameter of the thigh. The detail dimension of the soft tissue can be found in Table 2. The completed FE model is shown in Fig 53.

Table 2 Dimensions of the soft tissue in FE model.

Part	Dimension (mm)
Diameter of the thigh (proximal part)	160
Diameter of the thigh (distal part)	120
Length of the thigh (exclude distal disc)	200
Diameter of the distal disc	120
Thickness of the distal disc	5

Figure 52 Complete model of osseointegration residual limb.

4.1.6 Material mechanical property and element type

The material properties of all of the structures in the FE model have to be assigned to the respective elements. The source of the mechanical properties was from the literature, materials testing, and *in vivo* or *in vitro* studies.

Implant and abutment

Both implant and abutment were made of commercially pure (CP) titanium. The elastic modulus of CP titanium is 115GPa, and Poisson's ratio is 0.3. (Xu 2000) The density was 4400kg/m^3 . (Qi, 2003)

Bone (for femur and interface layer)

Previous reports indicate that the elastic modulus of bone is different in the radial and longitudinal direction and that these values vary along the femur. However, there is only a small variation within the femoral diaphysis. In this FE model, the first 16cm of femur was treated as cortical bone and the other part of the femur model was treated as cancellous bone. The cortical bone in this study was assumed to be transversely isotropic with a constant longitudinal elastic modulus of 18GPa and a shear modulus of 7GPa in all directions. (VanBuskirk, 1981) The radial and circumferential moduli are the same, 13GPa. (Cowin, 1997) The elastic modulus of the bone were tested in wet condition. The Poisson's ratio of the cortical bone was taken as 0.3. The density was 1800kg/m^3 . (Qi, 2003) The elastic modulus of cancellous bone

varies with its density. In this study the cancellous bone was assumed to be transversely isotropic with a constant longitudinal elastic modulus of 0.8GPa and a shear modulus of 0.1GPa in all directions. The radial and circumferential moduli are the same, 0.3GPa. The Poisson's ratio of the cancellous bone was taken as 0.33. (Ashman 1987) The density was 930kg/m³. (Qi, 2003) All elastic moduli of the bone were tested in wet condition.

Soft tissue

The soft tissue (muscle, skin, fat, ligaments, and tendons) of the residual limb is neither homogeneous nor isotropic. However, incorporation of homogeneous and isotropic material property assumptions greatly reduces the complexity of the model. Consequently, although many investigators in this field have speculated that precise representation of the tissue properties is critical to the accuracy. (Steege, 1988Silver-Thorn, 1992) The soft tissue has typically been assumed to be an elastic, homogeneous material. In this study soft tissue was assumed to be elastic and homogeneous with a constant elastic modulus of 60kPa. (Silver-Thorn, 1992) The Poisson's ratio of the soft tissue was taken as 0.49. The density was 919kg/m³. (Mendez, 1960)

4.1.7 Choosing element

FE methods require careful attention during the modeling of the physical problem and the creation of the FE mesh. In order to select the appropriate element to be used in the finite element model, several requirements should be met. Geometrical requirements necessitate that three-dimensional solid elements need to be used. Material requirements need that the element chosen should allow for orthotropic material properties to be assigned. Vibration analysis functions are needed.

According to above requirements, the element selected from the ANSYS element library is **Solid 92** as shown in Fig 55. **SOLID92** has a quadratic displacement behavior and is well suited for modelling irregular meshes.

The element is defined by ten nodes having three degrees of freedom at each node: translations in the nodal x, y, and z directions. The element also has plasticity, creep, swelling, stress stiffening, large deflection, and large strain capabilities. The element is suitable for the modal and harmonic analysis in this study.

Figure 53 SOLID92 Geometry

The whole model including the implant, the abutment as shown in Fig 56, the interface layer as shown in Fig 57, the femur as shown in Fig 58 and soft tissue as shown in Fig 59 was meshed by using automatic meshing generation. As shown in Fig.59, the mesh sizes of the elements varied with the dimensions of the edges in the model. For instance, the fine size element was meshed near the threaded interface with sharp edges. Meanwhile, the soft tissue was meshed with coarse size elements in order to reduce the cost of the computing time. The effect of the different mesh sizes of the FE model was not compared in this study. The total number of the elements in the model was 184,358.

Figure 54 (a) Meshed implant and abutment, (b) meshed interface, (c) meshed femur model with the implant and the abutment (d) The final meshed FE model for this study.

4.1.8 Constraint

The boundary conditions applied to the model were aimed to reproduce the mechanical constraints seen in a clinical situation. The femur head was fixed at one node in all directions as seen in Fig 56. Therefore, the femur model allowed rotation about a fixed point as the clinical situation as shown in Fig 57. The cross-section area of the proximal end of the thigh was fixed in all directions as seen on Fig 58.

Figure 55 The constrain of the femur-one key point on the top of the femur head was selected to fix in all directions.

Figure 56 The demonstration of the permitted movement of a hip joint (ball-and-socket joint); three axes plus all directions in between

Figure 57 The constrain of the thigh- the proximal part of the thigh was fixed with the area.

In the clinical test, the patient was supine. The patient's residual stump was rested on the bed to avoid muscular tension and to reduce possible effect of surrounding soft tissue on the NF of the implant system. Therefore, half of the residual stump model was fixed in all directions in the FE analysis.

4.2 Results

The data of the first 10 natural frequencies were computed and held as criteria for comparison. Commonly, high harmonics were associated with low energy, faster attenuation, and often difficult to detect. In this study, the fundamental frequencies of control model are shown in Table 3.

Table 3 The fundamental frequencies of the control model.

	1 st mode	2 nd mode	3 rd mode	4 th mode	5 th mode	6 th mode	7 th mode	8 th mode	9 th mode	10 th mode
NF(Hz)	238	240	1136	1145	2409	2454	2775	3078	4839	5142

The osseointegration trans-femoral model is not a single degree of freedom system so multiple vibration modes were presented; lateral, axial and torsion modes. For clarity, the first 10 vibration modes of the control model were exaggerated and are shown in Fig. 59. Fig 59 (a) and (b) shows the vibration of the model in the 1st and the 2nd vibration modes, which both corresponded to the first bending mode in XZ and YZ planes. Fig 59 (c) and (d) shows the vibration of the model in the 3rd and the 4th vibration modes, which both corresponded to the second bending mode in XZ and YZ planes. Fig 59 (e) demonstrates the vibration of the model in 5th vibration mode, which corresponded to stretch and compression mode. Fig 59 (f) and (g) shows the vibration of the model in the 6th and the 7th vibration modes, which both corresponded to the third bending mode in XZ and YZ planes. Fig 59 (h) demonstrates the vibration of the model in 8th vibration mode, which corresponded to rotation mode. Fig 59 (i) and (k) shows the vibration of the model in the 9th and the 10th vibration modes, which both corresponded to the forth bending mode in XZ and YZ planes.

(a)

(b)

(c)

(d)

(e)

(f)

(g)

(h)

Figure 58 The first tenth vibration modes of the control model. (a) First bending mode in XZ plane, (b) first bending mode in YZ plane, (c) second bending mode in XZ plane, (d) second bending mode in YZ plane, (e) Stretch and compression, (f) third bending mode in XZ plane, (g) third bending mode in YZ plane (h) rotation (i) forth bending mode in XZ plane (j) forth bending mode in YZ plane

However, some of these modes are not expected to be detectable *in vivo* or *in vitro* because of the limitations associated with instrumentation, soft tissue damping, and different wave propagation interference. The first 10 vibration modes typically dominate the vibrational motion and possess a majority of the energy; commonly they are closely examined. Therefore, the first 10 modes were used in this study.

To simulate the vibration in the frequency domain, the harmonic analysis was carried out. The frequency of the excitation force varied from 1Hz to 2600Hz. The excitation force was applied at the end of the abutment in X axis direction. There were 25 calculations. In each calculation, the frequency of the excitation force was increased by 104Hz. When the frequency of the excitation force approached the natural frequency of the model, the model had high amplitude vibration.

As shown in Fig. 60, there were three obvious peaks between 1Hz to 3000Hz. Therefore, the first peak A presented the value of the natural frequency of 1st modes as seen in Fig 58(a). The second peak B presented that of 3rd modes as seen in Fig 58(c). The third peak C presented that of 6th modes as seen in Fig 58(f). Noticeably, it is difficult to present the rest bending vibration modes since the amplitude of those modes were too small.

Figure 59 The spectrum of harmonic analysis; A, B and C are three obvious peaks between 1Hz to 3000Hz.

4.2.1 Effects of different elastic moduli

The original mechanical properties of the interface for this study were constant longitudinal elastic modulus of 18GPa and a shear modulus of 7GPa. The longitudinal elastic modulus was changed to 14GPa, 10GPa and 6GPa to simulate the decrease of mechanical properties of the interface (Papavasiliou 1997). The natural frequencies of the model with different mechanical property interface are listed in Table 4. The natural frequencies of the model decreased when the elastic modulus of the interface was reduced. The differences in NF are plotted in Fig 61. The results showed that in the low frequency range (<1000Hz) including 1st and 2nd modes the natural frequency was changed only by 2Hz (0.8%) when the elastic modulus of the interface dropped from 18GPa to 6GPa. But in the high frequency range (>3500Hz) including 9th and 10th modes, the natural frequency changed by 205Hz (3.9%). According to the detecting sensitivity, therefore, the 1st and the 2nd vibration modes are defined as the Dead Band. The 9th and the 10th vibration modes are defined as the Over Sensitive Band. The rest six vibration modes, from the 3rd to the 8th vibration modes, are defined as the Sensitive Band.

Table 4 The NF results of different elastic modulus interface.

<div>Vibration mode</div> <div>Elastic modulus</div>	1	2	3	4	5	6	7	8	9	10
18 GPa	238	240	1136	1145	2409	2454	2775	3078	4839	5142
14 GPa	236	239	1131	1141	2400	2438	2753	3050	4804	5098
10 GPa	236	239	1124	1134	2389	2425	2734	3016	4747	5026
6 GPa	236	239	1115	1126	2373	2402	2691	2958	4681	4937

Figure 60 Plot of NF differences between the different elastic modulus of the model and the classification of three vibration bands

4.2.2 Effects of the contact area

The results are presented according to the contact area at the middle, distal, and proximal portion of the implant, respectively.

(1) Simulation of middle interface failures.

To simulate the middle part interface failure, the related interface was removed from the model. Fig 62 shows an example of 25% failure in the middle part of the interface.

Figure 61 The simulation for a 25% failure in the central part of the interface

The natural frequency results are listed in Table 5. Fig 63 shows the first ten natural frequencies for the 75%, 50%, and 25% contact area failure (Papavasiliou 1997). In the Dead Band of the natural frequency (1st and 2nd mode), there was 1Hz (0.4%) difference. In the Sensitive Band of the natural frequency, a difference up to 338Hz (11.0%) was found. Even greater natural frequency differences were found in the Over Sensitive Band.

Table 5 The NF of different contact area in the middle part of interface.

<div>Vibration Contact Area</div> <div>Mode</div>	1	2	3	4	5	6	7	8	9	10
100%	238	240	1136	1145	2409	2454	2775	3078	4839	5142
75%	238	240	1134	1139	2389	2407	2667	2934	4742	4878
50%	237	239	1130	1135	2365	2388	2578	2879	4634	4704
25%	237	239	1125	1128	2344	2367	2505	2740	4502	4583

Figure 62 The effect of the middle interface failure.

(2) Proximal interface failures.

To simulate the proximal part failure of the interface, the related interface was removed. Fig 64 shows an example of 50% failure in the proximal part of the interface.

Figure 63 The simulation for a 50% failure in the proximal part of the interface

The natural frequencies of proximal interface failure are listed in Table 6. The models with various amounts of proximal interface failure are expected to be less stiff than that with corresponding middle failures. Fig 65 shows

the first 10 vibration modes with a contact area of 100%, 75%, 50% and 25%, respectively.

Table 6 The NF results of different contact area in the proximal part of interface.

<div>Vibration Mode</div> <div>Contact Area</div>	1	2	3	4	5	6	7	8	9	10
100%	238	240	1136	1145	2409	2454	2775	3078	4839	5142
75%	238	241	1111	1119	2380	2424	2770	3067	4587	4899
50%	238	241	1045	1054	2143	2156	2765	3035	3547	3686
25%	238	241	1040	1039	2127	2133	2764	3008	3506	3633

Figure 64 The effect of the proximal interface failure.

(3) Distal interface failures

To simulate the distal part failure of the interface, the distal part of interface was removed. Fig 66 shows an example of 75% failure in the distal part of the interface.

Figure 65 The simulation of a 75% contact failure from the distal end

The natural frequencies of distal interface failure are listed in Table 7. Fig 67 shows the results for different distal interface contact areas

Table 7 The NF results of different contact area in the distal part of interface. The contact area is from full contact to 25% contact.

<div>Vibration Contact Mode Area</div>	1	2	3	4	5	6	7	8	9	10
100%	238	240	1136	1145	2409	2454	2775	3078	4839	5142
75%	236	238	1105	1112	2332	2382	2564	2762	4071	4248
50%	233	235	1023	1031	2156	2202	2359	2374	2419	2611
25%	233	234	1019	1023	2149	2185	2348	2366	2417	2605

Figure 66 The effect of the distal interface failure.

4.2.3 Effects of the femur length, abutment length and soft tissue on the NF

According to the vibration Equation (27), it is anticipated that the NF would vary with the length of the femur and the surrounding soft tissue. A short femur would result in a high natural frequency. In this simulation, natural frequencies were obtained by using femur length of 24cm, 18cm, 16cm and 14cm as shown in Table 8. Fig 68 shows the natural frequencies of model with different femur length. Differences could be observed from the 1st mode natural frequency. There was a 32 Hz (13.4%) increase when a femur length reduction by 10cm. Fig 69 shows that there was a 206Hz (18.1%) increase in the 3rd mode when the femur length was reduced by 10cm.

Table 8 The NF results of different femur length.

Femur Length	Vibration Mode	1	2	3	4	5	6	7	8	9	10
24cm		238	240	1136	1145	2409	2454	2775	3078	4839	5142
18cm		238	252	1165	1212	2495	2566	2857	3220	5522	5716
16cm		252	279	1220	1293	2560	2732	2968	3430	6304	6743
14cm		270	306	1342	1412	2887	3136	3324	3653	7191	7724

Figure 67 The plot of the NFs against different femur length.

Figure 68 The NF results of 3rd vibration mode in Sensitive Band when the femur length changed from 14cm to 24cm.

The effect of soft tissue tests was simulated by changing the mechanical properties of the soft tissue surrounding the femur and the initial setting of the elastic modulus of the soft tissue was 60kPa. According to Silver-Thorn (1992), the elastic modulus of soft tissue could vary from 60kPa to 180kPa. Therefore in this study, the elastic modulus of soft tissue increased from 60kPa to 180kPa. The natural frequency results are listed in Table 9. Fig 70 shows the first 10th natural frequencies of the models with different elastic modulus. Fig 71 shows the changes in natural frequency results in the 3rd vibration mode.

Table 9 The NF results for the effect of different soft tissue condition.

<div>Elastic Modulus</div> <div>Vibration Mode</div>	1	2	3	4	5	6	7	8	9	10
60kPa	238	240	1136	1145	2409	2454	2775	3078	4839	5142
100kPa	240	248	1206	1299	2589	2623	2902	3245	5675	6044
140kPa	255	265	1323	1426	2775	2854	3245	3684	6728	6820
180kPa	284	297	1487	1634	3123	3232	4178	4352	7743	7924

Figure 69 The NF results of the first 10 vibration modes for the effect of different soft tissue condition.

Figure 70 The NF of the third vibration mode with soft tissue with different young's modulus

4.3 Discussion

The first 10 natural frequencies were computed and held as criteria for comparison as shown in Table 3. Accordingly, the first 10 vibration modes including bending, compression/ stretch, and torsion were presented in Fig 59. The harmonic analysis results are shown in Fig 60 in order to find out the amplitudes of bending vibration modes in XZ direction. Approximately, the vibrational amplitude from 1st modes was about four times higher than the amplitude from the 3rd vibrational mode; the vibrational amplitude of the 6th modes was less than half of the signal of the 3rd vibrational modes. In another word, it is better to use this technique with low frequencies since a clear signal could be acquired. Similar results were found in Qi et al. (2003). Consequently, he concluded that the first 6 vibration modes would be used *in vivo* and *in vitro* experiments because sufficient clear signals could be obtained.

4.3.1 Effects of different elastic modulus

In order to examine the changes in the natural frequency corresponding to different elastic modulus, the modulus of the interface was simulated from 18GPa, which was the cortical bone elastic modulus, to 6GPa. The results as shown in Fig 64 found that the first 10 natural frequencies dropped when the modulus of the interface reduced from 18GPa to 6GPa. A similar approach was taken by Meredith (1996). In their *in vitro* study, different mechanical properties of interface were simulated by using self-polymerising acrylic resin and the resin was allowed to polymerise around the implant fixture. A significant increase in the natural frequency was observed related to the increase in stiffness. Their study concluded that the natural frequency was increased during the curing period. Therefore, both studies confirmed that the natural frequency changed with different elastic modulus of the interfaces, as shown in Fig 61. More specifically, the interface with a high elastic modulus corresponded to a high natural frequency.

Also, three vibration bands were identified as shown in Fig 61. For the 1st vibration mode, the natural frequency was reduced by 2Hz when the elastic modulus of the interface was changed from 18GPa to 6GPa, whilst for the 2nd vibration mode, the natural frequency was reduced by 1Hz for the same change of the elastic modulus of the interface. In other words, for the 1st and 2nd vibration modes, the natural frequencies had not changed much when a mechanical property of the interface was changed from 18GPa to 6GPa. For this reason, the first two vibration modes were categorized into the Dead Band. Indeed, such a small change was difficult to detect because of the limitations associated with instrumentation, and inevitable movement. As a result, under *in vivo* or *in vitro* situation, the 1st and 2nd vibration modes might not be sufficiently helpful to detect the changes in the natural frequency. However, the details will be discussed in Chapter 6.

Next, the 3rd to the 8th vibration modes were categorized into the Sensitive Band. As shown in Fig 61, the natural frequencies corresponding to these vibration modes had obvious changes when the elastic modulus changed. For the 3rd vibration mode, the natural frequency changed by 21Hz (1.8%) when the elastic modulus of the interface changed from 18GPa to 6GPa. Under the same condition, the 4th vibration mode changed by 19Hz (1.7%), the 5th vibration mode changed by 36Hz (1.5%), the 6th vibration mode changed by 10Hz (2.1), the 7th vibration mode changed by 52Hz (3%) and the 8th vibration mode changed 120Hz (3.9%). When comparing with the Dead Band, the natural frequency changes in the Sensitive Band was greater. Hence, this vibration technique could be used as a diagnostic tool to assess the osseointegration bone-implant interface in the Sensitive Band.

Finally, regarding the 9th and the 10th vibration modes, because of notable large changes in the natural frequency, were categorized into the Over Sensitive Band which was between 3500Hz and to 6000Hz. Particularly for the 10th vibration mode, the change increased to 205Hz (4%) when the elastic modulus of the interface changed from 18GPa to 6GPa. The results indicate that the Over Sensitive Band could also be used to detect the interface changes. However, the higher frequency band was associated with low energy, and

faster attenuation. Therefore, further *in vitro* and *in vivo* experiments, it is necessary to determine which vibration band should be used. Again, the details will be discussed in Chapter 6.

Overall, as shown in Fig 61, the NFs were found decreased with a decreased elastic modulus of the interface. In terms of the general trend of the NF, the results were consistent with *in vitro* simulations reported by Meredith et al. (1998) and Huang et al. (2002).

4.3.2 Effects of different contact areas

In the study of effects of the contact area to the natural frequency, three different situations were carried out. Roberts (1988) found that “ clinically successful” implants had less than half of the osseointegration interface in direct contact with bone. Block et al. (1989) found similar levels of osseointegration for loaded implants in dogs, while Barth et al. (Barth 1990) reported integration levels approximating 80%. The percentage of osseointegration has not yet been fully investigated and one reason for this was the difficulty of assessing the percentage clinically. FE modelling makes it possible to simulate clinical situations of the implant with different percentages and patterns osseointegration (Papavasiliou 1997).

As shown in Table 5, the largest decreases in the natural frequency were often found in the Over Sensitive Band. In middle interface, the largest decrease was 559Hz (11%) when the contact area reduced from 100% to 25%. Under the same situation, the natural frequency decreased 1509Hz (29%) in proximal interface and 2422Hz (50%) in distal interface. Among these three different interface failure patterns, it is clearly that the middle interface failure had the smallest effect on the natural frequency as shown in Table 5 and Fig 63. However, the interface contact in all these three patterns had a greater effect on the natural frequency in comparison to various mechanical properties of the interface, especially in proximal and distal part of interface.

In the middle part of interface, the largest decrease was happened in both 8th and 10th vibration modes when the interface contact area had reduced 75% as shown in Fig 63. For the 1st and 2nd vibration modes, whatever the percentage of the interface contact area decreased, the natural frequencies had very small changes, which were less than 1%. However, in the Sensitive Band, the largest change of natural frequency was found at the 8th mode. There was a 338Hz (11.0 %) difference in the natural frequency. In addition, the results of middle interface contact area were spread similarly as those with different elastic modulus of the interface.

In the proximal part of interface, the different trend was observed, as shown in Table 6. As shown in Fig 65, the results of 50% and 25% contact area were almost overlapped each other. In other words, the natural frequency stayed around the same levels when the contact area reduced from 50% to 25%. Therefore, it probably indicated the difficulty in measuring the natural frequency changes when the contact area was less than 50% in the proximal part of osseointegration. On the other hand, the largest decrease was 1509Hz (29%), which was found in the 10th vibration mode when the interface contact area reduced 75%. Obviously, the proximal part of the interface had a greater effect on the natural frequency than the middle part of the interface.

In the distal part of the interface contact area, it was also found that the changes of natural frequencies did not follow the same trend as those of the middle part. As shown in Fig 67, even in the Over Sensitive Band, the natural frequency of the 10th vibration mode only changed 6Hz while the interface contact area reduced from 50% to 25%. Thus, the same difficulty could happen in measuring the natural frequency changes when the contact area was less than 50% in the distal part of osseointegration interface. Instead, the natural frequency showed a significant decrease in this part when the contact area reduced from 100% to 50%. As shown in Table 6, the natural frequency of the 3rd vibration mode, which is in the Sensitive Band, decreased by 117Hz (10%). Actually, the largest decrease in the Sensitive Band happened in the 8th vibration mode, which is 712Hz (23%). And, the largest natural frequency decrease in the Over Sensitive Band was found in the 10th vibration mode,

which was 2537Hz (49%). Therefore, it could suggest that, among these three patterns of interface contact, the distal part of the interface had the greatest effect on the natural frequency. Qi et al, (2003) also found the same result in their study.

4.3.3 Effects of the femur length and soft tissue

Based on Equation 27, it is anticipated that the natural frequency would vary with the length of the vibration system. A longer femur could result in a lower natural frequency. As shown in Fig 69, the different lengths of the femur had large differences in the natural frequencies of their implant system. For the 3rd vibration mode (within the Sensitive Band), the natural frequency of the implant with 24cm femur was 206Hz (18.1%) less than that with a 14cm femur. For the 4th vibration mode (also within the Sensitive Band), the difference under the same condition was 267Hz (23.3%) less. For the 8th vibration mode, the difference under the same condition was 575Hz (18.6%) less. Clearly, these results confirmed with the Equation 27. Also, it could suggest that the length of the femur played important role to the natural frequency of the implant system.

In the FE simulation of the effect of soft tissue on the NF of the implant system, it showed a large increase when the elastic modulus of the soft tissue increased from 60kPa to 180kPa. For the 3rd vibration mode (within the Sensitive Band), the natural frequency of the soft tissue with 60kPa was 351Hz (30.9%) less than that with 18kPa. Also, for the 6th vibration mode, the difference under the same condition was 778Hz (31.7%). It indicated that the soft tissue also had great effect on the natural frequency of the implant system.

4.4 Summary

Firstly, the purpose of the finite element modelling study was to find out what factors could influence the natural frequency of the implant system. Secondly, the results of simulating the NF of the osseointegration implant system could provide a better solution on designing the hardware for the further *in vivo* and *in vitro* experiments.

In this study, 10 modes of the NF of the model were obtained and several clinical conditions had been simulated, such as, mechanical properties of the interface, contact areas, femur length, and soft tissue conditions. Among these factors, femur length and soft tissue condition had been found to have the largest effect on the natural frequency. Next, the interface contact area also could play an important role on influencing the natural frequency of the implant system, especially the decreases which happened in the proximal and the distal parts of the implant interface. Thirdly, the mechanical properties of the interface had relatively little effect on the NF of the implant system. Finally, three NF bands were defined according to the detecting sensitivity.

In clinical applications, the first post surgery test should always be used as a baseline reference, regardless of the quality of the surgery. Since femur length and soft tissue condition are individual factors, the natural frequency results of the first post-surgery test are expected to vary from patient to patient. Normally, the length of the femur and the characteristic of the soft tissue would be regarded as fixed during the post surgery time. Hence, the vibration measurement could be used as diagnosis tool for accessing osseointegration bone-implant interface. The data of FE modelling concluded that larger osseointegration interface should result in higher natural frequency.

However, according to the literature review, the NF in the dead band (below 500 Hz) should be avoided (Qi, 2003). Meanwhile, the NF beyond 3000 Hz may not be detectable due to a number of factors may influence the technique, such as sensor sensitivity and soft tissue damping. Additionally, excitation

method had been widely used by the other research since it should not endanger a patient and would not be affected by the amount of the impulse.

Chapter 5 Measurement system development and Experiment procedure

In Chapter 2, it had been reviewed that the *in vivo* and the *in vitro* experiments using NF analysis were carried out widely and tested the osseointegration interface effectively. In Chapter 4, the FE modeling of osseointegration trans-femur implant system also showed that NF analysis would be useful in this study. In this chapter, the construction of the hardware and software for the vibration measurement are explained. The procedures of the *in vitro* and *in vivo* studies are given in details.

5.1 Design specification of the measurement system

As previously stated, the NF measurement system normally includes the mechanical excitation and the recording and analysis of the subsequent response. Recently, the impulse excitation method was widely accepted in clinical research because it was easy to apply and safe for clinical studies. Hence, the impulse excitation method was proposed for the present study. The minimum weight bearing for the osseointegration patient was 5kg. The impact time was assumed more than 0.001sec; therefore, the impulse should less than 0.05Ns.

For vibration measurement, the accelerometer should be placed on the end of the abutment to receive highest vibration signal. From the FE simulation results, it was found that the highest detectable vibration was 5142Hz in the 10th vibration mode. The measurement ranges should include the Dead Band, the Sensitive Band and the Over Sensitive Band. Therefore, the measurement range of the accelerometers needed to be over 6000Hz.

Moreover, in order to eliminate the effect on the vibration system, the mass of the accelerometer should be less than 5% of the vibration system. The FE

model was converted into a CAD software, Solid Edge, to calculate the mass of the implant system. The total mass of the implant, the abutment and the amputated femur was calculated as 0.192kg. Therefore, the weight of the accelerometer should be less than 10gm.

Finally, the AD converter was planned in order to convert the analog signal into digital signal. The analog signals collected from the accelerometer should be 6000Hz maximally. With the aim of displaying clear digital signal, the proposed sampling frequency was as 40,000Hz. Also, a tri-axis accelerometer should be applied. Therefore, an AD converter of 120Ksamples/sec throughput should be considered.

According to the NHS requirement, all devices for the clinical study need to be insulated from mains. Therefore battery with multiple outputs was needed to provide power supply for the measurement system. The proposed specifications of the NF measurement system are listed in Table 10.

Table 10 Proposed specifications for the vibration measurement system.

Excitation	Method	Impulse force
	Impulse	<0.05Ns
Accelerometer	Measurement range	>6000Hz
	Weight	<10gm
A/D converter	Throughput	>120ksamples/sec
Power supplier	Battery supply	

5.2 Excitation method

As stated before, the impulse excitation method was selected. The abutment was made to vibrate by an impulse, which was applied by dropping a steel ball attached to a thread as shown in Fig 72. The steel ball was 10g in weight and connected with a 100mm thread. It was released at 50mm height. Right after the first hit, the steel ball was held by hand to avoid rebound on the abutment. According to the calculation as shown in Appendix 1, it generated an impulse

maximum of 0.01Ns. The impulse was considered safe for the femur and the implant comparing with the initial bearing weight of 5kg.

Figure 71 The excitation method in this study.

5.3 Vibration measurement system

The vibration measurement system in this study consisted of the accelerometer, the signal conditioner, the A/D converter, the Laptop and the battery supplier as shown in Fig 73.

Figure 72 The vibration measurement system

5.3.1 Accelerometers

5.3.1.1 The function of the accelerometer

Most accelerometers generate an electrical output signal that is proportional to the induced force. This force is proportional to acceleration, according to Newton's law of motion, $F=ma$, where "F" is the induced and subsequently measured force, "m" is the mass creating the force, and "a" is the acceleration.

As shown in previous FE simulations, an impulse could generate several vibration modes of the osseointegration implant system. In this study, two types of accelerometers were used to record vibration signals on the abutment. One was single axial accelerometer to monitored the vibration in the impulse direction and the other one was tri-axial accelerometer to monitor the vibration in three axis.

Mechanical structure of shear mode piezoelectric accelerometers

Shear mode accelerometers bond the sensing crystals between a center post and a mass as shown in Fig 74. A compression ring applies a preload force required to create a rigid linear structure. Under acceleration, the mass causes a shear stress to be applied to the sensing crystals. By isolating the sensing crystals from the base and housing, shear accelerometers excel in rejecting thermal transient and base bending effects. Also, the shear geometry lends itself to small size, which minimizes mass loading effects on the test structure. With these characteristics, shear mode accelerometers offer optimum performance to meet the specification as stated before.

Figure 73 The structure of the Shear Mode ICP accelerometer.

Accelerometer sensing system

Piezoelectric accelerometers can be divided into two categories that define their mode of operation. One is electronic built-in accelerometer and the other is no electronics one. The first one is internal amplified accelerometer, which contains built-in microelectronic signal conditioning. The electronics within the accelerometers require excitation power from a constant-current regulated, DC voltage source. The second one is charge mode accelerometers, which contains the sensing element with no electronics.

5.3.1.2 The two accelerometers

As stated previously, there were two accelerometers were used. The first one was used to monitor the vibration in the impulse direction. The model 352A21 accelerometer was a high-resolution ceramic shear ICP accelerometer. It contained built-in signal conditioning electronics and required a constant-current-regulated DC voltage of 18 to 30 VDC power to operate. Also, it allowed to measure a large frequency range from 0.7Hz to 13000Hz. Furthermore, the weight of the accelerometer was 0.6g, which was less than

10g. Therefore, this model was selected in order to meet the specification of the measurement system. The key specifications are shown in Table. 11.

Table 11 Specification of Model 352A21 accelerometer.

Model	352A21
Sensitivity	10mV/g
Frequency Range ($\pm 10\%$)	0.7 to 13000Hz
Amplitude Range	± 500 g pk
Resolution	0.002grms
Connector	Side 3-56
Weight	0.6 gm

The second accelerometer was used to monitor the vibration in different direction. The model 356A24 was chosen because it was a tri-axial, low profile, ceramic shear ICP accelerometer. The key specifications are shown in Table 11.

Table 12 Specification of Model 356A24 accelerometer.

Model	356A24
Sensitivity	10mV/g
Frequency Range ($\pm 10\%$)	0.7 to 9000Hz
Amplitude Range	± 500 g pk
Resolution	0.002grms
Connector	8-36 4-Pin
Weight	3.1 gm

5.3.3 Signal conditioner

Signal conditioners are used to provide proper excitation power for accelerometers and prepare their measurement signals for readout, recording, and data acquisition. Each unit offers a per-channel null adjustment for zeroing the DC offset voltage inherent to accelerometers. The model 481A22 was used in this study. The key specification of the signal conditioner is listed in Table 13.

Table 13 Specification of the signal conditioner

Channels	4
Voltage gain	1:1
Temperature range	0-50C°
Power required	36VDC
Input and out connector	BNC JACK
Size	16cm×6.1cm×28cm
Weight	756 gm

5.3.4 A/D converter and Pcmcia Card

A data acquisition system actually takes and filters the signal and then converts it to a digital number using an analogue to digital converter. This digital value is then stored or processed by the computer. There were some aspects that needed to be considered before identify the most suitable A/D converter for this research.

Voltage signals

The signal from the accelerometer is $\pm 5V$. One of the first things to consider in choosing a data acquisition system was the voltage to be measured. It is very important to pick a board that covers the range required because it will have the greatest number of data points in the range and therefore the highest accuracy.

Resolution

Resolution refers to the number of divisions that make up the entire measurement range of the A/D. The number of divisions in the measurement range is equal to $2^{(\text{bits})}$. For example, a 14-bit A/D converter can resolve its input range into 2^{14} , or 16384, discrete values. The A/D resolution is spread over the entire input range of the A/D, so the fineness of the measurement depends not only on the resolution but also on the input range. The input signal steps = $10\text{V}/16384 = 0.6\text{mV}$

Input channels

Figure 74 The two different types of input channels.

The analog inputs on an A/D board can usually be configured as either Single-Ended or Differential. A single-ended input has one signal wire and one ground wire which may be shared across multiple inputs. The A/D board measures the difference between the signal input and the ground, or reference. A differential input has two signal wires and a ground wire. The A/D board measures the difference between the two signal wires, regardless of their voltage relative to ground as shown in Fig. 75. The differences of these input channels are listed in Table 13.

Table 14 The differences of single-ended and differential input channels.

Type	Single-Ended	Differential
Wires	2: Input and Reference	3: Input+, Input-, and Reference
Noise immunity	Less	More
Input channels	More	Less
Wiring cost	Lower	Higher

Throughput

Three elements specify A/D throughput: conversion time, acquisition time, and transfer time. Conversion time is the time required by the A/D converter to produce a digital value that corresponds to the analog input. Acquisition time corresponds to the time needed by the associated analog circuitry to acquire a signal. Transfer time corresponds to the time needed to transfer data from the interface to the computer memory. Throughput is the rate at which all three can be completed and is often the most important factor in choosing a data acquisition interface. The Nyquist Theorem specifies that an input signal should be sampled at least twice as fast as the input's highest frequency component. For example to accurately measure a 1 kHz signal, the minimum A/D throughput is 2 kHz. This avoids signal aliasing. Aliasing occurs when high frequency components appear in the digital values as erroneous lower frequencies. To measure a frequency range higher than 6000Hz accurately, the sampling frequency per channel in this research was 40 kHz. The maximum number of channels used was 3 (for the tri-axial accelerometer). Therefore the throughput for this study was:

$$40\text{kHz}\times 3\text{channels}=120\text{ksamples/sec.}$$

According to these requirements, InstruNet Model100 as shown in Fig. 76 and a Model 230 PCMCIA controller was used as A/D converter in this study. The key specifications are shown in Table. 15.

Figure 75 The InstruNet Model100 A/D Box.

Table 15 Specification of A/D box.

Key specification	
A/D conversion time	4 μ s (min)
A/D channels:	16 single ended, 8 differential
Resolution:	14-bit A/D
A/D ranges:	$\pm 5V$, $\pm 0.6V$, $\pm 80mV$, $\pm 10mV$
Throughput:	166ksamples/second max.

5.3.5 Battery supply

For clinical studies, the measurement system requires battery supply. A multiple usage battery supply was made for this study. -12v, +12v, +5v, and ground were used for the power supply for Pcmcia Card. +36v and ground were used for the signal condition and the accelerometers

5.4 Software for data collection and analysis

5.4.1 InstruNet World

InstruNet World is an application program, which can monitor and operate the InstruNet system. InstruNet World provides a spreadsheet like environment, as shown in Fig. 77, where one can set view channel parameters such as channel number, value input, units, and sensor type.

The screenshot shows the 'instruNet World' application window. At the top, there are buttons: 'Restore', 'Store', 'Open', 'Save', 'Clear', and 'Reset'. Below these is a table with columns: 'Channel', 'Addr', 'Value Input', 'Units', 'sRate', and 'Sensor'. The table contains 14 rows of data for various channels. At the bottom of the window, there are tabs: 'Network', 'Record', 'Test', and 'Script'. The 'Record' tab is currently selected.

Channel	Addr	Value Input	Units	sRate	Sensor
Ch1 Vin+	1/1/1	-1.204519	Volts	100	Voltage
Ch4 Vin+	1/1/1	-0.000354	Volts	100	Voltage
Ch7 Vin+	1/1/1	+0.007559	Volts	100	Voltage
Ch10 Vin+	1/1/1	+0.000106	Volts	100	Voltage
Ch13 Vin+	1/1/1	+0.001784	Volts	100	Voltage
Ch16 Vin+	1/1/1	-0.002689	Volts	100	Voltage
Ch19 Vin+	1/1/1	-0.000298	Volts	100	Voltage
Ch22 Vin+	1/1/1	+0.006824	Volts	100	Voltage
Ch2 Vin-	1/1/1	-0.000457	Volts	100	Voltage
Ch5 Vin-	1/1/1	+0.000857	Volts	100	Voltage
Ch8 Vin-	1/1/1	+0.007529	Volts	100	Voltage
Ch11 Vin-	1/1/1	+0.000068	Volts	100	Voltage
Ch14 Vin-	1/1/1	+0.000332	Volts	100	Voltage

Figure 76. Network page with signal from all the channels.

Each channel has its own row in the spreadsheet. The data of each channel can be recorded by a Record Page, as shown in Fig. 76. The Record Page can automatically create a separate display for each recorded channel.

The Setup dialog box, as shown in Fig. 79, was used to set the base sample rate, the number of points to be acquired per scan, the number of scans to be acquired and the recording mode (for example, oscilloscope or strip chart recorder). All InstruNet Networks were set up with one base sample rate (for example, number of points digitised per second). The data were recorded as a TXT file for FFT analysis.

Figure 77. Record page.

Figure 78 Record setup.

5.4.2 FFT analyzer

As introduced in Chapter 3, an FFT analyser stores an input signal waveform as data by digitally (discretely) sampling it, determines the Fourier coefficients in a short time, and displays the results. Since the FFT involved resolving a

signal into simple frequencies by expressing the value of frequency components (spectrum), the FFT analyser is also called a frequency analyser or spectrum analyser. In this study LabVIEW5.0 (Student Edition) was used to develop the FFT analyzer software as shown in Fig. 80. Vibration signals were firstly plotted as original collected. Useful data for analysis were then selected and analyzed by a high pass filter. The selected data were transformed by FFT into frequency domain. The highest peaks, which were the NFs, in the spectrum would be detected automatically.

Figure 79 The control panel of the FFT analyser.

5.5 Validation and calibration of the data collection system

Before the experiments, it was important to validate and calibrate the vibration measurement system. The accelerometers and the rest parts of the data acquisition system were tested separately.

5.5.1 Accelerometer calibration

Accelerometer calibration provides, with a definable degree of accuracy, the necessary link between the physical quantity being measured and the electrical signal generated by the sensor. Under normal conditions, piezoelectric sensors are extremely stable, and their calibrated performance characteristics do not change over time. PCB Piezotronics conservatively suggests a 24month recalibration cycle for these two accelerometers. The selected accelerometers were used in this study within the 24month recalibration cycles. The calibration data is shown in Table 16. The calibration certificate can be seen in Appendix 4.

Table 16 Accelerometer calibration data.

		Sensitivity (mV/g)	Output bias (VDC)	Transverse sensitivity	Discharge time constant (s)
352A21		9.12	8.3	1.3%	2.2
356A24	x	9.83	8.3	4.5%	1.3
	y	10.15	8.4	4.8%	1.2
	z	10.01	8.5	4.4%	1.3

5.5.2 Validation of the data collection system

The rest of the data acquisition system was tested using a signal generator, which was set to generate 1000Hz, 2000Hz, and 5000Hz sine wave outputs. The sampling rate for the data collection system was 40000 sample/sec. The sine wave was input to the InstruNet DAQ system and the data was processed by the FFT analyzer. For each tested frequency signal, twelve times measurements had been taken. The results are shown in Table 17. The standard

deviations of the collected data were less than 0.5% of the inputs. The results were acceptable.

Table 17 Validation of the data collection system results.

Standard	1	2	3	4	5	6	7	8	9	10	11	12	Mean	Sdev
1000Hz	1003	1000	1004	999	999	1001	998	999	1001	1003	1003	1003	1001.083	2.065
2000Hz	2001	1999	2000	2001	1999	1999	1999	1999	2000	2003	2004	2005	2000.75	2.137
5000Hz	5003	5001	4997	4998	4993	4997	5000	4997	5000	5001	5004	5004	4999.58	3.315

5.6 *In vitro* experiment

5.6.1 The setup of the *in vitro* experiments

The *in vitro* experiments were carried out by placing a threaded metallic bar as an implants in a synthetic femur model (Sawbones, Europe AB). And, a layer which simulated the variable characteristics of the interface between the femur and the implant during the rehabilitation program was constructed. Clinically, the size of implants and abutments were calculated according to the size of patients' femur, which were various from patient to patient. In this study, the implants were 17 mm and 18 mm in diameter. The length of the bar was 142 mm. It is divided into two parts: a threaded section (80 mm) and a connection bar (10 mm in diameter and 50 mm in length). The connection bar terminated with a hexagon end for prosthetic limb attachment. The bone implant system was wrapped with soft foam, which was to simulate the soft tissue around the femur. The whole model was held by a stand. The *in vitro* experiment setup is showed in Fig 81.

(b)

Figure 80 (a) A schematic diagram of the NF measurement system, (b) *In vitro* experiment setup.

The implant was made to vibrate using a controlled impact force obtained by swing a pendulum steel ball on the implant abutment. A similar method was used by Nokes et al. (1984) in the assessment of tibia fracture. In this study the input impulse was calculated to be less than 0.01Ns.

The vibration signals in this study were acquired by the single axial accelerometer (Model 352A21 accelerometer). The accelerometer was attached to the hexagon end of the abutment. The vibration signal was converted to a digital signal by a data acquisition board (iNET 100external A/D box, InstruNet). The digital data was transferred to a personal computer through a PCMCIA interface card (Model 230 PCMCIA controller, instruNet) and saved as TXT files. The data file was processed by using Fast Fourier Transform software.

5.6.2 The procedure of the *in vitro* experiments

Firstly, the NF tests were carried out by varying implant length exposed outside the femur (defined as abutment length hereafter) from 57mm to 73mm. The tests were carried out on two implant, 17mm and 18mm. The measurements for each implant were repeated 12 times.

Secondly, tests with three silicone rubbers (SILASTIC T4, MOLDSIL MIDDLE and MOLDSIL SOFT from Reptech) were carried out to simulate various states of the osseointegration layer between the implant and bone. Three silicone rubbers with different elastic modulus were used to simulate the bone implant interface. They were used to reflect the different stiffness, hence the mechanical characteristics of the osseointegration interface. The tests were carried out by using a 17mm diameter implant. To place silicone rubber interface, the synthetic femurs were drilled and tapped to produce an 18mm hole with thread. The silicone rubber was deposited on the 17 mm implant uniformly before it was inserted into the femoral model. The insertion was carried out carefully to make sure there was a 0.5mm silicone layer between bone and implant. In this way, three controlled interfaces, which reflected different interfaces, were created. The NFs were measured against the curing time of the silicone rubber to test how the interface stiffness affects the NF. Table 18 shows the tests of the *in vitro* experiments.

Table 18 The *in vitro* experiment setup

Test setting	
Test of different abutment length	17mm and 18mm diameter
Test of different interface layer	17mm diameter only

5.6.3 Testing the elastic modulus of three silicone rubbers

Three different silicone rubbers were selected to simulate the interface between the implant and the bone in the *in vitro* experiment. The silicone rubbers were measured by the material testing machine from Merlin Series 5500 Testing System as shown in Fig. 82. Each of the silicone rubbers was formed into a cylindrical specimen for the tensile test as shown in Fig. 83.

Figure 81. Mechanical properties test machine.

Figure 82. Tensile test of the silicone rubber.

Young's Modulus of MOLDSIL SOFT, MOLDSIL MIDDLE and Silastic T4 were 0.173MPa, 0.319MPa and 0.653MPa, respectively.

5.7 *In vivo* experiment

5.7.1 The preparation of *in vivo* experiment

The *in vivo* experiments were conducted at Queen Mary's Hospital, Roehampton London. A proposal, as shown in Appendix 2, about the clinical study was written and approved by the local ethical committee of NHS. 3 patients have been inspected by the vibration the measurement. The patients' information is listed in the Table 19. The information sheet of the study and the patient's consent form were given to each patient before the first test. All the patients who have participated in this research program have given their permissions and the consent form and the information sheet can be seen in Appendix 3 and 4.

Table 19 The clinical records of candidates.

Patient	Gender	Stage 1	Stage 2	Rehabilitation Time (day)	Body weight
A	Male	24/04/02	15/10/02	54	94.6kg
B	Male	24/09/03	24/03/04	67	78kg
C	Female	15/10/03	28/04/04	202	55kg

The vibration measurement procedure of the *in vivo* test was same as the *in vitro* study. Both a single-axial and a tri-axial accelerometer were used for the *in vivo* measurements. The tests were carried out when the patient was supine as shown in Fig 84. The patient's residual stump was rested on the bed to avoid muscular tension and to reduce any possible effect of surrounding soft tissue on the NF of the implant system. The excitation method used in the *in vivo* study was the same as for the *in vitro* study. A pendulum ball was used, as shown in Fig 85. The accelerometer was carefully attached on the flat surface of the end of the abutment with wax. For each test, the measurements were repeated 12 times and each test took 20 minutes.

Figure 83 The position of the patient in the *in vivo* experiments

Figure 84 The excitation method in the *in vivo* experiments

5.7.2 The procedure of in vivo experiment

The patients, who participated in the *in vivo* study, commenced weight bearing rehabilitation six weeks after the stage 2 operation. The weight bearing started with an application of a 20kg compression load through a short training prosthesis. Scales were used to monitor the loading as shown in Fig 86. This is graduated from 15 to 30 minutes twice a day. The amount of force applied through the prosthesis was increased by 10kg each week until the full body weight was reached. A pain scale form (See Appendix 5) had been completed by patients after each weight bearing. The pain index was from 0 to 10. 0 to 2 was 'ok', 3 to 5 was 'allowed', 6 to 10 was 'not allowed'. Pain or discomfort feeling during the weight bearing would require a medical review.

Figure 85 The weight bearing with a short training prosthesis

It took two to three months before candidates were able to apply a full body weight. After full body weight was reached, candidates were fitted with their full-length prosthesis. Gait rehabilitation commenced with the candidate partially weight bearing through the prosthesis between parallel bars. The *in vivo* studies were taken during this period.

5.7.3 The rehabilitation process of the three patients

For patient A, his weight bearing amount against time is plotted in Fig 87. The first test was carried out before the first weight bearing. The last test was taken on day 68 since the weight bearing exercises began. In the first three days, the test was carried out everyday. After the first three days, the test interval was increased to every two weeks. The weight bearing exercise reached the full body weight of the patient on day 50. During the entire period, 9 tests were carried out. The measurement was repeated twelve times in each test. Two accelerometers were used: a single-axial accelerometer and a tri-axial accelerometer.

Figure 86 Weight bearing record for the patient A.

Only the single-axial accelerometer was used in the tests of patient B. There were measurements taken before and after the weight bearing to compare the effect of the weight bearing to the natural frequency. In the first two days the test was carried out daily. After the first two days, the test interval was gradually increased to about every two weeks. With patient B, the weight bearing exercise reached the full body weight on day 36 and then the patient started walking with a prosthetic limb with two crutches as support. The two tests, before and after weight bearing, continued until day 67, whilst only one

test was carried out after that. Following up measurements continued to day 226. The weight bearing is plotted against time in Fig 88.

Figure 87 Weight bearing record for the patient B.

For patient C, the weight bearing exercise reached the full body weight of the patient at day 179. The weight bearing started with an application of a 5kg compression load through a short training prosthesis. The amount of force applied through the prosthesis was increased by 5kg each week. The first test was carried out before the first weight bearing exercise and single axial accelerometer was used. The measurements lasted for 258 days. The weight bearing exercise reached the full body weight of the patient on day 73, however due to pain, the weight bearing was reduced to 40kg. Instead of continued loading for 30 minutes, to reduce the pain, patient C performed the weight bearing for 1 to 5 minutes then took a break for a few seconds, however the total amount of time was still 30 minutes. The weight bearing gradually increased to full body weight and on day 179, patient C started to walk with a prosthetic limb. The actual progress of weight bearing is shown in Fig 89. From day 180, the patient started walk with prosthetic limb with aid of two crutches. There were two tests, before and after weight bearing, which continued to day 188. There were 14 tests carried out for patient C and because of slowness in rehabilitation, the whole vibration measurements lasted 258 days. From the first weight bearing to walking with a prosthetic limb, it took

202 days and there were around two weeks between every two tests. In total, 22 tests were carried out.

Figure 88 Weight bearing record for the patient C.

Chapter 6 Results

In previous chapters, the need of *in vitro* and *in vivo* study was outlined and the fundamental theories were explained. The FE modelling also effectively simulated the factors, which could affect the NF of the osseointegration implant system. Moreover, the results of FE modelling provided the frequency range for measurement in the *in vitro* and *in vivo* experiments. After introducing the vibration measurement system and the procedure of the *in vitro* and *in vivo* experiments in Chapter 5, the results are presented in this chapter.

6.1 *In vitro* results

6.1.1 Detection of NF

A typical signal in time domain is illustrated in Fig 90. The signal was obtained in 90th minute with SILASTIC T4 silicone rubber as interface in the *in vitro* experiment. The same signal in the frequency domain is shown in Fig 91. The natural frequency could be read clearly from the peak value window. There were three obvious NF peaks as shown in Fig 92. The first NF peak was at 289Hz. The second peak was at 2894Hz. The third peak was at 8113Hz. There was no other obvious NF peak in the figure.

Figure 89 A typical vibration signal of a tested implant model obtained in 90th minute with SILASTIC T4 silicone rubber as interface.

Figure 90. The same vibration signal in frequency domain obtained in 90th minute with SILASTIC T4 silicone rubber

Figure 91 Three obvious peaks in the spectrum window of the signal obtained in 90th minute with SILASTIC T4 silicone rubber as interface.

Another typical NF spectrum of the signal obtained at the 3000th minute with SILASTIC T4 silicone rubber as interface is shown in Fig 93. It is similar as the signal obtained at the 90th minute. Also, there were three obvious NF peaks in the spectrum. As shown in Fig 93, the first NF peak was at 303Hz. The

second peak was at 2970Hz. The third peak was at 8631Hz. Similarly, no other obvious NF peak was found.

Figure 92 Three obvious peaks in the spectrum window of the signal obtained in 3000th minute with SILASTIC T4 silicone rubber as interface.

However, in some measurements, there were only two obvious peaks as shown in Fig 94. The third peak was not clearly showed in the Over Sensitive Band. It is because higher harmonic peaks were associated with low energy, and faster attenuation, Therefore, the NF in the Over Sensitive Band would not be considered.

Figure 93 No obvious peaks in the Over Sensitive Band in some measurements. The signal was obtained in 3000th minute with SILASTIC T4 silicone rubber as interface.

Moreover, according to the results of the previous FE simulations, the NF changes in the Dead Band ($< 1000\text{Hz}$) was not detectable due to the limitations associated with instrumentation, and inevitable movement. Only the Sensitive Band was considered suitable for NF analysis in this study. To get the Sensitive Band natural frequency, a high pass filter was used in the signal analyser. A vibration signal spectrum of the filtered signal is shown in Fig 95. By applying high pass filter, the low frequency NF peak can be removed. Consequently, the NF in the Sensitive Band could be detected automatically.

Figure 94 A spectrum of filtered signal obtained in 3000th minute with SILASTIC T4 silicone rubber as interface.

6.1.2 The NF test of different abutment length and implant diameter

The *in vitro* measurement was repeated 12 times for both 17mm and 18mm diameter implants at the each different abutment lengths. The NF results of the 17mm and 18mm-diameter implants are listed in Table 20. The NF of the 18mm-diameter implant was higher than that of 17mm-diameter implant when their abutment lengths were same. For each implant, the NF decreased as the abutment length increased. To compare the two different diameter implants, a representative plot of NF against abutment length is shown in Fig 96.

Table 20 NF of 17mm and 18mm implant. (Mean NF $\pm$ SD, n=12)

Abutment Length (mm)	57	61	65	69	73
NF of 17mm implant	3258 $\pm$ 10	3237 $\pm$ 5	3180 $\pm$ 7	3107 $\pm$ 8	3080 $\pm$ 3
NF of 18mm implant	3405 $\pm$ 6	3318 $\pm$ 8	3248 $\pm$ 4	3194 $\pm$ 5	3131 $\pm$ 6

Figure 95 The NF values of different implant diameter and abutment length (Mean NF ± SD)

6.1.3 NF test on different elastic modulus interface

Three types of silicone rubber were used as interface layer between implant and bone in order to simulate how the natural frequency was affected by the different elastic modulus of osseointegration interface. The natural frequencies measured on the abutment with SILASTIC T4, MOLDSIL MIDDLE and MOLDSIL SOFT are listed in Table 21. The curing times of these three silicone rubbers were 24 hours. As shown in Fig 97, the natural frequencies of three types of silicone rubber were found steadily increased throughout the curing time.

Table 21 NF of abutment with different interfaces. ST, SILASTIC T4. MM, MOLDSIL MIDDLE. MS, MOLDSIL SOFT. (Mean NF ± SD, n=12)

Curing Time (minute)	30	90	150	210	330	450	570	690	1500	3000
NF of ST (Hz)	2844±12	2897±7	2902±17	2908±8	2903±22	2912±16	2897±22	2911±16	2963±43	2970±5
NF of MM (Hz)	2835±6	2850±11	2861±25	2865±10	2869±8	2873±17	2879±6	2886±4	2901±9	2907±6
NF of MS (Hz)	2834±30	2831±35	2838±21	2836±24	2849±10	2849±21	2860±15	2860±13	2882±13	2882±8

Figure 96 The NF values of using SILASTIC T4, MOLDSIL MIDDLE, and MOLDSIL SOFT silicone rubber as interface during the curing time (Mean NF $\pm$ SD)

Moreover, the NF of the abutment was found higher when the elastic modulus of interface layer was higher. In this study, the relationship between the NFs of the implant and the elastic modulus of the interface materials was found significantly correlated as shown in Fig 98 since the correlation coefficient is 0.9998 and $p < 0.05$.

Figure 97. Plot of NF values of three interface materials with different Young's Modulus at the time of 3000minutes (Mean NF $\pm$ SD)

6.2 *In vivo* results

6.2.1 Detection of NF

As stated in Chapter 5, the *in vivo* tests were carried out on three osseointegration trans-femoral implant patients. An original vibration signal spectrum is shown in Fig 99. The signal was taken from the measurement of patient C at day 1 after the weight bearing. The NF peaks are 322Hz, 1733Hz, and 2702Hz. They present the NFs of XZ direction in Dead Band, Sensitive Band and Over Sensitive Band as seen in Fig 59. Another vibration signal spectrum is shown in Fig 100. The NF peaks are 325Hz, 1750Hz and 2734Hz. In order to get the NF of the Sensitive Band, vibration signals for *in vivo* study were filtered by a high pass filter. The high pass filter was set at 700Hz. As shown in Fig. 100, the NF in the Sensitive Band could be detected automatically for all three patients.

Figure 98 A typical vibration signal spectrum taken from patient C at day 1 after the weight bearing.

Figure 99 A typical vibration signal spectrum taken from patient C at day 45 before the weight bearing.

(a)

(b)

(c)

Figure 100. A typical vibration signal spectrum from *in vivo* test. (a) Patient A (b) Patient B (c) Patient C.

Due to the differences of the soft tissue, bone condition, and abutment structure, the NF patterns of each patient were expected to be different. However, there were some common features among the NF patterns of all three patients. As shown in Fig. 101, there were three NF peaks found on each patient's vibration signal spectrum. They presented different vibration modes.

(a)

(b)

(c)

Figure 101. Common features in the FFT pattern. (a) Patient A (b) Patient B (c) Patient C.

The impulse excitation applied in this study made the abutment vibrate in three directions. In the test with patient A, a tri-axial accelerometer was used and the signals in the three axis were presented in Fig.102. The natural frequency of each direction was found different as shown in Fig. 103.

Figure 102. A vibration signal from tri-axial accelerometer taken from patient A on the first measurement.

(a)

(b)

(c)

Figure 103. The FFT of tri-axial accelerometer. (a) Horizontal direction, the same direction as the single axial accelerometer (b) Vertical direction (c) Axial direction.

In fact, the excitation force can generate the vibration in all directions as shown in Fig 104. The tri-axial accelerometer is able to record the vibration of three directions: x, y, and z, while the single one can only record the vibration of one direction. In this study, the single axial accelerometer was placed to measure the vibration of horizontal direction because in this direction the impact force generated the largest vibration signal.

Figure 104 A schematic diagram of how the excitation force produced vibrations in different directions.

6.2.2 Results of patient A

In the *in vivo* tests on Patient A, two accelerometers were used; a single-axial accelerometer and a tri-axial accelerometer. In total, nine tests were performed and the measurement was repeated twelve times in each test. The results of natural frequencies are listed in Table 22.

Table 22 *In vivo* natural frequency results of patient A. (Mean NF ± SD, n=12)

Time (day)		1	2	3	4	18	25	39	54	68
Single Accelerometer		1268±8	1249±11	1252±18	1252±15	1266±23	1273±15	1281±12	1280±8	1280±10
Tri-axial Accelerometer	X	1262±12	1231±12	1213±11	1215±13	1247±15	1257±13	1269±17	1265±12	1267±8
	Y	1735±23	1692±16	1695±31	1714±41	1740±38	1735±36	1757±25	1756±18	1758±37
	Z	1006±24	1013±21	1001±8	1017±18	1015±12	1016±20	1027±15	1031±17	1038±21

The results from the single axial accelerometer are presented in Fig. 105. At day 1, the mean value of the NF measured before weight bearing was 1268 ± 8 Hz. After the first weight bearing, the mean NF dropped significantly to 1249 ± 11 Hz ($p < 0.001$) at day 2. Then, the NF returned to 1268 Hz, which is same as the NF before the weight bearing at around day 24. At day 39, the NF increased gradually to its highest level about 1280 ± 8 Hz. The NF stayed at this level until the full body weight was reached at day 50. The correlation coefficient was 0.9092 and $p < 0.002$ between NFs and rehabilitation time (from day 2 to day 39). Beyond this point, the general trend of the NF was in steady state equilibrium. In addition, the correlation coefficient was 0.9458 ($p < 0.002$) between NFs and weight bearing during day 2 to day 50). Besides, the pain indexes were filled in all '0' after each time weight bearing. It could indicate that there were no pain and discomfort during his weight bearing rehabilitation.

Figure 105 The NF results from single accelerometer measurement of Patient A (Mean NF $\pm$ SD)

In addition, the tri-axial accelerometer results are shown in Fig 106. In average, the NF of single-axial accelerometer is 20 Hz higher than that of the x-axial direction of the tri-axial accelerometer. The correlation coefficient between

NFs of the single-axial accelerometer and those of the x-axial direction of the tri-axial accelerometer was 0.9329 ($p=0.0002$). And, the correlation coefficient between the NFs of the single axial accelerometer and those of the y-axial direction of the tri-axial accelerometer was 0.9621 ($p<0.0001$). The correlation coefficient between the NFs of the single-axial accelerometer and those of the z-axial direction of the tri-axial accelerometer was 0.7281 ($p<0.03$).

(a)

(b)

(c)

Figure 106. The results of the tri-axis accelerometer of Patient A. (a) Horizontal direction (b) Vertical direction (c) Axial direction.

6.2.3 Results of patient B

For patient B, only a single-axis accelerometer was used. Before day 67, the measurements were taken before and after the weight bearing on the same day in order to compare the effect of the weight bearing on the NF results. The mean values of natural frequencies are listed in Table 23. It was found that the NF dropped about 20Hz after weight bearing each time. Also, the trend of the NF was not in steady state equilibrium. For instance, the NF was found dropped from 1408Hz to 1382Hz at day 18. Then, the NF was 1402Hz when the full body weight bearing was reached at day 36. However, the NF was up and down since then. Until day 163, the NF raised up to 1412Hz, which was similar level as the NF of the first day of full body weight bearing.

Table 23 *In vivo* natural frequency results from patient B. (Mean NF ± SD, n=12)

Time (day)	1	2	11	18	32	53	60
Before WB	1373±25	1349±24	1408±25	1382±17	1402±14	1402±17	1393±18
After WB	1355±18	1330±14	1381±14	1370±20	1387±19	1393±17	1376±17
Time (day)	67	87	95	141	163	198	226
Before WB	1394±15	1395±15	1390±21	1383±24	1412±14	1420±16	1429±18
After WB	1383±23						

Moreover, Fig 107 shows the NF results of patient B before weight bearing. The natural frequency results of before and after weight bearing are plotted in Fig 108. The correlation coefficient was 0.5733 ($p=0.1785$) between NF before the weight bearing (except the 1st measurement) and bearing weight. However, the correlation coefficient was 0.8262 ($p<0.05$) between the NF and the rehabilitation time during day 141 and day 226.

Figure 107 The *in vivo* NF results of the patient B (before weight bearing).

Figure 108 The natural frequencies of the patient B (before and after weight bearing).

The pain index record is plotted in Fig 109. To compare the pain index results with the NF, the pain indexes were averaged between the day of last test and

the day before the new test. The time-averaged pain index is plotted in Fig 110. The correlation coefficient is 0.3400 ($p=0.4$) between the NF and average pain index. The result shows no significant relationship between the NF values and pain index of patient B. Noticeably, on day 18, the average pain index reached to a high value; meanwhile, the NF had a sharp drop of 26Hz.

Figure 109 Pain scale data of patient B

Figure 110 The graph of pain scale data averaged between the day of last test and the day before new test.

6.2.4 Results of patient C

For patient C, the measurements were taken before and after the weight bearing to compare the effect of weight bearing to the natural frequency. The natural frequency results of patient C are listed in Table 24. It was found that the NF was up and down throughout 258 days. No obvious trend was observed in general. Yet, it was found that the NF dropped more than 20Hz after weight bearing each time. There were 8 times when the difference was more than 20Hz among 19 tests. The highest decrease was 37Hz on day 79.

And, Fig 111 shows the NF results before each weight bearing. The natural frequencies of before and after each weight bearing were plotted in Fig 112. The correlation coefficient was 0.1887 ($p=0.5007$) between NF before the weight bearing (except the 1st measurement) and weight bearing. It illustrated that no statistically significant trend was found in the results.

Table 24 *In vivo* natural frequency results of patient C. (Mean NF $\pm$ SD, n=12)

Time (day)	1	8	17	28	31	38	45	59
Before WB	1759 $\pm$ 30	1750 $\pm$ 16	1760 $\pm$ 23	1744 $\pm$ 21	1716 $\pm$ 10	1745 $\pm$ 21	1757 $\pm$ 10	1742 $\pm$ 16
After WB	1733 $\pm$ 20	1723 $\pm$ 18	1735 $\pm$ 24	1722 $\pm$ 19	1699 $\pm$ 16	1721 $\pm$ 17	1740 $\pm$ 17	1716 $\pm$ 17
Time (day)	79	89	95	109	130	146	166	179
Before WB	1762 $\pm$ 11	1750 $\pm$ 17	1747 $\pm$ 18	1719 $\pm$ 15	1743 $\pm$ 15	1755 $\pm$ 16	1745 $\pm$ 12	1726 $\pm$ 12
After WB	1725 $\pm$ 18	1735 $\pm$ 14	1728 $\pm$ 20	1716 $\pm$ 21	1728 $\pm$ 18	1725 $\pm$ 22	1726 $\pm$ 25	
Time (day)	180	182	188	202	223	258		
Before WB	1727 $\pm$ 12	1741 $\pm$ 17	1737 $\pm$ 12	1736 $\pm$ 17	1745 $\pm$ 16	1744 $\pm$ 20

Figure 111 The *in vivo* NF measurement results of the patient C before weight bearing.

Figure 112 A plot of the natural frequency results of patient C (before and after weight bearing).

The pain index record of patient C is plotted in Fig 113. To compare the pain index with natural frequency, the pain indexes were averaged between the day of last test and the day before the new test. The pain index in average is plotted in Fig 114. The correlation coefficient was -0.1887 ($p=0.0769$) between NF before the weight bearing (except the 1st measurement) and the pain index. It was considered not statistically significant.

Interestingly, there were three noticeable increases happened in pain scales of patient C as shown in Fig 113. The first increase happened during day17 to day 28 when the NF dropped 16Hz. After this peak, though, the NF continuously dropped 28Hz until day 31. Next, the second increase was found during day 79 to day 95 when the NF decreased 15Hz. Again, the NF continuously dropped 28Hz until day 109. Finally, the last increase was observed during day 146 to day 166 while the NF had a drop of 10Hz.

Figure 113 A plot of the pain scale data of patients C

Figure 114 The graph of the pain scale data averaged each time between the last and present test.

Chapter 7 Discussion

Based on the results presented in Chapter 6, the discussion is developed in this chapter according to the *in vitro* and *in vivo* studies separately. Firstly, the discussion of the *in vitro* study is made in terms of the effect of different implant size and different elastic modulus of the interface. Then, the discussion in the *in vivo* study is made according to the three clinical factors in the rehabilitation program, rehabilitation time, weight bearing and pain index. In the end, the limitation of the study is mentioned.

7.1 *In vitro* study

7.1.1 Detection of NF

In FE study, there were 10 vibration modes demonstrated because the modeling simulation was able to show the entire vibration modes in different direction, such as bending, and torsion. In comparison, there were only a limited number of vibration modes detected in the *in vitro* study due to the direction of the excitation and the limited sensitivity of the accelerometers. Similar results were found by other researchers. In a paper by Nakatsuchi *et al.* (1996), vibrational modes were examined by simulating callus consolidated on an excised human tibia. Two separate modes, which were the vibration in the lateral and the anterior posterior direction, were found. The vibrational mode in the lateral direction was identified as the primary mode to their study since the one in the anterior posterior direction was relatively weaker. The analysis based on the vibrational mode of the lateral direction effectively investigated bone fracture healing. In this *in vitro* study, the accelerometer was placed in the same direction of the excitation. Similarly, the vibration in the lateral direction was considered as the primary vibration mode for assessing the implant-bone interface.

In the FE study of the NF on osseointegration trans-femoral implant, three vibration bands were defined according to the sensitivity of detecting. Even though, Qi (2003) suggested that the NF in the low frequency range (the Dead Band) could generate large and clear signal, which could be easily detected. The results in the FE study, however, showed that the NF in the low frequency range was not sensitive to response the property change of the osseointegration implant system. Also, in the experiment on the curing time of T4 silicone rubber interface, as shown in Fig 92 and 93, the Dead Band natural frequency had 14Hz change only, while the Sensitive Band had 74Hz difference, and the Over Sensitive Band had 518Hz difference. This finding was consistent with FE study. In fact, the small change in the Dead Band was not able to entirely detect under the *in vivo* or *in vitro* situation because of the limitations associated with instrumentation, and inevitable movement. As a result, the high pass filter, which was set at 800Hz, was applied in order to achieve automatic NF detection in the signal analysis software.

Moreover, as shown in Fig 92 and Fig 93, three obvious frequency peaks were detected in the signal obtained in 90th and 3000th minutes. This was constant with the findings of the FEA results in this study as shown in Fig 59. However, the third harmonic peak in the Over Sensitive Band was not obvious in some measurements as shown in Fig 94. Although, increasing the excitation force may improve the signal in the Over Sensitive Band. Yet, it may bring risk in the clinical situation. Hence, the NF in the Over Sensitive Band was considered not reliable to the *in vivo* and *in vitro* studies. Consequently, the natural frequency result of the Sensitive Band (1000Hz-3500Hz) was finally identified as the most suitable range for the *in vitro* experiments studies.

7.1.2 Implant size

According to the vibration Equation (27), there were two factors affecting the NF, l and I , the length of the abutment and the moment of inertia. The NF increases as the moment of inertia increases. When the abutment length was fixed, the implant with a larger diameter had a higher NF. This was because the

increase of the implant diameter caused an increase in the moment of inertia of the implant system. In the *in vitro* study, the NF of the 18mm-diameter implant was higher than that of 17mm-diameter implant when their abutment lengths were same as shown in Fig 96. Clearly, there was an agreement drawn between Equation 1 and the *in vitro* study.

For a vibration system such as bone-implant, the NF could have a correlation with the effective length of the vibration beam. The simulation in FE study was to change of the femur length in order to change the effective length of the vibration beam. The NF was found decrease when femur length increased. These results were consistent with the findings of previous researchers' study (Meredith, 1997; Meredith, 1998; Huang, 2001). The increase of the abutment length in the *in vitro* study was actually to increase the effective length of the vibration beam, which was equal to the femur length plus the abutment length. As shown in Fig 99 the natural frequency of the abutment decreased linearly when the abutment length was increased. Both the FEA and *in vitro* study were in agreement with the vibration theory in Chapter 3.

7.1.3 Interface between the implant and the bone

Implant stability is related to the conditions of the implant-bone interface. As the elastic modulus of the interface represents a major factor of the implant-bone interface, three silicone rubbers with different elastic modulus were used in the *in vitro* study. Only shore hardness was given by the silicone rubber company for the mechanical property. Therefore, the three silicone rubbers were tested by a mechanical property test machine to get the elastic modulus. As shown in Fig 97, in 3000th minute of curing time, the NF with MOLDSIL SOFT interface (0.173MPa) was 98Hz (3.2%) less than that with SILASTIC T4 interface (0.653MPa). Moreover, as shown in Fig 98, linear relationship was found between the natural frequencies and young's modulus of the silicone rubber interface. In other words, the result of *in vitro* study showed that the NF

with higher elastic modulus of the interface is higher than those with lower elastic modulus of the interface. .

Similarly, in FE study, the NF result of the 8th vibration mode (about 3000Hz) was found decreased 120Hz (3.8%) when the elastic modulus of the interface reduced from 18GPa to 6GPa. This finding confirmed that the natural frequency changed with different elastic modulus of the interfaces as shown in Fig 61. In detail, the interface with a high elastic modulus corresponded to a high natural frequency. In other words, the findings in both *in vitro* and FE studies supported each other.

In the *in vitro* study, to simulate a full contact situation of the implant model, an interface layer with changing modulus was used. As the silicone rubber layer between the femur and implant cured, the interface changed from liquid to solid. Fig 97 illustrates the natural frequency changes against curing time. It showed that the NF measurement was sensitive to the changes in the interface.

These findings agreed with the theory well and were consistent with the results of *in vitro* simulations reported by Meredith et al. (Meredith 1996) and Huang et al. (Huang 2000). In the *in vitro* study by Meredith (1996), different mechanical properties of interface were simulated by using self-polymerising acrylic resin. The resin was allowed to polymerise around the implant fixture. During the curing period, the natural frequency was found increased.

Indeed, this *in vitro* study not only confirmed the previous studies, but also formed a valid basis for further investigation of the osseointegration interface. The study proved that the accelerometer was sensitive to the changes of the mechanical properties as the silicone rubber hardens progressively. It indicated that the NF analysis could be a potential method to monitor changes at the implant-bone interface during the rehabilitation.

7.2 *In vivo* study

7.2.1 The detection of NF

In this *in vivo* study, the abutment and the implant system had a much more complex structure. Therefore, the FFT spectrum of the *in vivo* study showed more complex features. Compared to the NF spectrum of *in vitro* measurement, the peaks of the NF from *in vivo* signal were much less obvious. As shown in Fig 98, there were three obvious peaks in the Dead Band and the Sensitive Band, 322Hz, 1733Hz and 2702Hz. There was no peak after 3500Hz (the Over Sensitive Band). This finding was consistent with the results of FE study and *in vitro* experiments that the NF in the Over Sensitive Band was difficult to detect under small excitation force. Also, in the measurement on the day45 of patient C, as shown Fig 99, the Dead Band natural frequency had 3Hz change only, while the Sensitive Band had 17Hz difference, and the Over Sensitive Band had 32Hz difference. This finding was consistent with FE study and the *in vitro* experiments. Again, the natural frequencies in Sensitive Band were found more suitable to determine the changes in stability as a reflection of interface between the implant and the bone. Thus, the high pass filter was set at 800Hz to establish automatic NF detection by the signal analysis software. However, due to inevitable muscle movements during the tests, the standard deviations of the *in vivo* results were found higher than that of the *in vitro* results.

The typical NF spectrums of the vibration signal from three patients are shown in Fig 100. Because of the soft tissue effect, bone condition, and the implant dimension, the NF patterns of each patient were different. However, in the Sensitive Band, there was common feature found for all three patients. As shown in Fig 101, three NF peaks were able to be observed in each signal spectrum for all three patients. Similarly, in Cawley's paper (1998), the different vibration modes were discovered in their *in vivo* study. The beam transducer was suggested to prevent a double peak in the natural frequency. It was achieved by changing depth-width ratio of the transducer. In his conclusion, if the thickness and the width were similar, two natural frequencies

in different directions would appear. In this study, the abutment of the osseointegration trans-femoral implant patient was a cylinder with a square end. The flat surface was not vertical. The accelerometer recorded not only the vibration on horizontal plane, which was the same as the excitation direction, but also vertical plane and the axial direction.

For a cantilever beam vibration, it would generate different vibrations including those in the vertical, the horizontal and axial direction. Each direction had its own vibration mode. A typical vibration signal from a tri-axial accelerometer was shown in Fig 102. Similar as the finding of the FE study, these three most apparent NF peaks were natural frequencies of the abutment in different directions. The NF peak of the x-direction signal was 1284Hz as shown in Fig 106(a). The NF peak of 1691Hz had the highest amplitude in the y-direction as shown in Fig 106(b). The NF peak of 977Hz had the highest amplitude in the z-direction as shown in Fig 106(c). Therefore, the three NF peaks were found to be the vibration in x, y and z-axial direction respectively.

Moreover, both single and tri-axial accelerometers were used in Patient A's measurement throughout 9 tests. The NF of single-axial accelerometer was 20Hz higher than that of the x-direction of the tri-axial accelerometer. According to the Equation 1, high moment inertial resulted in a high NF. The tri-axial accelerometer was 2.5gm heavier than the single axial accelerometer in weight. Therefore, the NFs of the tri-axial accelerometer were expected lower than those of the single-axial accelerometer due to its lower moment inertial.

The correlation coefficient between NF from the single-axial accelerometer and the x direction of the tri-axial accelerometer, which is same direction of the single axial accelerometer, was 0.9329 ($p=0.0002$). It meant that the NF trend from the single-axial accelerometer was statistically correlated with the one from the x direction of the tri-axial accelerometer. Similarly, the correlation coefficient between NF from the single axial accelerometer and the y direction

was 0.9621 ($p < 0.0001$). And, the correlation coefficient between the NF from the single-axial accelerometer and the z direction of the tri-axial accelerometer was 0.7281. ($p < 0.03$) Therefore, the vibration signals of all three directions from the tri-axial accelerometer were statistically in strong correlation with that from the single-axial accelerometer. In other words, the NF results from the single axial accelerometer demonstrated the similar trend as those from the tri-axial accelerometer. However, the tri-axial accelerometer had greater influence on the NF due to the heavier weight attached to the end of the abutment. Hence, only the NF results obtained from the single accelerometer would be assess in the following discussion.

7.2.2 NF and rehabilitation time

The process of osseointegration could be affected by many factors, including the surgical technique and the conditions of the bone (Ericsson, 1994). Furthermore, clinical observations indicated that the final healing was affected by individual differences and operation conditions (Friberg, 1995). Both *in vivo* and *in vitro* studies in the previous research showed that implants with a lower natural frequency value had poorer initial stability and a longer healing period than those with well-fitted conditions (Huang, 2003). Additionally, the previous *in vivo* studies showed that the stability of the implant was affected by the healing time. (Meredith, 1994; Huang, 2003) In this study, the rehabilitation time varied from patient to patient due to different individual condition. Three osseointegration trans-femoral implant patients were followed up during their training program in order to discuss the relationship between NF and rehabilitation time.

For patient A, as shown in Fig 105, the NF results dropped after first day loading. With the training continuing, the natural frequencies increased steadily. A 31Hz increase of the NF was found between the first day weight bearing and the day of the full body weight bearing. On day 38, the NF had reached the highest level about 1280 ± 7.8 Hz and stayed still afterwards as shown in Fig 105. The results demonstrated that during the rehabilitation

program the stability of the implant was improved gradually during the first 38 days and kept stable since then. The correlation coefficient between the NF and rehabilitation time (from day 2 to day 39) was 0.9092 and $p < 0.002$. It could indicate the NF was in very strong correlation to the rehabilitation time in case of Patient A. It could be a reflection of the osseointegration process of patient A.

For patient B, as listed in Table 23, a 59Hz increase of the NF was observed during day 2 to day 11. At day 36, patient B finished the weight bearing exercises and walked with a prosthetic limb with aid of two crutches. The results could explain that the weight bearing stimulated the bone to grow and improved the elastic modulus of the interface. Interestingly, after reaching the full body weight bearing, the NF results were found up and down slightly until day 141 as shown in Fig 107. It could indicate that even though the weight bearing had reached full body weight, the osseointegration interface might not entirely achieve a stable level.

Patient C had a longer rehabilitation program because of the multi-fracture on femur. She commenced weight bearing rehabilitation two month after the stage 2 operation. In general, the NF of the patient C showed a more complex pattern than the other two patients as shown in Fig 111. Since the first day of weight bearing, the NF results were found up and down throughout the whole rehabilitation program. The correlation coefficient was 0.05676 ($p = 0.8408$) between the NF (except 1st measurement) and the rehabilitation time (from day 1 to day 166). There was no obvious trend found in the NF results of Patient C.

However, the strong negative correlation was found between the NF and rehabilitation time during day 79 to day 109. In other words, there was a statistically significant decrease found around that period. Indeed, on day 73, the weight bearing exercise reached the full body weight. However because of unbearable pain, the weight bearing was reduced. The decrease in the NF results could signify the vulnerable period in the rehabilitation program (Balshi 2005). Similarly, Glauser et al. (Glauser, 2004) stated that there were micro-

fractures occurring in the surrounding bone due to the loading in the first 3 months.

Conversely, since then, the correlation was found strongly relationship between the NF and rehabilitation time during day 109 to day 146. The correlation coefficient was 0.9937 ($p < 0.02$). The statistically significant increase in the NF results could explain that the adequate weight bearing exercises helped the osseointegration taking place during this period.

Finally, during day 146 to day 179, another decrease in the NF was found about 29Hz. On day 179, eventually, patient C started to use prosthetic limb with the aid of two crutches. Since then, the NF was kept in an equilibrium level of 1735Hz as shown in Fig 114. It could be a reflection that the vulnerable period was over at last and the osseointegration process had reached to the stable level.

Comparing the results of the three patients, the NF patterns varied from patient to patient. It could illustrate that the osseointegration process varied from each other. The vulnerable period was found in the NF results of patient B and patient C. And, both patient A and patient C reached their equilibrium level at the final stage of rehabilitation program. However, a common occurrence seen in all three patients was that the implant stability had been approached through the rehabilitation time went by.

7.2.3 NF and the weight bearing

In general, a common drop of the NF was observed after the first weight bearing rehabilitation exercise on all three patients. It clearly indicated that the weight bearing could bring considerably effect on the NF results. It was consistent with the research by Crismani (2006), which suggested that one possible reason for this phenomenon was a decrease of the mechanical stability of the implants because of the surrounding hard tissue relaxation. In clinical situation, the threaded implant created a lateral compression to the bone tissue

during insertion. It was likely that relaxation of inbuilt stresses between the implant and bone occurs after implant placement, which could result in a drop in the NF value (Glauser 2004).

After the initial drop of patient A and patient B, the NFs rose back while the weight bearing increased. This finding was consistent with the dental research by Crismani (2006), which concluded that a formation of new woven bone between the bony walls and the implant surface occurs predominantly as the first stage of the osseointegration process. Similarly, Balshi et al. (2005) observed that a common occurrence seen throughout their dental study was a significant decrease in stability from the time of implant placement to 30 days post surgery, followed by an increase in stability approaching the original stability level.

However, the NF results of patient C displayed a different pattern as shown in Fig 112. The NF went up and down within the range of 1716Hz to 1762Hz before the first time the full body was reached. Due to the unbearable pain, the weight bearing was reduced in the rehabilitation program. Similar as the results in the rehabilitation time above, this finding could be explained that the loading might induce micro-fracture in the surrounding bone and changed the stiffness of the interface. Likewise, Sullivan (2003) pointed out that the osseointegration interface might not be able to stand a heavy load at the two months healing following the 2nd stage operation. The initial loss of the NF after weight bearing indicated that the osseointegration interface could be weakened due to weight bearing.

Comparing the results of before and after each time weight bearing of patient B and Patient C, the average drop of patient B was 16Hz, while that of patient C was 21.5Hz. In comparison, the largest NF drop of patient B was 27Hz, whereas that of patient C was 37Hz. According to the different rehabilitation progress between the two patients, it could reasonably suggest that certain level of drop should be considered carefully, such as over 20Hz. The rehabilitation program, as a result, need to be adjusted if necessary.

7.2.4 NF and pain index

Generally speaking, there was no statistically significant relationship found between pain index and the NF results in all three patients. In another words, the pain index could not directly reflect to the dynamic situation of osseointegration interface. However, it was noticed that the pain index might be able to give support evidence to explain the changes in the NF results.

Firstly, because the pain index results of the patient A were all zero, the relationship between the NF and pain index could not be determined. However, it could be a positive sign of a stable osseointegration interface.

Secondly, the patterns of pain index patient B and C were both more complex than those from patient A. Similar phenomenon was found in the two patients that the NF started a deep drop when the pain index results increased obviously. Particularly for patient C, the results showed that the NF value decreased while the patient C felt painful; the NF value increased while she felt less painful during the weight bearing excises. Even though, the relationship between the NF and the pain index is not very strong as the correlation coefficient was -0.1887 between NF (except 1st measurement) and the pain index. ($p=0.0769$) The reason might be that the pain index was the result of patients' subjective feelings, which could cause discrepancies in the measurement.

In generally speaking, the rehabilitation time could have significant correlation with the NF. It was found typically in the NF pattern of patient A. However, in the cases of patient B and patient C, the general trend was not obvious. Even though, both weight bearing and the pain index may not reflect the osseointegration process directly. They could give support evidences to explain the reasons why the NF did not follow the typical trend in the cases of patient B and patient C.

7.3 Limitations

Noticeably, the NF results obtained in all three studies are relatively different to each other. For instance, the range of NF in the Sensitive Band of the *in vitro* study was found between 2800Hz to 3450Hz, which was higher than that of the *in vivo* study, which was between 1240Hz to 1760Hz. Understandably, the clinical situation could not be entirely simulated in the *in vitro* study due to the limitations of the experiment setup.

Firstly, the femur implant model was wrapped by a soft foam and held by a stand in the *in vitro* study. But the head of the femur model was not constrained. The model could vibrate around the axis of middle part of the femur. That would effectively reduce the length of the vibration beam significantly. On the other hand, in the *in vivo* study, the osseointegration system femur-implant was supported by the soft tissue and fixed on three-direction movement by the socket of pelvis. Also, the similar clinical constraint of the situation was simulated in the FE simulation. The femur head was fixed at one node in all direction so that the femur allowed rotation about the centre of the femur head. Consequently, these factors could cause the difference in the NF results of the *in vitro*, the FEA and the *in vivo* studies.

Furthermore, the materials used in the *in vitro* study were different to those in the *in vivo* study and FEA. For instance, the implant and the abutment used in the *in vitro* were made of stainless steel, which had doubled the elastic modulus (205GPa) than the real commercial pure titanium (115GPa). And, the plastic bone used in the *in vitro* study had a lower density than real bone. This could increase the moment inertial of the implant system. As a result, the NF results of the *in vitro* study could be affected inevitably.

Similarly, the differences in the NF results could be observed in FEA and the *in vivo* study. Firstly, the non-linear damping effect and deformation of soft tissue were not considered in FE study. Also, only two bulk properties were assigned to the soft tissue in the FE study, Young's modulus and Poisson's ratio. In fact, soft tissues are irregular geometry and complex components, with complex

material properties such as non-homogeneity, anisotropy, viscoelasticity, and time dependency. Although, it had been well proved from other researches that the complexity of soft tissues could be simplified by considering them as one uniform material in FE modelling. Undoubtedly, these differences could cause the changes of mechanical properties of the soft tissue, consequently, the NF results.

In addition, in the FE model, the bone and the implant were simulated fully connected without any gap. In clinical situation, the gap of thread contact between the implant and the bone should exist. It was assumed that the bone would have grown onto the implant and filled the gap after 8 months of recovery. Again, this could affect the results of NF in FE study.

Lastly, in the in vivo study, there were only three patients participate in the experiments. Each patient had different pattern of rehabilitation progress in terms of the length of the training program and the feeling of pain. The limit number of patients could bring difficulties in finding the similarities and contrasts.

Chapter 8 Conclusions and future work

8.1 Contribution of the research

The term of osseointegration has been used to define a direct structural and functional connection between living bone and the surface of a loading carrying implant. As reviewed in the literature, the vibration analysis had been widely proved as a useful tool to determine the implant stability and the level of osseointegration. This study was aimed to use the NF of the osseointegration trans-femoral implant system to determine the change of mechanical characteristics as a reflection of interface between the implant and the bone. To achieve this goal, FEA, *in vitro* and *in vivo* studies had been carried out.

8.1.1 Contribution of FEA study

In FEA study, several clinical conditions were simulated, such as, size of contact area, mechanical properties of the interface, femur length, and soft tissue properties. Firstly, the results demonstrated that both larger contact area and higher elastic modulus of osseointegration interface would result in higher natural frequency and vice versa. Moreover, the findings illustrated that femur length and soft tissue properties also had considerable effect on the natural frequency. In clinical situation, femur length and soft tissue properties were individually factors. Therefore, the difference of the NF results between each patient found in the *in vivo* study was expected. Finally, the results also identified that the Sensitive Band (between 1000Hz to 3500Hz) and the Over Sensitive Band (>3500Hz) could be more suitable for detecting the changes of NF while the mechanical properties of the interface changed.

8.1.2 Contribution of the *in vitro* study

In the *in vitro* study, the hardware and software of the vibration measurement were constructed and tested successfully based on the literature review and the

findings in FEA study. The experiments were carried out by using silicone interface layers to simulate implant-bone interfaces with different elastic modulus. The different length of the abutment and the different diameter of the implant were also applied to find out how the NF changed correspondingly.

Similarly as FEA study, the *in vitro* findings showed that there was a linear correlation between the NF and the elastic modulus of the interface. Meanwhile, the NF decreased linearly when the length of the abutment increased, which was constant with the Equation 27. Noticeably, the NF results showed a considerable increase against curing time while the silicone rubber hardened from liquid to solid. In other words, the finding suggested that the NF analysis could be a potential method to monitor the changes of the implant-bone interface during the rehabilitation program.

8.1.3 Contribution of *in vivo* study

At last, the *In vivo* measurements were carried out on three patients over the period of weight bearing and walking. The clinical factors, which might reflect the osseointegration situation, such as rehabilitation time, weight bearing and pain index, were assessed based on the NF results

Firstly, three patients had participated in this study and their rehabilitation time were different. A gradual NF increase was found with patients who finished the rehabilitation program within normal time. However, the standard deviation of the results were found larger than the changes in the NFs. No certain conclusion can be drawn on the relationship between the NF and the rehabilitation time.

Secondly, the NF decreases were commonly observed after the first rehabilitation weight bearing in all three patients and the NFs after weight bearing were significant lower than the NF before the weight bearing. The results, therefore, indicate that the weight bearing had a significant effect on the NF of the osseointegration system, and consequently the bone-implant

interface and the weight bearing might induce micro-fracture in the surrounding bone and then changed the stiffness of the interface. Particularly, the weight bearing caused unbearable pain and discomfort to one of the three patients whose NF results showed a drop over 20Hz each time after weight bearing. Hence, it could suggest that a longer rehabilitation programme should be taken into the consideration if the drop was found above 20 Hz.

At last, the relationship between the NF and the pain index was found not significantly strong. In other words, the pain index could not directly reflect to the dynamic situation of osseointegration interface. However, it was noticed that the pain index might be able to give support evidence to explain the irregular changes in the NF results.

In conclusion, the correlation found between the NF and mechanical characteristics, such as contact area and elastic modulus, of osseointegration interface in all of above three studies suggested that the NF analysis is an effective method to reveal the osseointegration condition during the rehabilitation. Consequently, the technique is able to help to manage and adjust the rehabilitation programme for each patient individually according to their NF results. Furthermore, this study may serve as a useful reference for future clinical studies elaborating the NF characteristic of osseointegration trans-femoral implant.

8.2 Future work

With the aim and objectives achieved in present study, the quantitative results may have limited application. The following three pieces of works are recommended for future study.

The first one is that further development and investigations are needed in terms of studying the NF analysis' performance *in vivo* and the degree of correlation with established techniques including radiography and histology. To develop this part, firstly, the entire osseointegration process should be monitored by the vibration measurement since the first stage surgery was taken. It could provide

more significant data to analyze the changes of osseointegration. Secondly, more rehabilitation patients are required to participate in the *in vivo* study so that the intro-patient comparison could be made. Finally, the results obtained by the established techniques, such as radiography, should be compared correspondingly with those gained by the NF measurement so that better understanding of the development of the osseointegration could be drawn.

Next, further improvement of the software and hardware design is needed in order to obtain more accurate data in the *in vivo* experiments. To improve this aspect, in the *in vitro* experiment, the mechanical properties of the implant and abutment should be simulated closer to the clinical situation so that it could offer more accurate prediction and better understanding for the *in vivo* experiment. Moreover, other transfer functions should be considered in order to obtain the changes of the NF in high frequency range. For instance, discrete wavelet analysis had been proved to be sensitive to the changes of the NF in the Over Sensitive Band (Qi, 2003).

The last work is that the viability and effectiveness of the FEM results need to be improved so that more systematic evaluation of osseointegration process could be drawn. To enhance this part, the more sophisticated clinical situation, such as damping and viscoelastical effect, should be considered in the simulation in order to improve the similarity of the FE model to the real structure of the osseointegration trans-femoral implant. Moreover, the thinner osseointegration interface should be considered in the simulation since in clinic situation the thickness of osseointegration is often about 10-50 μ m. It could enhance the accuracy of the study whilst it may cost longer calculation time due to the significant increased number of the element.

References

- Albrektsson, T., Jacobsson, M. (1987). "Bone-metal interface in osseointegration." Journal of Prosthetic Dentistry **57**: 597-607.
- Andrade, J. D. (1985). "Principles of protein adsorption. In: Andrade JD editor. Surface and interfacial aspects of biomedical polymers." New York: Plenum Press: 1-80.
- Ashman, R., Corin, JD., Turner, CH., (1987). "Elastic properties of cancellous bone: Measurement by an ultrasonic technique." Journal of Biomechanics **20**(10): 979-986.
- Balshi, S. F., Allen, F.d., Wolfinger, G.J., Balshi, T.J., (2005). "A resonance frequency analysis assessment of maxillary and mandibular immediately loaded implant." The International Journal of Oral & Maxillofacial implant **20**(4): 584-594.
- Barth, E., Johansson, C., and Albrektsson, T., (1990). "Histologic comparison of ceramic and titanium implants in cats." International Journal of Oral & Maxillofacial Implants **5**: 227-231.
- Bergkvist, R. "osseointegration: case report on prosthetic treatment in transfemoral amputation. In: Conference Book of 9th World Congress ISPO, Amsterdam. /edited by Soede M. -Copenhagen, ISPO." 136-137.
- Björsten, L.-M. (1991). "The bone-implant interface in osseointegration. In: Rydevik B, Branemark P-I, Skalak R, editors. International Workshop on Osseointegration in Skeletal Reconstruction and Joint Replacement." The Institute for Applied Biotechnology, Goteborg, Sweden: 25-31.
- Black, J. (1992). "Biological performance of materials." New York: Marcel Dekker.
- Block, M. S., Finger, I.M., Fontenot, M.G., and Kent, J.N., (1989). "Loaded hydroxylapatite-coated and grit-blasted titanium implants in dogs." International Journal of Oral & Maxillofacial Implants **4**: 219-225.
- Branemark, P.-I. (1959). "Vital microscopy of bone marrow in rabbit." Scand J Clin Lab Invest: suppl138.
- Branemark, P.-I. (1983). "Osseointegration and its experimental studies." J Prosthetic Dentistry **50**: 399-410.

- Branemark, P.-I. (1994). "Osseointegration in orthopaedic surgery. In Dolci, G., Favero, G. A., editors." Proceedings of First World Congress of Osseointegration, Venice: 17-18.
- Branemark, P.-I. (1999 May 2). "Introductory remarks. In: Williams E, Rydevik B, Branemark P-I, editors. Osseointegration from molecule to man." Institute for Applied Biotechnology, Proceedings of Interdisciplinary Conference; Goteborg, Sweden: Artisten.
- Branemark, P.-I., Hansson, B.O., Adell, R., Breine, U., Lindstrom, J., Hallen, O., and Ohman, A (1977). "Osseointegrated implants in the treatment of the edentulous jaw." Stockholm: Almqvist and Wiksell: 132.
- Branemark, R. (1996). "A biomechanical study of osseointegration. In-vivo measurements in rat, rabbit, dog and man." Dissertation Research, Gothenburg, Sweden; ISBN 91-628-2267-5.
- Branemark, R., Branemark, P-I., Rydevik, B., and Myers, R. R. (2001). "Osseointegration in skeletal reconstruction and rehabilitation." J Rehabil Res Dev **38**: 175-181.
- Branemark, R., Branemark, PI., Rydevik, B., Myers, RR.. (2001). "Osseointegration in skeletal reconstruction and rehabilitation." Journal of Rehabilitation Research and Development **38 (2)**: 175-81.
- Brunski, J. B. (1991). "Biomechanics of dental implants." In: Davies JE, editor. The bone-biomaterial interface. Toronto: Univerisity of Toronto Press,: 391-405.
- Carlsson, L., Rostlund, T., Albrektsson, B., and Albrektsson, T. (1988). "Removal torques for polished and rough titanium implants." International Journal of Oral and Maxillofacial Implants **3**: 21-24.
- Cawley, P., Pavlakovic, B., Alleyne, D. N., George, R., Back, T., and Meredith, N. (1998). "The design of a vibration transducer to monitor the integrity of dental implants." Proc Inst Mech Eng [1]. **212(4)**: 265-72.
- Collier, R. J., Nadav, O., and Thomas, T. G. (1982). "The mechanical resonances of a human tibia. I. In-vitro." J. Biomechanics **15**: 545.
- Cowin, S., Lou, G., (1997). "Modelling of ingrowth into implant cavities along the bone-implant interface." International Workshop on Osseointegration in Skeletal Reconstruction and Joint Replacement.: 57-68.

- Crismani, A. G., Bernhart, T., Schwart, K., Celar, A.G., Bantleon, H.P., Watzek, G., (2006). "Ninety percent success in palatal implants loaded 1 week after placement: a clinical evaluation by resonance frequency analysis." Clinical Oral Implants Research **17**: 445-450.
- Cunningham, J. L., Kenwright, J., and Kershaw, C. J. (1990). "Biomechanical measurement of fracture healing." Journal of Medical Engineering & Technology **14**: 92-101.
- Davy, D. T., Connolly, J. F. "The biomechanical behaviour of healing canine radii and ribs." J. Biomechanics **15**: 245-247.
- Dencker, H., Moberg, E. (1967). "Early diagnosis of interposition in fractures with the aid of a vibrator." Acta Orthop. Scand **38**: 396.
- Dhert, W. J., Verheyen, C.C., Braak, L.H., de Wijn, J.R., Klein, C.P., De Groot, K., and Rozing, P.M. (1992). "A finite element analysis of the push-out test: influence of test conditions." Journal of Biomedical Materials Research **26**: 119-130.
- Ericsson, I., Johansson, C.B., Bystedt, H., and Norton, M.R. (1994). "A histomorphometric evaluation of bone-to-implant contact on machine-prepared and roughened titaniumdental implants. A pilot study in the dog." Clinical Oral Implants Research **5**: 202-206.
- Ferguson, A. (1959). Historical use of metals in the human body. In "Engineering Bone and Joint Surgery". Baltimore, The williams &Wilkins Company.
- Friberg, B., Sennerby, L., Roos, J., Lekholm, U. (1995). "1995) Identification of bone quality in conjunction with insertion of titanium implant. A pilot study in jaw autopsy specimens." Clinical Oral Implants Research **6**: 213-219.
- Georgiou, A. P., Cunningham, J. L. (2001 May). "Accurate diagnosis of hip prosthesis loosening using a vibrational technique." CLIN BIOMECH. **16**(4): 315-323.
- Glauser, R., Sennerby, L., Meredith, N., Ree, A., Lundgren, A.K., Gottlow, J., Hammerle, C.H.F., (2004). "Resonance frequency analysis of implants subjected to immediate or early functional occlusal loading. Successful vs. failing implants." Clinical Oral Implants Research **15**: 428-434.

- Guzelsu, N., Saha, S. (1986). "A theoretical analysis of piezo-electric wave propagation in long bones." In Advances in Bioengineering, (ASME, New York). 61.
- Hagert, C. G., Eiken, O., Ohlsson, N.M., Aschan, W., and Movin, A. (1975). "Metacarpophalangeal joint implants. I. Roentgenographic study on the silastic finger joint implant, Swanson design." Scand J Plast Reconstr Surg **9**: 147-57.
- Ham, A. W., Harris, W. R. (1972). "Repair and transplantation of bone. The biochemistry and Physiology of Bone (Edited by Bourne R)." Academic Press, New York: 337-339.
- Hight, T. K., Piziali, R. L., and Nagel, D. A. (1980). "Natural frequency analysis of a human tibia." J. Biomechanics **13**: 139.
- Hobatho, M. C., Darmana, R., Pastor, P., Barrau, J. J., Laroze, S., and Morucci, J. P. (1991). "Development of three dimensional finite element modelling of human tibia using experimental modal analysis." J. Biomechanics **24**: 371-383.
- Holmes, D. C. L., J.T., (1997). "Influence of bone quality on stress distribution for endosseous implants." Journal of Oral Implantology **23**: 104-111.
- Huang, H. M., Chiu, C. L., Yeh, C. Y., and Lee, S. Y. (2003 Oct.). "Factors influencing the resonance frequency of dental implants." J. Oral Maxillofac Surg **61**(10): 1184-48.
- Huang, H. M., Chiu, C. L., Yeh, C. Y., Lin, C. T., Lin, L. H., and Lee, S. Y. (2003 Aug.). "Early detection of implant healing process using resonance frequency analysis." Clin Oral Implants Res. **14**(4): 437-43.
- Huang, H. M., Lee, S. Y., Yeh, C. Y., and Lin, C. T. (2002 Feb.). "Resonance frequency assessment of dental implant stability with various bone qualities: a numerical approach." Clin Oral Implants Res. **13**(1): 65-74.
- Huang, H. M., Lee, S. Y., Yeh, C. Y., Wang, M. S., Chang, W. J., and Lin, C. T. (2001 Oct.). "Natural frequency analysis of periodontal conditions in human anterior teeth." Ann Biomed Eng. **29**(10): 915-20.
- Huang, H. M., Pan, L. C., Lee, S. Y., Chiu, C. L., Fan, K. H., and Ho, K. N. (2000 Sep.). "Assessing the implant/bone interface by using natural frequency analysis." Oral Surg Oral Med Oral Pathol Oral Radiol Endod **90**(3): 285-291.

- Johansson, C. B., Sennerby, L., and Albrektsson, T. (1991). "A removal torque and histomorphometric study of bone tissue reactions to commercially pure titanium and vitallium implants." International Journal of Oral & Maxillofacial Implants 6: 437-441.
- Jurist, J. M. (1970). "In vivo determination of the elastic response of bone II ulnar resonant frequency in osteoporotic, diabetic and normal subjects." Phy. Med. Biol. 15: 427-434.
- Kaneko, T. (1987). "Assessment of the interfacial rigidity of bone implants from vibration signals." Journal of Materials in Science 22: 3495-3502.
- Kaneko, T. (1989). "Comparison between acoustic and mechanical tapping methods for assessing the inter-facial states of bone implants." Journal of Materials in Science 24: 2820-2824.
- Kaneko, T. (1991 Aug). "Pulsed oscillation technique for assessing the mechanical state of the dental implant-bone interface." BIOMATERIALS 12(6): 555-560.
- Kasemo, B., Lausmaa, J. (1985). Tissue integrated prostheses. Osseointegration in clinical dentistry. Chicago, Quintessence Publishing Co., Inc.,.
- Khalil, T. B., Viano, D. C. (1982). "Generalized resonances of the isolated femur." No. GMR 3560R, Biomedical Sciences Department, General Motors Research Laboratories, Warren, Michigan, 26 March.
- Khalil, T. B., Viano, D. C., and Taber, L. A. (1980). "Vibrational characteristics of the embalmed human femur." In Advances in Bioengineering, (ASME, New York). 57.
- Lee, S. Y., Huang, H. M., Lin, C. Y., and Shih, Y. H. (2000 April). "In vivo and in vitro natural frequency analysis of periodontal conditions: an innovative method." J. Periodontol 71(4): 632-40.
- Lekholm, U., and Zarb, G.A., (1985). "Patient selection and preparation." In: Branemark, P.I., Zarb, G.A. & Albrektsson, T., eds. Tissue-integrated prostheses: osseointegration in clinical denisity.(Chicago: Quintessence Publishing): 199-200.
- Lewis, J. L., Tarr, R. R. (1975). "Fracture healing assessment by impact response." J. bone Jt Surg 57a: 576.

- Li, P. L. S., Jones, N. B., and Gregg, P. J. (1995 July). "Loosening of total hip arthroplasty-diagnosis by vibration analysis." J Bone Joint Surg BR 77b(4): 640-644.
- Li, P. L. S., Jones, N. B., and Gregg, P. J. (1996). "Vibration analysis in the detection of total hip prosthetic loosening." MED. ENG. PHYS. 18(7): 596-600.
- Lippmann, R. K. (1932). "The use of auscultatory percussion for the examination of fractures." J. Bone Jt Surg 14: 118.
- Lundborg, G. (1997). "Osseointegrated metacarpophalangeal joint prostheses in the hand. In: Branemark P-I, Rydevik BL, Skalak R, editors. Osseointegration in skeletal reconstruction and joint replacement." Carol Stream, IL: Quintessence Publishing Co: 167-74.
- McLean, J. (1994). "The evolution of glass-ionomercements:- A personal View." 2nd International Symposium on Glass-Ionomers.
- Meredith, N. (1994). "The application of modal vibration analysis to study bone healing in vivo." J Dent Res. 73(4): 793.
- Meredith, N. (1998). "Assessment of implant stability as a prognostic determinant." International Journal of Prosthodontics 11: 491-501.
- Meredith, N. (1998). "Assessment of implant stability as a prognostic determinant." Int. J. Prosthodont 11(5): 491-501.
- Meredith, N., Alleyne, D., and Cawley, P. (1996). "Quantitative determination of the stability of the implant-tissue interface using resonance frequency analysis." Clin Oral Implants Res. 7: 261-267.
- Meredith, N., Book, K., Friberg, B., Jemt, T., and Sennerby, L. (1997). "Resonance frequency measurements of implant stability in vivo. A cross-sectional and longitudinal study of resonance frequency measurements on implants in the edentulous and partially dentate maxilla." Clin Oral Implants Res. 8: 226-233.
- Meredith, N., Sennerby, F., Alleyne, D., Sennerby, L., and Cawley, P. (1997). "The application of resonance frequency measurements to study the stability of titanium implants during healing in the rabbit tibia." Clin Oral Impl Res. 8: 234-243.

- Nakatsuchi, Y., Tsuchikane, A., and Nomura, A. (1996). "Assessment of fracture healing in the tibia using the impulse response method." J. Orthop Trauma **10**(1): 50-62.
- Nakatsuchi, Y., Tsuchikane, A., and Nomura, A. (1996). "The vibrational mode of the tibia and assessment of bone union in experimental fracture healing using the impulse response method." Med. Eng. Phy. **18**(7): 575-583.
- Nanci, A., Mckee, M.D., Zalzal, S., (1998). "Ultrastructural and immunocytochemical analysis of the tissue response to metal implants in the rat tibia." In: Davidovitch Z, Mah J, editors. Biological mechanisms of tooth eruption , resorption and replacementby implant. Boston: Harvard Society for the Advancement of Orthodontics,(487-500).
- Natali, A. N., Pavan, P.G., Schileo, E., Williams, K.R., (2006). "A numerical approach to resonance frequency analysis for the investigation of oral implant osseointegration." Journal of Oral Rehabilitation **33**: 674-681.
- Nikiforidis, G., Bezerianos, A., Dimarogonast, A., and Sutherland, C. (1990). "Dimarogonast A, and Sutherland C. Monitoring of fracture healing by lateral and axial vibration analysis." J. Biomechanics **23**(4): 323-330.
- Nishimura, T. (1991). "Processing and applications of titanium and titanium alloys." Titanium-From Mining to Biomaterials: 9-20.
- Nokes, L., Mintowtczyz, W. J., Fairclough, J. A., Mackie, I., and Williams, J. (1985). "Vibration analysis in the assessment of conservatively managed tibial fractures." Journal of Biomedical Engineering **7**(1): 40-44.
- Nokes, L. D. M., Mintowl-Czyz, W. J., Fairclough, J. A., Mackie, I., Howard, C., and Williams, J. (1984). "Natural frequency of fracture fragments in the assessment of tibial fracture healing." J. Biomed Eng **6**: 227-229.
- Noyes, D. H., Solt, C. W. (1973). "Measurement of mechanical mobility of human incisors with sinusoidal forces." J. biomech. **6**: 439-442.
- Papavasiliou, G., Kamposiora, P., Bayne, S.C., and Felton, D.A., (1997). "3D-FEA of osseointegration percentages and patterns on implant-bone interfacial stresses." Journal of dentistry **25**(6): 485-491.

- Park, J. B. (1980). "Polymeric materials. Biomaterials,." Plenum Press, New York: 86.
- Perren, S. M. (1979). "Physical and biological aspects of fracture healing with special reference to internal fixation." Clin. Othop. Rel. Res **138**: 175-195.
- Peters, A. (1982). "Monitoring of the rate of fracture healing by acoustic percussion." Presented at the Second International Symposium on Internal Fixation of Fractures, Lyon, France.
- Prendergast, P., and Taylor, D., (1994). "Prediction of bone adaptation using damage accumulation." Journal of Biomechanics **27**: 1067-1076.
- Pugh, J. W., Streitman, A. I., and Stanoch, J. F. (1975). "Studies of the impact response of human limbs." 28th A.G.E.M.B. Meeting New Orleans, USA: 171.
- Puleo, D. A., Nanci, A., (1999). "Understanding and controlling the bone-implant interface." BIOMATERIALS **20**: 2311-2321.
- Qi, G., Mouchon, W. P., and Tan, T. E. (2003 June). "How much can a vibrational diagnostic tool reveal in total hip arthroplasty loosening?" CLINICAL BIOMECHANICS **18**(5): 444-458.
- Rasmusson, L., Meredith, N., Cho, I. H., and Sennerby, L. (1999). "The influence of simultaneous versus delayed placement on the stability of titanium implants in onlay bone grafts." Int. J. Oral Maxillofac Implants **28**: 224-231.
- Roberts, W. E. (1988). "Bone tissue interface." Journal of Dental Education **52**: 804-809.
- Sennerby, L., Thomsen, P., and Ericson, L.E. (1992a). "A morphometric and biomechanic comparison of titanium implants inserted in rabbit cortical and cancellous bone." International Journal of Oral and Maxillofacial Implants **7**: 62-71.
- Sennerby, L., Thomsen, P., and Ericson, L.E. (1992b). "Ultrastructure of the bone-titanium interface in rabbits." Journal of Materials Science - Materials in Medicin **3**: 262-271.
- Shirazi-Adl, A. (1992). "Finite element stress analysis of push-out test. Part 1: Fixed interface using stress compatible elements." Journal of Biomechanical Engineering **114**: 111-118.

- Silver-Thorn, M., Childress, D.S. (1992). "Use of a generic, geometric finite element model of the below-knee residual limb and prosthetic socket to predict interface pressure." In: Proceedings of the 7th World Congress of ISPO: 272.
- Singh, V. R., Yadav, S., and Adya, V. P. (1989). "Role of natural frequency of bone as a guide for detection of bone fracture healing." J. Biomed. Eng. **11**: 457-461.
- Skalak, R. (1991). "Overview of previous development and biomechanics of osseointegration. In: Rydevik B, Branemark P-I, Skalak R, editors." International Workshop on Osseointegration in Skeletal Reconstruction and Joint Replacement. The Institute for Applied Biotechnology, Goteborg, Sweden.
- Skalak, R. (1997). "Biomechanics of osseointegration. In: Branemark P-I, Rydevik BL, Skalak R, editors. Osseointegration in skeletal reconstruction and joint replacement." Carol Stream, IL: Quintessence Publishing Co: 45-56.
- Southam, J. C., Selwyn, R. (1971). "Structural changes around screws used in the treatment of fractured human mandibles." Br J Oral Surg **8**: 211-21.
- Steege, J., Childress, D.S., (1988). "Finite element modelling of the below-knee socket and limb: phase 2." In: Modeling and control issues in biomechanical systems, ASME: 121-130.
- Sullivan, J., Uden, M., Robinson, K. P., and Sooriakumaran, S. (2003). "Rehabilitation of the trans-femoral amputee with an osseointegrated prosthesis: the United Kingdom experience." Presthetics and Othotics International **27**: 114-120.
- Sundgren, J. E., Bodo, P., and Lundstrom, I. (1986). "Auger electronspectroscopic studies of the interface between human tissue and implants of titanium and stainless steel." J Coll Int Sci **110**: 9-20.
- Swanson, A. B. (1972). "Flexible implant arthroplasty for arthritic finger joints: Rationale, technique and results of treatment." J Bone Joint Surg **54A**: 435-55.
- Thompson, G. A., Young, D. R. and Orne, D. (1976). "In-vivo determinations of mechanical properties of the human ulna by means of mechanical

- impedance tests: experimental result and improved mathematical models." Med. Biol. Engng, **14**: 253.
- Thomsen, J. J. (1990). "Modelling human tibia structural vibrations." J. Biomechanics **23**: 215-228.
- Tjellstrom, A., Jansson, K., and Branemark, P-I. (1992). "Craniofacial defects. In: Worthington, P., Branemark, P-I., editors. Advanced Osseointegration Surgery: Applications in the Maxillofacial Region." Chicago: Quintessence: 293-312.
- Torres-Moreno, R., Solomonidis, S. E., Jones, D (1992). "Three-dimensional finite element analysis of the above-knee residual limb." In: Proceedings of the 7th World Congress of ISPO **274**.
- VanBuskirk, W., Ashman, RB., (1981). Elastic Moduli of Bone, Applied Mechanics Division, AMD. New York.
- Vannah, W. M. (1990). "Indentor tests and finite element modelling of bulk muscular tissue in vivo." Dissertation. Chicago: North-Western University.
- Wennerberg, A., Albrektsson, T., Andersson, B., and Krol, J.J. (1995). "A histomorphometric and removal torque study of screw-shaped titanium implants with three different surface topographies." Clinical Oral Implants Research **6**: 24-30.
- Williams, K. R. and A. D. C. Williams (1997). "Impulse response of a dental implant in bone by numerical analysis." Biomaterials **18**(10): 715-719.
- Wong, A. T. C., Goldsmith, W. and Saebman, J. L. (1976). "Flexural wave propagation in discontinuous models and 'in-vivo' tibia." J. Biomechanics **9**: 689.
- Worthington, P. (1991). "Current status of osseointegration. In: Rydevik B, Branemark P-I, Skalak R, editors. International Workshop on Osseointegration in Skeletal Reconstruction and Joint Replacement." The Institute for Applied Biotechnology, Goteborg, Sweden: 17-22.
- Worthington, P. (1997). "History, development, and current status of osseointegration as revealed by experience in craniomaxillofacial surgery. In: Branemark P-I, Rydevik BL, Skalak R, editors. Osseointegration in skeletal reconstruction and joint replacement." Carol Stream, IL: Quintessence Publishing Co: 25-44.

- Xu, W., Crocombe, AD., Hughes, SC., (2000). "Finite element analysis of bone stress and strain around a distal osseointegrated implant for prosthetic limb attachment." Proc Inst Mech Eng H. **214**: 595-602.
- Yamada, H. "Strength of Biological Materials (Edited by Evans, FG)." Robert E. Krieger, Huntington, NY.: 73-74, 83-87.
- Yamagishi, M., Yoshimura, Y. (1955). "The biomechanics of fracture healing." J. Bone Jt Surg **37a**: 1035-1066.
- Ysander, M., Branemark, R., Olmarker, K., and Myers, R. R. (2001). "Intramedullary osseointegration. Development of a rodent model and study of histology and neuropeptide changes around titanium implants." J Rehabil Res Dev **38**: XXX-XXX.

Appendix 1

The velocity of the steel ball just before hitting the abutment could be calculated from energy equation:

$$E_k = E_p$$

$$\frac{1}{2}mv^2 = mgh$$

$$v = \sqrt{2gh}$$

An impulse is defined as the integral of a force with respect to time:

$$I = \int F dt$$

An impulse may also be regarded as the change in momentum of an object to which a force is applied. The impulse may be expressed in a simpler form when both the force and the mass are constant:

$$I = F\Delta t = m\Delta v$$

where

Δv is the change in velocity produced by the force in the considered time interval

Therefore, in this study the impulse can be calculated as:

$$I = m\Delta v = m\sqrt{2gh} = 0.01\text{kg}\sqrt{2 \times 9.8\text{m/s}^2 \times 0.05\text{m}} \approx 0.01\text{Ns}$$

Appendix 2

Proposal of the clinical study of using resonant frequency to determine the degree of osseointegration of the trans-femoral implant

Introduction

Mechanical vibration analysis, a traditional non-destructive testing tool, has been widely used by mechanical engineers to inspect for structural integrity. Now it is extensively used in biomedical research, including the determination of bone density and the stability of orthopedic implant, the assessment of bone healing after fracture, and diagnosing total hip arthroplasty loosening.

Osseointegrated trans-femoral implant is a new orthopedic anchoring method to attach prosthetic limbs. The aim of this clinical study is using vibration measurement to determine the degree of osseointegration of the trans-femoral implant. The degree of osseointegration is a key factor, which dominates the clinical successful rate and rehabilitation process of the amputee who uses trans-femoral implant for limb attachment. An early detection of the degree of the osseointegration, right after the second surgery and throughout the rehabilitation stage, could provide quantitative and consistent data to assist rehabilitation planning and prevent early failure of the implant.

In previous research the vibration measurement was applied successfully in several osseointegration areas. In 1998, Dr N Meredith took vibration measurement on 9 patients who had a total of 56 dental implants placed in the maxilla. [1] The resonance frequency of the implant transducer system increased for 50 of the 56 implants from a mean value of 6473 Hz to a mean value of 7915 Hz. Two implants had failed to integrate, and the resonance frequency of these had fallen. These results suggest that NF analysis is an effective method for assessing the implant-bone integration. In 2001, in the study of the natural frequency (NF) of the central incisors, Dr Huang found the NF at 1.24 ± 0.11 kHz that the periodontal disease is significantly lower than that of teeth in a healthy condition (1.34 ± 0.20 kHz; $p < 0.01$) [2]. In 2001, AP Georgiou used this technique on patients and compared vibration-testing results with radiographs in 23 patients with total hip replacements experiencing hip pain without X-ray evidence of loosening. The authors concluded that the vibration testing method was more sensitive than the traditional X-ray. [3]

In-vitro experiment

From a mechanical standpoint, an implant that is free to vibrate at one end and fixed at the other is a simple cantilever. The resonance frequency (f) of such a beam can be expressed as equation (1):

$$f = \alpha \sqrt{\frac{EI}{\rho l^4}} \quad (1)$$

Where f = Natural Frequency (Hz); l =effective length of beam (mm); E = Young's modulus (GNm⁻²); I =moment of inertia; ρ = the vibration mass per unit 1 and α is a constant related to the boundary conditions. [12,13]

It is anticipated that if the length of the abutment remained unchanged the natural frequency (NF) of the system would vary with the length of abutment and the surrounding bony tissue.

Experiments were carried out to place a metallic implant in a plastic femur model. The plastic femora models (Sawbones Europe AB) were used for all of the experiments because the principles of the experiments of vibration of a rod such as the femur would be the same for real bone as for an artificial material. Even though the mass, density and stiffness of real bone are quite different from this model; these differences should be reflected only in the absolute value of the resonant frequencies. We have been more concerned with changes in the vibration signals than in their absolute values. [5]

The abutment was triggered to vibrate with an impact force, which was applied by dropping a thread-attached nut. The similar method was used by L Nokes in the assessment of tibia fracture. In his study a standard impulse of 0.045Ns was applied to the tibial tubercle by dropping onto a 28g steel ball down a Perspex tube[4]. In this study the input impact force was calculated to be less than 0.01Ns by momentum theory. The output vibrational signals were observed and recorded by piezo-electric accelerometer (MTN/1800 Modal Analysis Accelerometer, Monitran LTD), which was attached on the top of the abutment by two-sided tape. The vibrational signal was converted to digital signal by A/D data acquisition board (iNET 100external A/D box, instruNet). These digital data was transferred to a personal computer through a PCMCIA interface card (Model 230 PCMCIA instruNet controller, instruNet) and saved as Excel files. In addition Fast Fourier Transform was calculated with personal software, which was programmed with LabView 5.01. The response signal spectrum was then translated from a time domain to a frequency domain.

Safety

As explained in the above section, the NF analysis is noninvasive and nondestructive technique. The same method has been used for dental patients [11] Comparing to the dental implant the Osseointegrated trans-femoral implant is expect to stands more vigorous load condition and has virtually no problem to the NF testing. To make sure that the test can be safely carried out, the time of testing is selected at after the second surgery. This allows at least 6 months osseointegration period to safeguard the implant system. To make sure the measurement can be done safely, the first test is suggested to be carried out after 5% body weight load exercises are completed. Less than 0.33Ns excitation force will be applied to the abutment, which is much less than 5% body weight. 4 12V-batteries and 4 1.2V-batteries will be used to provide all the power supply of the equipments during these measurements for electrical safety reason. The specifications of the equipment used for the measurement,

including an accelerometer, a laptop computer, an excitation instrument, batteries, are given in the equipments section.

Proposed in-vivo experiment procedure

Refer to the previous study [10], It is suggested that the measurement will be carried out twice a month throughout the load bearing rehabilitation period.

The procedure of the test is given as following

1. Stick the accelerometer on the abutment with two-side tape.
2. Excite the abutment with the dropping ball.
3. Run software to record the signal from accelerometer to the laptop computer.
4. Repeat the excitation and data recording 5 more times.
5. Take the accelerometer off.

The total measurement process will take about 30 minutes to one hour.

Reference

- [1] Meredith N; Assessment of implant stability as a prognostic determinant. *The international journal of prosthodontics*; 11:5 491-501. 1998.
- [2] Huang HM, Lee SY, Yeh CY, Wang MS, Chang WJ and Lin CT. Natural frequency analysis of periodontal conditions in human anterior teeth. *Biomedical engineering Society*. 29: 915-920, 2001.
- [3] Georgiou AP, Cunningham JL. Accurate diagnosis of hip prosthesis loosening using a vibrational technique. *Clinical biomechanics*, 16: 315-323, 2001.
- [4] Nokes L, Mintowt-Czyz WJ, Fairclough JA, Machie I, Williams J. Vibration analysis in the assessment of conservatively managed tibial fracture. *J. Biomed Eng*. 7: 40-44, 1985.
- [5] Li PLS, Jones NB, Gregg PJ. Loosening of total hip arthroplasty diagnosis by vibration analysis. *The journal of bone and joint surgery*, 77-B: 640-644, 1994.
- [6] Johansson CB, Sennerby L, Albrektsson T, A removal torque and histomorphometric study of bone reaction to commercially pure titanium and vitallium implant. *Int J Oral Maxillofac Implants*, 6:437-41, 1991.
- [7] Carlsson L, Rostlund T. Removal torques for polished and rough titanium implants. *Int J Oral Maxillofac Implants* 1988; 3:21-4.
- [8] Wennerberg A, Albrektsson T, Albrektsson B, Kroll JJ. A histomorphometric and removal torque study screw-shaped titanium implants with three different surface topographies. *Clin Oral Implants Res* 1996; 6:24-30.
- [9] Lee SY, Huang HM, and Lin CY. In vivo and in vitro natural frequency analysis of periodontal conditions, an innovative method. *J. Periodontal* 74: 632-640, 2000.
- [10] Meredith N; Assessment of implant stability as a prognostic determinant. *The international journal of prosthodontics*; 11:5 491-501. 1998.
- [11] Meredith N, Alleyne D, Cawley P. Quantitive determination of the stability of the implant-tissue interface using resonance frequency analysis. *Clinical oral implants research*. 7: 261-267; 1996.

[12] Thomson WT. Free vibration. In: Thomson WT, ed. Theory of vibration with applications, 3rd edition, p.17. New York: Prentice-Hall Inc.

[13] Meredith N, Alleyne D, and Cawley P. 1996a. Quantitative determination of the stability of the implant-tissue interface using resonance frequency analysis. *Clinical Oral Implants Research*. 7:261-267.

Appendix 3

Wandsworth

Primary Care Trust

REFERRERS INFORMATION SHEET

Detecting the Degree of Osseointegrated Trans-Femoral Implant by Vibration Measurement

Background

Mechanical vibration analysis, a traditional non-destructive testing tool, has been widely used by mechanical engineers to inspect for structural integrity. Now it is extensively used in biomedical research, including the determination of bone density and the stability of orthopedic implant, the assessment of bone healing after fracture, and diagnosing total hip arthroplasty loosening.

Osseointegrated trans-femoral implant is a new orthopedic anchoring method to attach prosthetic limbs. The aim of this clinical study is using vibration measurement to determine the degree of osseointegration of the trans-femoral implant. The degree of osseointegration is a key factor, which dominates the clinical successful rate and rehabilitation process of the amputee who uses trans-femoral implant for limb attachment. An early detection of the degree of the osseointegration, right after the second surgery and throughout the rehabilitation stage, could provide quantitative and consistent data to assist rehabilitation planning and prevent early failure of the implant. In this study we propose to investigate the relationship between the natural frequency of abutment and the stability of the implant.

Aims

To determine the normal natural frequency range of the abutment from the osseointegrated trans-femoral implant patients.

To find out the variable natural frequency of the abutment during the rehabilitation program.

To find out the relationship between the natural frequency of the abutment and the osseointegration degree in the interface between the implant and the bone. In this study the vibration measurement will be applied on 11 osseointegrated trans-femoral implant patients. For each patient we will do the measurement after the abutment being tightened with torque wrench. The project is funded by Royal Society Research Grant. This funding body is supporting Natural Science equipment, consumables. The project has been reviewed by Royal Society.

The procedure of the test is given as following:

Stick the accelerometer on the abutment with two-side tape.

Measurement the effective length of the abutment.

Excite the abutment with the dropping nut.

Run software to record the signal from accelerometer to the laptop computer.

Repeat the excitation and data recording 11 more times.

Take the accelerometer off.

The whole process is not more than 30munites.

Project Investigators

Dr David Ewins, Queen Mary's Hospital and University of Surrey

Mr. Fei Shao, University of Surrey

Dr Wei Xu, Queen Mary's Hospital and University of Surrey

Contact for Further Information

Mr. Fei Shao, PhD student

Centre for Biomedical Engineering

University of Surrey

GUILDFORD

Surrey, UK

GU2 7XH

Tel: 01483 682981

Fax: 01483 306039

Email: f.shao@surrey.ac.uk

Appendix 4

Wandsworth

Primary Care Trust

Gait Laboratory
 Clinical Biomedical Engineering Centre (CBEC)
 Roehampton Rehabilitation Centre
 Queen Mary's Hospital
 London SW15 5PN

Tel: 020 8355 2175
 Fax: 020 8355 2952
 email: gaitlab@swlct.sthames.nhs.uk

CONSENT FORM

Title of Project: Detecting the Degree of Osseointegrated Trans-Femoral
 Implant by Vibration Measurement

Name of Researcher: Mr Fei Shao

- I the undersigned voluntarily agree to take part in the above study.
- I have read and understood the Information Sheet provided. I have been given a full explanation by the investigators of the nature, purpose, location and likely duration of the study, and of what I will be expected to do. I have been given the opportunity to ask questions on all aspects of the study and have understood the advice and information given as a result.
- I understand that sections of any of my medical notes may be looked at by responsible individuals from Queen Mary's Hospital or from regulatory authorities where it is relevant to them taking part in research. I give permission for these individuals to have access to their records.
- I understand that I am free to withdraw from the study at any time without needing to justify my decision and without prejudice.
- I understand that I have been asked if the photographs taken /videos recorded of my child can be used for teaching purposes and I do/do not* give my permission for this.
- I confirm that I have read and understood the above and freely consent to participating in this study. I have been given adequate time to consider my participation and agree to comply with the instructions and restrictions of the study.

* delete as appropriate

 Name of Parent/Carer Date Signature

 Researcher Date Signature

 Witness name Date Signature

1 copy for patient; 1 for researcher; 1 to be kept with medical records

Appendix 5

~ Calibration Certificate ~

Per ISO 16063-21

Model Number: 352A21Serial Number: 46863Description: ICP® AccelerometerMethod: Back-to-Back Comparison CalibrationManufacturer: PCB

Calibration Data

Sensitivity @ 100.0 Hz	9.12 mV/g	Output Bias	8.3 VDC
	(0.930 mV/m/s ²)	Transverse Sensitivity	1.3 %
Discharge Time Constant	2.2 seconds	Resonant Frequency	\$6.5 kHz

Sensitivity

Data Points

Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)
10.0	-1.2	300.0	0.1	7000.0	1.2
15.0	-0.5	500.0	0.2	10000.0	2.5
30.0	-0.4	1000.0	0.2		
50.0	-0.0	3000.0	-0.7		
REF. FREQ.	0.0	5000.0	0.9		

Mounting Surface: Twignon Adapter Fastener: Cyanacrylate Adhesive

Prober Orientation: VerticalAcceleration Level: 100 g (981 m/s²)

*The acceleration level may be limited by shaker displacement at low frequencies. If the limit level cannot be obtained, the calibration system used the following limits to set the vibration amplitude: Acceleration Level (g) = 0.010 + (3 sec)²

Condition of Unit

As Found: n/aAs Left: New Unit, In Tolerance

Notes

1. Calibration is NIST Traceable thru Project 822/267400 and PTB Traceable thru Project 1055.
2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc.
3. Calibration is performed in compliance with ISO 9001, ISO 10012-1, ANSI/NCCL Z540-1-1994 and ISO 17025.
4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%.

Technician: Thomas WitnauerDate: 01/12/04

PAGE 1 of 1

PCB PIEZOTRONICS
VIBRATION DIVISION

3425 Walden Avenue Depew, NY 14043
TEL: 885-634-0913 FAX: 716-685-3886 www.pcb.com

cat # 31557E145R 06

~ Calibration Certificate ~

Per ISO 18063-21

Model Number: 356A24

Serial Number: 31202 (x axis)

Description: ICP® Triaxial Accelerometer

Method: Back-to-Back Comparison Calibration

Manufacturer: PCB

Calibration Data

Sensitivity @ 100.0 Hz9.83 mV/g

Output Bias8.3 VDC

(1.003 mV/m/s²)

Transverse Sensitivity4.5 %

Discharge Time Constant1.3 seconds

Sensitivity

Temperature: 70 °F (21 °C)

Relative Humidity: 24 %

dB

Hz

10.0

100.0

1000.0

8000.0

Data Points

Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)
10.0	-0.4	300.0	0.1	7000.0	1.7
15.0	-0.5	500.0	0.1	9000.0	3.0
30.0	-0.1	1000.0	0.2		
50.0	0.0	3000.0	0.3		
REF. FREQ.	0.0	5000.0	1.0		

Mounting Surface: Stainless Steel

Fastener: Cyanacrylate Adhesive

Acceleration Level (max): 10.0 g (98.1 m/s²)

Future Orientation: Inverted Vertical

The acceleration level may be limited by shock or displacement at low frequencies. If the limit level cannot be obtained, the calibration system uses the following formula to set the vibration amplitude: Acceleration Level (g) = 0.010 x (Hz)².

The gravitational constant used for calculations by the calibration system is: 1 g = 9.8066 m/s².

Condition of Unit

As Found: n/a

As Left: New Unit. In Tolerance

Notes

1. Calibration is NIST Traceable thru Project 822/267400 and PTB Traceable thru Project 1055.

2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc.

3. Calibration is performed in compliance with ISO 9001, ISO 10012-1, ANSI/NCSL Z540-1-1994 and ISO 17025.

4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.

5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%.

Technician: Michael Ruchalski

Date: 02/24/04

PCB

ACCREDITED

Cert No 1802.01

PCB PIEZOTRONICS™

VIBRATION DIVISION

3425 Walden AvenueDepew, NY 14043

TEL: 838-684-0013 • FAX: 716-685-3886 • www.pcb.com

PAGE 1 of 1

call • 3150432131.64

181

F Shao

~ Calibration Certificate ~

Per ISO 16063-21

Model Number: 356A24
Serial Number: 31202 (y axis)
Description: ICP® Triaxial Accelerometer Method: Back-to-Back Comparison Calibration
Manufacturer: PCB

Calibration Data

Sensitivity @ 100.0 Hz 10.15 mV/g Output Bias 8.4 VDC
(1.035 mV/m/s²) Transverse Sensitivity 4.8 %
Discharge Time Constant 1.2 seconds

Sensitivity

Data Points

Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)
10.0	-0.6	300.0	0.2	7000.0	2.3
15.0	-0.7	500.0	0.4	9000.0	3.7
30.0	-0.2	1000.0	0.3		
50.0	0.0	3000.0	0.5		
REF. FREQ.	0.0	5000.0	1.2		

Mounting Surface: Stainless Steel Fastener: Cynacrylate Adhesive
Acceleration Level (ms²): 100 g (98.1 ms²)
*The acceleration level may be limited by shaker displacement at low frequencies. If the rated level cannot be obtained, the calibration system uses the following formula to set the vibration amplitude: Acceleration Level (g) = 0.010 x (Hz)^{1.5}
*The gravitational constant used for calculations by the calibration system is: 1 g = 9.80665 ms⁻²

Fixture Orientation: Inverted Vertical

Condition of Unit

As Found: n/a
As Left: New Unit, In Tolerance

Notes

1. Calibration is NIST Traceable thru Project 822/267400 and PTB Traceable thru Project 1055.
2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc.
3. Calibration is performed in compliance with ISO 9001, ISO 10012-1, ANSI/NCCL Z540-1-1994 and ISO 17025.
4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%.

Technician: Michael Ruchalski M.R. Date: 02/24/04

3425 Walden Avenue Depew, NY 14043
TEL: 888-684-0013 FAX: 716-685-3886 www.pcb.com

~ Calibration Certificate ~

Per ISO 16063-21

Model Number: 356A24Serial Number: 31202 (z axis)Description: ICP® Triaxial AccelerometerMethod: Back-to-Back Comparison CalibrationManufacturer: PCB

Calibration Data

Sensitivity @ 100.0 Hz 10.01 mV/g Output Bias 8.5 VDC
 (1.021 mV/(m/s²)) Transverse Sensitivity 4.4 %
 Discharge Time Constant 1.3 seconds

Sensitivity

Data Points

Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)
10.0	-0.5	300.0	0.1	7000.0	2.8
15.0	-0.4	500.0	0.2	9000.0	4.5
30.0	-0.1	1000.0	0.4		
50.0	-0.0	3000.0	1.1		
REF. FREQ.	0.0	5000.0	1.3		

Mounting Surface: Stainless Steel Fastener: Cyanoacrylate Adhesive
 Acceleration Level (msy): 100 g (98.1 m/s²)

Fixture Orientation: Vertical

*The acceleration level may be limited by shock or displacement at low frequencies. If the kinetic level cannot be obtained, the calibration system uses the following formula to set the vibration amplitude: Acceleration Level (g) = 0.010 × (Hz)^{0.5}.
 *The gravitational constant used for calculations by the calibration system is: 1 g = 9.80665 m/s².

Condition of Unit

As Found: n/aAs Left: New Unit, In Tolerance

Notes

1. Calibration is NIST Traceable thru Project 822/267400 and PTB Traceable thru Project 1055.
2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc.
3. Calibration is performed in compliance with ISO 9001, ISO 10012-1, ANSI/NCSL Z540-1-1994 and ISO 17025.
4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%.

Technician: Michael Ruchalski MR Date: 02/24/04

PCB PIEZOTRONICS
 VIBRATION DIVISION

3425 Walden Avenue Depew, NY 14043
 TEL: 888-684-0013 FAX: 716-685-3886 www.pcb.com

PAGE 1 of 1

call - 31424810632

~Certificate of Calibration~

Model Number: 482A22	N.I.S.T Project #: 6720012
Serial Number: 1750	Calibration Date:02/16/2004
Description: Signal Conditioner	Recalibration Date:
Test Procedure:AT-104-18	Calibration Technician:Cynthia Niedzwiecki <i>CN</i>
Temperature: 68°F 20°C	Relative Humidity: 8%

PASS

Channel	Volts	Current	Gain X1*	Gain X10*	Gain X100*
1	25.7	3.96	1.00	N/A	N/A
2	25.8	3.97	1.00	N/A	N/A
3	25.8	3.90	1.00	N/A	N/A
4	25.8	3.98	1.00	N/A	N/A

As Received: N/A

As Left: New unit in tolerance.

Special Notes:

This document certifies that the equipment referenced above meets published specifications. The calibration procedure is in compliance with ISO 10012-1, and former MIL-STD-45662A and is traceable to NIST. *Measurement uncertainty (95% confidence level w/coverage factor of 2) for scale factors is +/- 0.2%.

This certificate may not be reproduced, except in full, without written approval of PCB Piezotronics, Inc.

Calibration Station 29
3425 Walden Avenue Depew, New York, USA 14043-2495

For any questions concerning this certificate, please call PCB at (716) 654-0001 and ask for an application engineer.
3159808920.58

Appendix 6

TRANSFEMORAL OSSEOINTEGRATION

LOADING PROGRAMME WITH THE SHORT TRAINING PROSTHESIS (STP)

NAME: David Cox

D.O.B: 12/10/54

DATE OF STAGE (2) SURGERY: 24/3/04

BODY WEIGHT (WITHOUT PROSTHESIS): 94.6kg

DATE	24/5/04		25/5/04		26/5/04		27/5/04		28-5-04	29-5-04		30-5-04	
AM or PM	am	pm	am	pm	am	pm	am	pm		AM	PM	AM	PM
LOADING WEIGHT	20kg 15 mins	20kg 15 mins	20kg 20 mins	20kg 20 mins	20kg 25 mins	20kg 25 mins	20kg 30 mins	20kg 30 mins		20kg 30 mins	20kg 30 mins	20kg 30 mins	20kg 30 mins
TIME LOADING	0	0	0	0	0	0	0	0		0	0	0	0
PAIN SCALE 0-10 'BEFORE'	0	0	0	0	0	0	0	0		0	0	0	0
'DURING'	0	0	0	0	0	0	0	0		0	0	0	0
'LIFTING OFF'	0	0	0	0	0	0	0	0		0	0	0	0
COMMENTS	'back, achy, less than am.	back ache - less than am.	slight back ache - better than am.	slight back ache - better than am.	V.V. slight back ache.		'back better, Nr. 002			Back BETTER.			

KEY: PAIN SCALE (0-10) 0 'ok' 1-2 'allowed' 3-5 'not allowed' - seek advice

JAN. 2003

Wandsworth
Primary Care Trust

TRANSFEMORAL OSSEointegration

LOADING PROGRAMME WITH THE SHORT TRAINING PROSTHESIS (STP)

NAME: Steven Singh

D.O.B: 5/5/78

DATE OF STAGE (2) SURGERY: 26/5/04

BODY WEIGHT (WITHOUT PROSTHESIS): 78kg

DATE	12/7/04		13/7/04		14/7/04		15/7/04		16/7/04		17/7/04		18/7/04	
AM or PM	am	pm	am	pm			am	pm	am	pm	am	pm	am	pm
LOADING WEIGHT	20kg	20kg	20kg	20kg	/		20kg	20kg	20kg	20kg	20kg	20kg	20kg	20kg
TIME LOADING	15 mins	15 mins	20 mins	20 mins			25 mins	25 mins	30 mins	30 mins	30 mins	30 mins	30 mins	30 mins
PAIN SCALE 0-10	0	0	0	0			0	0	0	0	0	0	0	0
'BEFORE'	0	0	0	0			0	0	0	0	0	0	0	0
'DURING'	0	0	0	0			0	0	0	0	0	0	0	0
'LIFTING OFF'	0	0	0	0			0	0	0	0	0	0	0	0
COMMENTS														

KEY: PAIN SCALE (0-10) 0 'ok' 2 'allowed' 5 'not allowed' - seek advice 10

JAN. 2003

Wandsworth
Primary Care Trust

TRANSFEMORAL OSSEOINTEGRATION

LOADING PROGRAMME WITH THE SHORT TRAINING PROSTHESIS (STP)

NAME: Gemma Trotter.

D.O.B: 9/7/82

DATE OF STAGE (2) SURGERY: 28/4/04

BODY WEIGHT (WITHOUT PROSTHESIS): 55kg

DATE	14/6/04		15/6/04		16/6/04		17/6/04		18/6/04		19/6/04		20/6/04	
AM or PM	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm
LOADING WEIGHT	5kg	no loading	5kg	5kg	5-8kg	8-10kg	6-8kg	8-10kg	8-10kg	8-10kg	8-10kg	8-10kg	8-10kg	8-10kg
TIME LOADING	15 mins	—	15 mins	20 mins	25 mins	20 mins	15 mins	30 mins	30 mins	30 mins	30 mins	30 mins	30 mins	30 mins
PAIN SCALE 0-10 'BEFORE'	0	—	0	0	0	0	0	0	0	0	0	0	0	0
'DURING'	2	—	1	1	1/2	0	1/2	0	0	0	0	0	0	0
'LIFTING OFF'	2	—	1	1	1/2	0	1/2	0	0	0	0	0	1	0
COMMENTS	"wag"	—	'feeling'	'nervous feeling - not pain'	'feeling'	'feeling'	'feeling'	'feeling'	'feeling'	'feeling'	'feeling'	'feeling'	'feeling'	'feeling'

KEY: PAIN SCALE (0-10) 0 — 2 — 5 — 10
 'ok' 'allowed' 'not allowed' — seek advice

JAN. 2003

Appendix 7

```
//-----
#include <vcl.h>
#include <stdio.h>
#include <string.h>
#include <math.h>
#pragma hdrstop

#include "newproject.h"
//-----
#pragma package(smart_init)
#pragma resource "*.dfm"
FILE *rp,*wp;
TForm1 *Form1;
int SW=0;
//-----
__fastcall TForm1::TForm1(TComponent* Owner)
: TForm(Owner)
{
}
//-----

void __fastcall TForm1::Button1Click(TObject *Sender)
{
char header,character,comma;
```

```

AnsiString x,y,z;
int X1,Y1,X2,Y2,I,n,np;
float unit_convertor,c;
int point_counter;
int maxx,maxy,minx,miny,i;
float midx,midy;
double distance;
int a[1000],b[1000];
int A[2000],B[2000];
float X[2000],Y[2000],Z[2000];
Label5->Caption="Insert distance for new points!";
comma=',';
point_counter=1;
c=0.005;
I=1;
n=0;
x="";
y="";
i=1;
z=Edit4->Text;
if ((rp = fopen(z.c_str(), "rt")) == NULL)
    Application->MessageBox("The file does not
    exist","Warning",MB_OK | MB_ICONEXCLAMATION);
else
    if ((wp = fopen("resample.txt", "wt"))== NULL)
        Application->MessageBox("The file does not
        exist","Warning",MB_OK);
    else
    {
        header=fgetc(rp); //extracts the header from file
        header=fgetc(rp);
        header=fgetc(rp);
        header=fgetc(rp);
        header=fgetc(rp);
        while (header!='\n') //extracts the header from file
        {
            character=header;
            x+=character;
            header=fgetc(rp);
        }
        np=StrToInt(x); //get how many points in the file
        Label4->Caption=IntToStr(np); //output point number on window
        Label10->Caption="Number of points=";
        x="";
        unit_convertor=StrToFloat(Edit3->Text); //get unit convertor
        character=fgetc(rp);
        while(character!=';') //get x number
        {
            x+=character;
            character=fgetc(rp);
        }
    }

```

```

    }
    a[I]=StrToInt(x);    // give x to a[I]
    A[point_counter]=a[I];
    character=fgetc(rp);
    character=fgetc(rp);

    while(character!=';')    // get y number
    {
        y+=character;
        character=fgetc(rp);
    }
    b[I]= StrToInt(y);
    B[point_counter]=b[I];
    character=fgetc(rp);

/*fprintf(wp,"%d%c%f%c%f",point_counter,comma,X[point_counter],comma,
Y[point_counter]);
    fputc('\n',wp); */
    I++;
    n++;
    point_counter++;
    while(!feof(rp) && I < np)
    {
        x="";
        y="";

        character=fgetc(rp);
        while(character!=';')
        {
            x+=character;
            character=fgetc(rp);
        }
        a[I]=StrToInt(x);
        character=fgetc(rp);
        character=fgetc(rp);
        while(character!=';')
        {
            y+=character;
            character=fgetc(rp);
        }
        b[I]= StrToInt(y);
        character=fgetc(rp);
        distance=sqrt(((a[I]-a[I-n])*(a[I]-a[I-n]))+((b[I]-b[I-n])*(b[I]-
b[I-n]))));

        if (distance >= StrToFloat(Edit1->Text))
        {
            A[point_counter]=a[I];
            B[point_counter]=b[I];

```

```

/*fprintf(wp,"%d%c%f%c%f",point_counter,comma,X[point_counter],comma,
Y[point_counter]);
        fputc('\n',wp); */
        point_counter++;
        n=0;
    }
    I++;
    n++;
}
maxx=A[1];
minx=A[1];
maxy=B[1];
miny=B[1];
while (i<=point_counter)
{
    if (A[i]<A[i+1]&& A[i]<minx)
    {
        minx=A[i];
    }
    if(A[i]>A[i+1]&& A[i]>maxx )
    {
        maxx=A[i];
    }
    if (B[i]<B[i+1]&& B[i]<miny)
    {
        miny=B[i];
    }
    if(B[i]>B[i+1]&& B[i]>maxy )
    {
        maxy=B[i];
    }

    i++;
}
midx=(maxx+minx)/2;
midy=(maxy+miny)/2;
i=1;
while(i<point_counter)
{
    X[i]=(A[i]-midx)*unit_convertor;
    Y[i]=(B[i]-midy)*unit_convertor;
    Z[i]=Y[i]+c;
    fprintf(wp,"%d%c%f%c%f",i,comma,X[i],comma,Z[i]);
        fputc('\n',wp);
        i++;
}
Label21->Caption=FloatToStr(minx);
Label22->Caption=FloatToStr(maxx);

```



```

        Label23->Caption=FloatToStr(miny);
        Label24->Caption=FloatToStr(maxy);
        Label17->Caption="X offset=";
        Label18->Caption=FloatToStr(midx);
        Label19->Caption="Y offset=";
        Label20->Caption=FloatToStr(midy);

        Label12->Caption=IntToStr(point_counter-1); //output point
number on window
        Label11->Caption="Number of selected points=";
        Label3->Caption="ready for conversion";
        Button1->Enabled=true; //enable button1
        fclose(rp);
        fclose(wp);
    }
    Button1->Enabled=false;
    Button2->Enabled=true;
}
//-----

void __fastcall TForm1::Edit1Change(TObject *Sender)
{
    Label9->Caption="Insert z value for ANSYS file!";
    Edit2->Enabled=true;
}
//-----

void __fastcall TForm1::Edit2Change(TObject *Sender)
{
    Edit3->Enabled=true;

    Label8->Caption="Insert Unit conversion ratio!";
    SW++;
}
//-----

void __fastcall TForm1::Button2Click(TObject *Sender)
{
    char header,character,comma;
    /*AnsiString z; */
    float zaxis;
    comma=',';
    zaxis= StrToFloat(Edit2->Text); //get Z data
    /*z=Edit5->Text; */
    if ((rp = fopen("resample.txt", "rt")) == NULL)
        Application->MessageBox("The file does not exist","Warning",MB_OK |
MB_ICONEXCLAMATION);
    else

```

```

    if ((wp = fopen("ANSYSCODE.txt", "wt")) == NULL)
        Application->MessageBox("The file does not
exist", "Warning", MB_OK);
    else
    {
        fseek(wp, 0, SEEK_SET);
        character=fgetc(rp);
        fputc('/', wp);
        fputc('p', wp);
        fputc('r', wp);
        fputc('e', wp);
        fputc('p', wp);
        fputc('7', wp);
        fputc('\n', wp);
        while (!feof(rp))
        {

            if (character!='\n')
            {
                fputc('k', wp);
                fputc(',', wp);
                while (character!='\n')
                {
                    fputc(character, wp);
                    character=fgetc(rp);
                }
            }
            else
            {
                fprintf(wp, "%c%f", comma, zaxis);
                fputc('\n', wp);
                character=fgetc(rp);
            }
        }
        fclose(rp);
        fclose(wp);
    }
    Button2->Enabled=false;
}
//-----

```

```

void __fastcall TForm1::Label5Click(TObject *Sender)
{
    Label6->Enabled=true;
}
//-----

```

```

void __fastcall TForm1::Edit3Change(TObject *Sender)

```

```
{

    Label16->Caption="Input file name!";
    Edit4->Enabled=true;

}
//-----

void __fastcall TForm1::StartClick(TObject *Sender)
{
    Label5->Caption="Insert distance for new points!";
    Edit1->Enabled=true;
}
//-----

void __fastcall TForm1::Button3Click(TObject *Sender)
{
    Form1->Close();
}
//-----

void __fastcall TForm1::Edit4Change(TObject *Sender)
{
    Button1->Enabled=true;
}
//-----

void __fastcall TForm1::Button4Click(TObject *Sender)
{
    char header,character,comma;
    AnsiString x,y,z;
    int X1,Y1,X2,Y2,I,n,np,i;
    float unit_convertor,c,zaxis;
    int point_counter;
    float midx,midy;
    double distance;
    int a[1000],b[1000];
    int A[5000],B[5000];
    float X[5000],Y[5000],Z[5000];
    comma=',';
    point_counter=StrToInt(Edit5->Text);
    I=1;
    n=0;
    x="";
    y="";
```

```

i=1;
z=Edit4->Text;
zaxis= StrToFloat(Edit2->Text);
if ((rp = fopen(z.c_str(), "rt")) == NULL)
    Application->MessageBox("The file does not
exist", "Warning", MB_OK | MB_ICONEXCLAMATION);
else
    if ((wp = fopen("resample.txt", "wt")) == NULL)
        Application->MessageBox("The file does not
exist", "Warning", MB_OK);
    else
    {
        header=fgetc(rp); //extracts the header from file
        header=fgetc(rp);
        header=fgetc(rp);
        header=fgetc(rp);
        header=fgetc(rp);
        while (header!='\n') //extracts the header from file
        {
            character=header;
            x+=character;
            header=fgetc(rp);
        }
        np=StrToInt(x); //get how many points in the file
        Label4->Caption=IntToStr(np); //output point number on window
        Label10->Caption="Number of points=";
        x="";
        unit_convertor=StrToFloat(Edit3->Text); //get unit convertor
        character=fgetc(rp);
        while(character!=';') //get x number
        {
            x+=character;
            character=fgetc(rp);
        }
        a[I]=StrToInt(x); // give x to a[I]
        A[point_counter]=a[I];
        character=fgetc(rp);
        character=fgetc(rp);

        while(character!=';') // get y number
        {
            y+=character;
            character=fgetc(rp);
        }
        b[I]= StrToInt(y);
        B[point_counter]=b[I];
        character=fgetc(rp);

        /*fprintf(wp, "%d%c%f%c%f", point_counter, comma, X[point_counter], comma,
        Y[point_counter]);

```

```

fputc('\n',wp); */
I++;
n++;
point_counter++;
while(!feof(rp) && I < np)
{
    x="";
    y="";

    character=fgetc(rp);
    while(character!=';')
    {
        x+=character;
        character=fgetc(rp);
    }
    a[I]=StrToInt(x);
    character=fgetc(rp);
    character=fgetc(rp);
    while(character!=';')
    {
        y+=character;
        character=fgetc(rp);
    }
    b[I]= StrToInt(y);
    character=fgetc(rp);
    distance=sqrt(((a[I]-a[I-n])*(a[I]-a[I-n]))+((b[I]-b[I-n])*(b[I]-
b[I-n]))));

    if (distance >= StrToFloat(Edit1->Text))
    {
        A[point_counter]=a[I];
        B[point_counter]=b[I];

/*fprintf(wp,"%d%c%f%c%f",point_counter,comma,X[point_counter],comma,
Y[point_counter]);
        fputc('\n',wp); */
        point_counter++;
        n=0;
    }
    I++;
    n++;
}
i=1;
midx=StrToFloat(Label18->Caption);
midy=StrToFloat(Label20->Caption);
i=StrToInt(Edit5->Text);
while(i<(point_counter))
{
    X[i]=(A[i]-midx)*unit_convertor-0.08*zaxis;
    Y[i]=(B[i]-midy)*unit_convertor+0.005-0.05*zaxis;
    fprintf(wp,"%d%c%f%c%f",i,comma,X[i],comma,Y[i]);

```

```

        fputc('\n',wp);
        i++;

    }

    Label12->Caption=IntToStr(point_counter-1); //output point
number on window
    Label11->Caption="Number of selected points=";
    Label3->Caption="ready for conversion";
    Button1->Enabled=true; //enable button1
    fclose(rp);
    fclose(wp);
    }
    Button4->Enabled=false;
    Button5->Enabled=true;
}
//-----

void __fastcall TForm1::Button6Click(TObject *Sender)
{
    Button4->Enabled=true;
}
//-----

void __fastcall TForm1::Button5Click(TObject *Sender)
{
    char header,character,comma;
    float zaxis;
    comma=',';
    zaxis= StrToFloat(Edit2->Text); //get Z data
    if ((rp = fopen("resample.txt", "rt")) == NULL)
        Application->MessageBox("The file does not exist","Warning",MB_OK |
MB_ICONEXCLAMATION);
    else
        if ((wp = fopen("ANSYSCODE.txt", "wt")) == NULL)
            Application->MessageBox("The file does not
exist","Warning",MB_OK);
        else
        {
            character=fgetc(rp);
            fputc('/',wp);
            fputc('p',wp);
            fputc('r',wp);
            fputc('e',wp);
            fputc('p',wp);
            fputc('7',wp);
            fputc('\n',wp);
            while (!feof(rp))
            {

```

```
    if (character!='\n')
    {
        fputc('k',wp);
        fputc(',',wp);
        while (character!='\n')
        {
            fputc(character, wp);
            character=fgetc(rp);
        }
    }
    else
    {
        fprintf(wp,"%c%f",comma,zaxis);
        fputc('\n',wp);
        character=fgetc(rp);
    }
}
fclose(rp);
fclose(wp);
}
Button5->Enabled=false;
}
//-----
```

Appendix 8

From a mechanical standpoint, an implant system is a simple cantilever. Consider the free body diagram of an element of the abutment shown in Fig. 36.

Figure 115. The force analysis of an element in a vibrational beam.

where $M(x, t)$ is the bending moment, $V(x, t)$ is the shear force, $w(x, t)$ is the displacement and $f(x, t)$ is the external force per unit length of the abutment. Since the inertia force acting on the element of the abutment is

$$\rho A(x) dx \frac{\partial^2 w(x, t)}{\partial t^2} \quad (3)$$

the force equation of the motion in the z direction gives

$$-(V + dV) + f(x, t) dx + V = \rho A(x) dx \frac{\partial^2 w(x, t)}{\partial t^2} \quad (4)$$

where ρ is the mass density and $A(x)$ is the cross-sectional area of the beam. The moment equation of motion about the y axis passing through the point 0 in Fig 37 leads to

$$(M + dM) - (V + dV) dx + f(x, t) dx \frac{dx}{2} - M = 0 \quad (5)$$

By writing

$$dV = \frac{\partial V}{\partial x} dx \quad \text{and} \quad dM = \frac{\partial M}{\partial x} dx$$

and disregarding terms involving second powers in dx , Eqs.(4) and (5) can be written as

$$-\frac{\partial V(x,t)}{\partial x} + f(x,t) = \rho A(x) \frac{\partial^2 w(x,t)}{\partial t^2} \quad (6)$$

$$\frac{\partial M(x,t)}{\partial x} - V(x,t) = 0 \quad (7)$$

By using the relation $V = \partial M / \partial x$ from Eq. (6) becomes

$$-\frac{\partial^2 M(x,t)}{\partial x^2} + f(x,t) = \rho A(x) \frac{\partial^2 w(x,t)}{\partial t^2} \quad (8)$$

From the elementary theory of the bending of beams, the relationship between bending moment and deflection can be expressed as

$$M(x,t) = EI(x) \frac{\partial^2 w(x,t)}{\partial x^2} \quad (9)$$

where E is Young's modulus and $I(x)$ is the moment of inertia of the beam cross section about the y axis. Inserting Eq.(9) into Eq.(8), we obtain the equation of motion for the forced lateral vibration of a nonuniform beam:

$$\frac{\partial^2}{\partial x^2} \left[EI(x) \frac{\partial^2 w(x,t)}{\partial x^2} \right] + \rho A(x) \frac{\partial^2 w(x,t)}{\partial t^2} = f(x,t) \quad (10)$$

For a uniform beam, Eq.(10) reduces to

$$EI \frac{\partial^4 w(x,t)}{\partial x^4} + \rho A(x) \frac{\partial^2 w(x,t)}{\partial t^2} = f(x,t) \quad (11)$$

For free vibration, $f(x, t)=0$, and so the equation of motion becomes

$$c^2 \frac{\partial^4 w(x,t)}{\partial x^4} + \frac{\partial^2 w(x,t)}{\partial t^2} = 0 \quad (12)$$

where

$$c = \sqrt{\frac{EI}{\rho A}} \quad (13)$$

Since the equation of motion involves a second order derivative with respect to time and a fourth order derivative with respect to x , two initial condition and four boundary conditions are needed for finding a unique solution for $w(x, t)$. Usually, the values of lateral displacement and velocity are specified as $w_0(x)$ and $\dot{w}_0(x)$ at $t=0$, so that the initial conditions become

$$\begin{aligned} w(x, t=0) &= w_0(x) \\ \frac{\partial w}{\partial t}(x, t=0) &= \dot{w}_0(x) \end{aligned} \quad (14)$$

The free vibration solution can be found using the method of separation of variables as

$$w(x, t) = W(x)T(t) \quad (15)$$

Substituting Eq. (15) into Eq. (12) and rearranging leads to

$$\frac{c^2}{W(x)} \frac{d^4 W(x)}{dx^4} = -\frac{1}{T} \frac{d^2 T(t)}{dt^2} = a = \omega^2 \quad (16)$$

where $a = \omega^2$ is a positive constant. Equation (16) can be written as two equations:

$$\frac{d^4 W(x)}{dx^4} - \beta^4 W(x) = 0 \quad (17)$$

$$\frac{d^2 T(t)}{dt^2} + \omega^2 T(t) = 0 \quad (18)$$

where

$$\beta^4 = \frac{\omega^2}{c^2} = \frac{\rho A \omega^2}{EI} \quad (19)$$

The solution of Eq. (18) can be expressed as

$$T(t) = A \cos \omega t + B \sin \omega t \quad (20)$$

where A and B are constants that can be found from the initial conditions. For the solution of Eq. (17), we assume

$$W(x) = C e^{sx} \quad (21)$$

where C and s are constants, and derive the auxiliary as

$$s^4 - \beta^4 = 0 \quad (22)$$

The roots of this equation are

$$s_{1,2} = \pm \beta, \quad s_{3,4} = \pm i\beta \quad (23)$$

Hence the solution of Eq. (17) becomes

$$W(x) = C_1 e^{\beta x} + C_2 e^{-\beta x} + C_3 e^{i\beta x} + C_4 e^{-i\beta x} \quad (24)$$

where C_1 , C_2 , C_3 and C_4 are constants. Equation (24) can also be expressed as

$$W(x) = C_1 \cos \beta x + C_2 \sin \beta x + C_3 \cosh \beta x + C_4 \sinh \beta x \quad (25)$$

or

$$\begin{aligned} W(x) = & C_1 (\cos \beta x + \cosh \beta x) + C_2 (\cos \beta x - \cosh \beta x) \\ & + C_3 (\sin \beta x + \sinh \beta x) + C_4 (\sin \beta x - \sinh \beta x) \end{aligned} \quad (26)$$

where C_1 , C_2 , C_3 and C_4 , in each case, are different constants. The constants C_1 , C_2 , C_3 and C_4 can be found from the boundary conditions. The natural frequencies of the abutment are computed from Eq. (19) as

$$F = \frac{\omega}{2\pi} = \frac{\beta^2}{2\pi} \sqrt{\frac{EI}{\rho A}} = \frac{(\beta l)^2}{2\pi} \sqrt{\frac{EI}{\rho A l^4}} = \frac{\alpha^2}{2\pi} \sqrt{\frac{EI}{\rho A l^4}} \quad (27)$$

Where F = Natural Frequency (Hz); l =effective length of beam; A =area of cross section; E = Young's modulus; I = moment of inertia; ρ = the vibration mass per unit l, and α is a constant related to the boundary conditions.